

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110年 09 月

	零號冰箱		四號冰箱		五號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		1715-0116	
備用	2052-0129		1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max
2021/9/1	5.1	5.8	4.9	5.4	4.8	5.5	21.7	22.2
2021/9/2	4.8	5.7	4.8	5.3	4.8	6	22.1	22.3
2021/9/3	4.7	5.6	4.8	5.3	4.5	5.7	22.1	22.5
2021/9/4	4.7	5.3	4.8	5.3	4.8	5.6	22	22.3
2021/9/5	4.7	5.2	4.8	5.3	4.8	5.6	21.7	22
2021/9/6	4.6	5.4	4.8	5.3	4.8	5.7	21.7	22
2021/9/7	4.8	5.4	4.8	5.3	4.8	5.7	21.7	22.2
2021/9/8	4.7	5.4	4.8	5.7	4.7	5.9	21.9	22.3
2021/9/9	4.7	5.3	4.7	5.3	4.7	5.6	21.7	22.1
2021/9/10	4.6	5.5	4.7	5.2	4.8	5.9	21.8	22.2
2021/9/11	4.7	5.6	4.7	5.1	5.1	5.9	22	22.2
2021/9/12	4.7	5.6	4.7	5.1	5.1	5.9	21.9	22.2
2021/9/13	4.8	5.7	4.5	5.1	5.1	6.2	21.7	22.2
2021/9/14	4.9	5.6	4.5	4.9	5.1	5.9	21.7	22.3
2021/9/15	4.9	5.6	3.8	4.9	5.1	6.1	21.7	22
2021/9/16	4.7	5.6	3.8	4.4	5.3	6.1	21.7	22
2021/9/17	4.9	5.7	3.8	4.5	5.2	6.1	21.7	21.9
2021/9/18	4.9	5.7	3.9	4.5	5.2	6	21.4	21.7
2021/9/19	4.9	5.7	3.9	4.4	5.2	6	21.3	21.4
2021/9/20	4.9	5.7	3.9	4.4	5.2	6	21.1	21.3
2021/9/21	4.9	5.7	3.9	4.4	5.3	6	20.9	21.1
2021/9/22	4.9	5.7	3.9	4.5	5.3	6	20.9	21.3
2021/9/23	4.8	5.4	4.1	5.1	5.2	6	21.2	21.7
2021/9/24	4.7	6.1	4	4.7	5.2	6.3	21.4	22.2
2021/9/25	4.8	5.6	4	4.7	5.3	5.9	21.7	21.9
2021/9/26	4.8	5.4	4	4.7	5.3	5.9	21.3	21.6
2021/9/27	4.6	5.9	4	4.9	5.3	6.1	21.3	22
2021/9/28	4.7	5.4	4.2	4.9	5.1	6	21.4	22.1
2021/9/29	4.7	5.3	4.2	4.9	5.1	6	21.8	22.1
2021/9/30	4.7	5.6	4.2	4.9	5.1	6.6	21.7	21.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 溫度均在正常範圍內
藥學部 邱保祥 110
師(三)級藥師 001
藥學部 邱子芹 110
師(三)級藥師 001