

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 10 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2021/10/1	2.9	4.7	3.7	4.4	4.3	5.6	-20.4	-18.9	-81.3	-79.6	19.3	19.7
2021/10/2	3.7	4.7	3.9	4.4	4.3	4.5	-20.4	-20.2	-81.4	-79.7	19	19.3
2021/10/3	3.7	4.7	3.9	4.4	4.3	4.5	-20.4	-20.2	-81.5	-79.4	18.4	19
2021/10/4	3.7	4.7	3.7	4.5	4.3	5.8	-20.7	-19.6	-81.4	-79.7	18.4	19.4
2021/10/5	3.6	4.9	3.7	4.7	4.2	5.7	-20.7	-20.2	-81.5	-79	19.1	20.1
2021/10/6	3.9	5.1	3.6	4.7	4.2	5.7	-20.6	-19.8	-81.5	-79.3	19.2	19.8
2021/10/7	4	5.3	3.5	4.7	4.3	5.1	-20.3	-19.8	-81.5	-79	19.3	20.2
2021/10/8	3.9	5	3.6	4.8	4.3	4.9	-20.2	-19.9	-81.5	-79.6	19.3	19.8
2021/10/9	3.9	4.8	4	4.7	4.3	4.7	-20.2	-19.9	-81.5	-79.2	19.2	19.4
2021/10/10	3.9	4.7	4	4.7	4.3	4.7	-20.3	-19.9	-81.5	-79.6	19.1	19.4
2021/10/11	3.9	4.7	3.9	4.6	4.3	4.7	-20.3	-19.9	-81.5	-79.7	18.6	19.1
2021/10/12	4	4.8	3.9	5.2	4.2	4.8	-20.3	-19.9	-81.5	-79.3	18.9	19.9
2021/10/13	3.8	5	3.9	5.4	4.4	4.8	-20.4	-19.9	-81.3	-72.7	19.3	20
2021/10/14	3.7	4.7	3.9	5.7	4.5	4.9	-20.4	-19.9	-81.2	-79.6	19.4	20.4
2021/10/15	3.7	4.8	4	6.4	4.4	4.9	-20.4	-19.7	-81.3	-79.3	19.3	20.1
2021/10/16	4	4.8	3.9	4.7	4.6	4.9	-20.1	-19.8	-81.4	-79.4	19.2	19.4
2021/10/17	4	4.8	3.9	4.7	4.6	4.9	-20.1	-19.8	-81.4	-79.8	18.8	19.2
2021/10/18	4	4.9	3.9	6.1	4.4	4.9	-20.4	-19.8	-81.4	-79.6	18.7	19.2
2021/10/19	3.8	5.2	4	7.1	4.4	5	-20.4	-19.8	-81.3	-78.6	18.7	19.3
2021/10/20	3.9	4.8	4.1	6.8	4.4	4.9	-20.4	-19.7	-81.4	-79.5	18.7	19.3
2021/10/21	4.1	5.2	4.3	7.4	4.3	4.9	-20.4	-19.5	-81.3	-79.7	18.9	20.4
2021/10/22	3.6	4.8	4.3	7.4	4.4	4.9	-20.4	-19.9	-81.2	-79.7	20.1	20.8
2021/10/23	3.8	4.6	4.3	5.1	4.7	4.9	-20.2	-19.9	-81.2	-79.4	20.1	20.2
2021/10/24	3.9	4.6	4.3	5.1	4.7	4.9	-20.3	-19.9	-81.2	-79.7	19.9	20.1
2021/10/25	3.7	4.8	4.3	7.3	4.4	5.1	-20.4	-19.8	-81.2	-79.1	19.9	20.9
2021/10/26	3.7	5.7	4.3	7.3	4.3	5.5	-20.4	-18.5	-81.2	-78.9	20.1	21
2021/10/27	4.3	5.1	3.7	7.3	4.4	5.6	-20.2	-19.7	-81.2	-77.4	20.1	20.8
2021/10/28	4.2	5.3	3.7	4.7	4.5	5.4	-20.2	-19.6	-81.2	-75.7	19.6	20.6
2021/10/29	4	5	3.9	4.7	4.6	5.7	-20.4	-19.9	-81.3	-76.9	19.4	20.5
2021/10/30	3.9	4.8	4.2	4.7	5.5	5.7	-20.4	-20.1	-81.4	-78.9	19.9	20.4
2021/10/31	4	4.8	4.2	4.6	5.5	5.7	-20.4	-20.2	-81.4	-79.1	19.9	19.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 (10/10/10)

藥學部 邱子芹 (10/10/10)

總藥師: 藥學部 廖志峰 (10/10/10)

藥學部 廖志峰 (10/10/10)

業管主管: 藥學部 何沁沁 (10/10/10)

藥學部 何沁沁 (10/10/10)