

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 11 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2021/11/1	4.2	5.2	4	4.5	5.1	5.9	-20.9	-18.5	-81.4	-78.9	19.2	20.3
2021/11/2	3.7	5.1	3.6	5.2	4.9	5.8	-20.9	-19	-81.5	-79.3	18.9	19.4
2021/11/3	4	4.8	3.7	4.5	5	5.8	-19.8	-19.1	-81.5	-79.8	18.7	19.4
2021/11/4	3.9	4.9	3.7	4.6	4.6	5.3	-19.8	-19.1	-81.5	-79.8	18.8	19.5
2021/11/5	3.9	4.9	3.9	4.6	4.7	5.3	-19.7	-19.3	-81.4	-79.6	18.8	19.5
2021/11/6	3.9	4.8	4.1	4.5	5.1	5.3	-19.6	-19.6	-81.5	-79.7	18.7	18.9
2021/11/7	3.9	4.8	4.1	4.5	5.1	5.3	-19.6	-19.6	-81.4	-79.4	18.7	19.3
2021/11/8	3.7	4.8	3.7	4.5	4.6	5.4	-19.7	-19.3	-81.5	-79.2	18.1	19.3
2021/11/9	3.8	4.9	3.7	4.6	4.4	5.3	-19.8	-19.6	-81.4	-79	18.4	19.1
2021/11/10	3.9	5.1	4.1	4.7	4.9	5.4	-19.9	-19.7	-81.5	-77.6	18.6	19.2
2021/11/11	3.9	4.8	3.8	4.6	4.4	5.3	-19.9	-19.6	-81.4	-79	18.6	19.1
2021/11/12	3.9	4.8	4.2	4.7	4.9	5.3	-20.1	-19.6	-81.3	-79.7	18.6	19.1
2021/11/13	3.9	4.8	4.2	4.7	4.9	5.3	-19.9	-19.8	-81.4	-79.2	18.6	18.8
2021/11/14	3.9	4.8	4.2	4.9	4.9	5.3	-19.9	-19.7	-81.3	-79.7	18.4	19.3
2021/11/15	3.7	4.9	3.8	4.6	4.1	5.4	-19.9	-19.1	-81.2	-79.7	18.4	19.1
2021/11/16	3.7	5	4.1	4.6	4.7	5.1	-20.3	-18.5	-81.2	-79.6	18.6	19.2
2021/11/17	3.9	5.1	3.9	4.8	4.3	5.9	-20.7	-19.9	-81.2	-79.3	18.6	19.3
2021/11/18	3.3	4.8	3.9	4.4	4.3	4.7	-20.4	-19.6	-81.2	-75.1	18.8	19.3
2021/11/19	3.7	4.9	4	4.7	4.3	4.8	-20.2	-19.6	-81.2	-79.1	18.8	19.6
2021/11/20	3.8	4.7	4.4	4.7	4.3	4.7	-20.4	-19.3	-81.2	-79.6	18.8	19
2021/11/21	3.8	4.8	4.3	4.7	4.3	4.7	-20.4	-20.2	-81.2	-79.5	18.6	18.8
2021/11/22	3.8	5.1	4.3	4.8	4.3	5.3	-20.7	-20.1	-81.2	-79.3	18.6	19.2
2021/11/23	3.9	5.1	4.4	4.8	4.4	4.9	-20.7	-20	-81.2	-79	18.6	19.3
2021/11/24	3.9	4.9	4.3	4.7	4.4	4.9	-20.6	-20	-81.2	-78.7	18.6	19.2
2021/11/25	4	5.4	4.3	5.1	4.4	4.9	-20.6	-20.1	-81.2	-79.2	18.6	19.2
2021/11/26	4	5.1	4.3	5.2	4.4	4.9	-20.3	-19.8	-81.2	-79.7	18.7	19.2
2021/11/27	3.9	4.9	4.4	4.7	4.4	4.9	-20.3	-20	-81.2	-78.8	18.6	18.8
2021/11/28	4	4.9	4.4	4.7	4.4	4.8	-20.3	-20	-81.2	-79.7	18.4	18.6
2021/11/29	3.7	5	4.4	4.9	4.4	4.9	-20.4	-20	-81.2	-79.6	18.4	19.2
2021/11/30	3.8	4.9	4.4	5.2	4.4	4.9	-20.4	-20	-81.2	-79.5	18.6	19.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 (三)級藥師 110/201  
藥學部 邱子芹 (三)級藥師 110/131

總藥師: 藥學部 廖志峰 (二)級藥師 110/1530

業管主管: 藥學部 何沁沁 (一)級藥師 110/21