

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110年 11 月

	零號冰箱		四號冰箱		五號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		1715-0116	
備用	2052-0129		1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max
2021/11/1	5.9	6.2	4.7	5.3	4.3	5.7	21.3	21.5
2021/11/2	5.1	6.2	4.7	5.6	4.3	5.6	21.1	21.3
2021/11/3	4.9	6.6	4.9	5.8	3.6	5.6	19.7	21.3
2021/11/4	4.7	5.6	4.9	5.6	4.2	6.2	21.1	21.5
2021/11/5	4.8	5.6	5	5.6	4.2	5.6	21.1	21.6
2021/11/6	4.8	5.4	5.1	5.5	4.4	5.6	21.1	21.3
2021/11/7	4.8	5.4	5.1	5.5	4.4	5.6	20.9	21.2
2021/11/8	4.7	5.4	5	5.6	4.5	5.7	20.7	20.9
2021/11/9	4.7	5.4	4.9	5.6	4.5	5.7	20.6	20.7
2021/11/10	4.7	5.4	5.1	5.6	4.5	5.6	20.4	20.7
2021/11/11	4.9	5.4	5.1	5.6	4.4	5.7	20.4	20.7
2021/11/12	4.7	5.3	5	5.6	4.2	5.6	20.5	20.7
2021/11/13	4.7	5.3	5	5.5	4.4	5.5	20.5	20.6
2021/11/14	4.7	5.3	5	5.5	4.4	5.5	20.3	20.5
2021/11/15	4.7	5.3	5	5.8	4.4	5.5	20.2	20.5
2021/11/16	4.7	5.4	5.1	5.6	4.3	5.6	20.3	20.8
2021/11/17	4.5	5.6	5.1	5.6	4.3	5.4	20.6	20.9
2021/11/18	4.5	5.8	5.1	5.6	4.4	5.8	20.7	21.1
2021/11/19	4.8	5.7	5.1	5.7	4.3	5.4	20.8	21.3
2021/11/20	4.8	5.4	5.1	5.6	4.3	5.3	20.9	21.1
2021/11/21	4.8	5.4	5.1	5.6	4.3	5.3	20.6	20.8
2021/11/22	4.8	5.4	5.1	5.7	4.3	5.4	20.4	20.7
2021/11/23	4.8	5.4	5	5.6	4.3	5.4	20.6	20.8
2021/11/24	4.8	5.5	4.8	5.5	4.3	5.4	20.4	20.8
2021/11/25	4.9	5.9	4.8	5.4	4.3	5.5	20.6	21.1
2021/11/26	5.1	5.7	4.8	5.5	4.2	5.4	20.7	21.1
2021/11/27	5.1	5.7	4.8	5.3	4.3	5.4	20.6	20.9
2021/11/28	5.1	5.6	4.8	5.3	4.3	5.4	20.4	20.6
2021/11/29	5.1	5.7	4.8	5.5	4.3	5.4	20.3	20.6
2021/11/30	5.1	5.6	4.9	5.7	4.2	5.4	20.5	20.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥
師(三)級藥師 ¹¹⁰/₂₀₁
藥學部 邱子芹
師(三)級藥師 ¹¹⁰/₂₀₁

總藥師: 藥學部 廖志峰
師(二)級藥師 ²⁰¹/₁₅₃₀

業管主管: 藥學部 何沁沁
師(一)級藥師 ¹¹⁰/₂₀₂

溫度均在正常範圍內