

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 110 年 11月

序號	六號冰箱		室溫	
	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2021/11/1	4.5	6.6	19.2	19.8
2021/11/2	4.5	4.9	19.2	19.7
2021/11/3	4.4	5.1	19.3	20.3
2021/11/4	4.4	4.7	19.2	20.2
2021/11/5	4.3	4.7	19.4	20.8
2021/11/6	4.3	4.7	19.6	20.2
2021/11/7	4.3	4.7	19.2	20.8
2021/11/8	4.4	4.8	19.1	20.3
2021/11/9	4.5	4.8	18.6	19.3
2021/11/10	4.4	4.9	18.6	19.3
2021/11/11	4.5	4.9	18.3	19.3
2021/11/12	4.6	5	18.4	18.8
2021/11/13	4.7	4.9	18.6	18.8
2021/11/14	4.7	4.9	18.4	18.8
2021/11/15	4.6	4.9	18.3	19.2
2021/11/16	4.7	4.9	18.6	19.8
2021/11/17	4.6	4.9	19	20.3
2021/11/18	4.5	5.2	19.5	20.1
2021/11/19	4.5	5.1	19.3	19.9
2021/11/20	4.5	4.8	19.2	19.5
2021/11/21	4.5	4.8	18.9	19.5
2021/11/22	4.5	4.8	18.8	19.7
2021/11/23	4.5	4.9	18.6	19.1
2021/11/24	4.5	5.3	18.6	19.4
2021/11/25	4.5	4.9	18.8	19.5
2021/11/26	4.5	4.9	18.9	19.6
2021/11/27	4.5	4.9	18.8	19.3
2021/11/28	4.5	4.9	18.8	19.4
2021/11/29	4.5	4.9	19	19.5
2021/11/30	4.5	4.9	18.9	19.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥 (110)
師(三)級藥師 (20)
藥學部 邱子芹 (10)
師(三)級藥師 (20)

總藥師:

藥學部 廖志峰 (20)
師(二)級藥師 (530)

業管主管:

藥學部 何沁沁 (10)
師(一)級藥師 (202/11)