

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 12 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2021/12/1	3.5	5	4.3	4.9	4.6	5	-20.3	-19.2	-81.1	-78.6	19.4	19.9
2021/12/2	3.4	5.6	4.4	4.8	4.4	4.9	-20.4	-19.6	-81.1	-79.1	19.8	20.3
2021/12/3	3.8	5	4.5	5.2	4.4	4.8	-20.4	-20	-81.1	-78	19.7	20.2
2021/12/4	3.9	5.1	4.7	5.1	4.4	4.8	-20.4	-20.2	-81.1	-79.9	19.2	19.7
2021/12/5	4.1	5.2	4.7	5.1	4.4	4.8	-20.4	-20.2	-81.1	-79.8	19.1	19.2
2021/12/6	4.1	5.3	4.6	5.3	4.4	4.9	-20.6	-20.2	-81.1	-79.8	19.1	19.4
2021/12/7	3.9	5.3	4.3	5.1	4.4	4.8	-20.8	-20.4	-81.1	-79.8	19.2	19.5
2021/12/8	3.9	5.2	4.2	5.5	4.4	4.8	-20.9	-20.4	-81.1	-79.7	19.2	19.6
2021/12/9	4.1	5.4	4.1	4.6	4.4	4.9	-20.7	-20.2	-81.1	-79.6	19.3	19.7
2021/12/10	3.8	5.4	4	4.7	4.6	5.1	-20.7	-19.4	-81.1	-79.8	19.5	19.9
2021/12/11	4.2	5.2	4.1	4.4	4.6	5.1	-20.7	-20.4	-81.2	-79.6	19.7	19.7
2021/12/12	4.2	5.2	4.1	4.4	4.6	5.1	-20.7	-20.4	-81.2	-79.8	19.7	19.7
2021/12/13	4.2	5.3	4	4.5	4.6	5.1	-20.7	-20.2	-81.2	-79.3	19.4	19.6
2021/12/14	4.2	5.4	4	4.8	4.7	5.2	-20.8	-19.7	-81	-79.3	19.4	19.8
2021/12/15	4.2	5.4	4.1	4.6	4.7	5.2	-21	-19.7	-80.9	-79.1	19.5	19.8
2021/12/16	4.1	5.4	4	4.6	4.5	5.1	-21.2	-20.7	-81.1	-78.8	19.5	19.8
2021/12/17	4.1	5.5	4.1	4.9	4.5	5.1	-21.2	-20.8	-80.9	-78.6	19.5	19.8
2021/12/18	4.2	5.3	4.1	4.4	4.5	4.9	-21.2	-20.9	-81.1	-78.9	19.1	19.5
2021/12/19	4.2	5.4	4.2	4.4	4.5	4.8	-21.2	-20.9	-81.1	-78.8	18.8	19.1
2021/12/20	4.1	5.5	4.2	4.7	4.5	4.9	-21.2	-20.9	-81.4	-77.2	18.8	19.3
2021/12/21	4.2	5.3	4.2	4.9	4.5	4.9	-21.2	-20.7	-81.2	-78.3	19.2	19.6
2021/12/22	4.2	5.2	4.5	5.1	4.4	5.1	-20.9	-20.5	-81.2	-78.9	19.3	19.7
2021/12/23	4.2	5.3	4.6	5	4.6	5.1	-20.8	-20.2	-81.4	-78.9	19.4	19.7
2021/12/24	4.2	5.3	4.5	5.3	4.6	5.1	-20.9	-20.4	-81.4	-78.5	19.5	19.8
2021/12/25	4.2	5.5	4.5	5.7	4.7	5.7	-20.9	-20.7	-81.3	-78.5	19.4	19.6
2021/12/26	4.2	5.3	4.5	4.8	4.7	5	-20.9	-20.7	-81.4	-79.1	18.9	19.3
2021/12/27	4.2	5.3	4.5	4.9	4.7	5.1	-21.2	-20.8	-81.4	-78.7	18.8	19
2021/12/28	4.3	5.3	4.5	4.9	4.4	5.1	-21.2	-20.7	-81.4	-79.1	18.8	19.2
2021/12/29	4.2	5.3	4.2	4.9	4.4	5.1	-20.9	-20.2	-81.4	-79.1	19.1	19.4
2021/12/30	4.2	5.3	4.2	4.8	4.8	5.2	-20.9	-20.2	-81.4	-79.1	19	19.2
2021/12/31	4.3	5.3	4.2	4.7	4.9	5.2	-21.2	-20.8	-81.4	-78.5	18.7	19

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 111 0103 總藥師: 藥學部 廖志峰 1500 業管主管: 藥學部 何沁沁 110 0004 281

藥學部 邱子芹 111 0103