

臺北榮民總醫院藥學部
臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110年12月

	零號冰箱		四號冰箱		五號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		1715-0116	
備用	2052-0129		1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max
2021/12/1	4.9	5.7	5.1	5.8	4.2	5.3	20.6	21.2
2021/12/2	4.9	5.7	4.9	5.5	4.2	5.8	21.1	21.4
2021/12/3	4.9	5.7	4.8	5.4	4.1	5.3	20.7	21.4
2021/12/4	4.9	5.6	4.6	5.3	4.1	5.7	20.2	20.7
2021/12/5	4.9	5.6	4.6	5.1	4.7	5.7	20.1	20.3
2021/12/6	4.6	5.6	4.6	5.1	4.7	5.7	20.1	20.7
2021/12/7	4.6	6.1	4.6	5.2	4.7	5.8	20.3	20.8
2021/12/8	4.9	5.6	4.7	5.3	4.7	5.7	20.2	20.7
2021/12/9	4.9	5.8	4.7	5.4	4.7	6	20.5	21
2021/12/10	4.8	5.5	4.7	5.3	4.9	6	20.6	21.1
2021/12/11	4.9	5.5	4.7	5.3	4.9	5.8	20.7	20.9
2021/12/12	4.9	5.5	4.7	5.2	4.9	5.7	20.6	20.8
2021/12/13	4.9	5.6	4.7	5.3	4.9	5.7	20.4	20.8
2021/12/14	4.9	5.6	4.7	5.3	4.8	5.9	20.5	20.9
2021/12/15	4.9	5.8	4.7	5.3	4.7	5.7	20.7	21.3
2021/12/16	4.9	5.6	4.7	5.4	4.7	5.8	21.1	21.6
2021/12/17	4.7	5.4	4.7	5.3	4.9	6	20.7	21.2
2021/12/18	4.7	5.3	4.7	5.2	4.9	5.8	20.3	20.7
2021/12/19	4.7	5.3	4.7	5.2	4.9	5.8	19.9	20.3
2021/12/20	4.7	5.3	4.7	5.3	4.9	6.1	19.9	20.6
2021/12/21	4.7	5.3	4.7	5.3	4.9	6	20.2	20.6
2021/12/22	4.7	5.6	4.7	5.3	5.1	6	20.3	20.7
2021/12/23	5.1	5.6	4.7	5.3	5.1	6.1	20.5	20.7
2021/12/24	4.9	5.6	4.7	5.3	4.9	5.9	20.6	20.9
2021/12/25	4.9	6.1	4.7	5.5	5.2	6.3	20.5	20.9
2021/12/26	4.9	5.5	4.7	5.2	5.2	5.9	19.8	20.4
2021/12/27	4.9	5.6	4.7	5.3	5.1	5.9	19.6	19.8
2021/12/28	5.1	5.5	4.7	5.3	5.1	6	19.7	20.2
2021/12/29	4.9	5.6	4.7	5.3	4.9	6.1	20.1	20.5
2021/12/30	4.9	5.6	4.7	5.3	4.9	6.2	20.2	20.4
2021/12/31	4.9	5.6	4.7	5.3	5.2	5.9	19.9	20.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師 :

藥學部邱保祥
師(三級藥師) 0103

總藥師 :

藥學部廖志峰
師(二級藥師) 0103
1503

業管主管 :

藥學部何沁心
師一級藥師 009
34