

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111年 01 月

	零號冰箱		四號冰箱		五號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		1715-0116	
備用	2052-0129		1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max
2022/1/1	4.9	5.6	4.7	5.3	5.2	5.9	19.9	20.9
2022/1/2	4.9	5.6	4.7	5.3	5.1	5.9	20.4	20.8
2022/1/3	4.9	5.7	4.7	5.3	5.3	5.9	20.3	20.5
2022/1/4	4.8	5.7	4.7	5.2	5.1	5.9	20.3	20.7
2022/1/5	4.9	5.5	4.7	5.3	5.1	5.9	20.4	20.7
2022/1/6	4.9	5.6	4.7	5.3	5.1	6.2	20.4	20.9
2022/1/7	4.9	5.6	4.7	5.3	5.2	5.9	20.2	20.5
2022/1/8	4.8	5.5	4.7	5.2	5.2	5.9	19.9	20.2
2022/1/9	4.9	5.4	4.7	5.3	5.2	5.8	19.7	19.9
2022/1/10	4.9	5.6	4.7	5.3	5.1	5.8	19.7	20.2
2022/1/11	4.9	5.6	4.7	5.2	4.9	5.8	19.8	20.2
2022/1/12	4.8	5.4	4.7	5.3	4.6	5.7	19.5	19.9
2022/1/13	4.9	5.6	4.7	5.2	4.7	5.5	19.7	19.9
2022/1/14	4.9	5.4	4.7	5.2	4.7	5.3	19.4	19.8
2022/1/15	4.9	5.4	4.7	5.1	4.7	5.3	19.7	19.9
2022/1/16	4.9	5.4	4.7	5.2	4.7	5.3	19.7	19.9
2022/1/17	4.9	5.4	4.5	5.2	4.7	5.2	19.7	20.5
2022/1/18	4.9	5.4	4.4	5.1	4.7	5.4	20.3	21.1
2022/1/19	4.9	5.4	4.5	5.2	4.7	5.9	20.6	21
2022/1/20	4.9	5.5	4.7	5.3	4.8	5.8	20.6	21
2022/1/21	4.9	5.4	4.7	5.3	4.8	5.4	20.7	21.1
2022/1/22	4.9	5.4	4.7	5.3	4.7	5.6	20.8	21.1
2022/1/23	4.9	5.4	4.7	5.3	4.7	5.3	20.8	21.1
2022/1/24	4.9	5.4	4.7	5.3	4.7	5.3	20.6	21.1
2022/1/25	4.9	5.4	4.7	5.3	4.7	5.4	20.9	21.2
2022/1/26	4.9	5.4	4.3	5.4	4.6	5.3	20.9	21.2
2022/1/27	4.9	6.1	4.4	6.1	4.2	5.6	21.1	21.4
2022/1/28	5.6	6.1	4.7	5.3	4.3	5	21.1	21.4
2022/1/29	5.5	5.9	4.7	5.1	4.4	5	20.8	21.2
2022/1/30	5.5	5.9	4.7	5.1	4.4	5	20.3	20.8
2022/1/31	5.5	5.9	4.7	5.1	4.4	5.1	19.9	20.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥
師(三)級藥師 藥學部 邱子芹
師(三)級藥師

總藥師: 藥學部 廖志峰
師(二)級藥師

業管主管: 藥學部 何沁沁
師(一)級藥師