

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111 年 02 月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	0520-362	0520-292	1504-0305	1405-0193	1215-0006	1213-0007						
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/2/1	4.3	5.2	4	4.4	4.4	4.7	-20.9	-20.7	-80.6	-79.7	18.6	18.8
2022/2/2	4.3	5.2	4	4.4	4.4	4.7	-20.9	-20.7	-80.6	-79.8	18.7	18.9
2022/2/3	4.2	5.2	4	4.4	4.4	4.7	-20.9	-20.7	-80.6	-79.8	18.7	18.8
2022/2/4	4.2	5.1	3.7	4.4	4.3	4.9	-20.9	-20.7	-80.6	-79.7	18.6	19.1
2022/2/5	4.2	5.1	3.7	4.3	4.3	4.7	-20.9	-20.8	-80.6	-79.5	18.4	18.6
2022/2/6	4.2	5.2	3.7	4.3	4.3	4.7	-20.9	-20.8	-80.6	-79.8	18.3	18.4
2022/2/7	4.2	5.2	3.7	4.8	4.4	4.8	-20.9	-19.9	-80.6	-79.8	18.4	19.2
2022/2/8	4.2	5.2	4.2	4.8	4.4	4.9	-20.9	-20.4	-80.7	-79.8	18.6	19.3
2022/2/9	4.1	5.2	4.3	4.9	4.3	4.9	-20.9	-20.5	-80.7	-79.4	18.8	19.6
2022/2/10	4.1	5.1	4.3	5.1	4.3	4.8	-20.8	-20.4	-80.6	-79.7	18.9	19.6
2022/2/11	3.8	5.1	4.2	4.8	4.4	4.9	-20.8	-20.4	-80.6	-79.4	19	19.9
2022/2/12	3.9	4.8	4.3	4.7	4.5	4.9	-20.8	-20.6	-80.6	-79.6	19.2	19.4
2022/2/13	3.9	4.8	4.3	4.7	4.5	4.9	-20.9	-20.6	-80.6	-79.5	19	19.2
2022/2/14	3.8	4.8	4.3	4.8	4.4	4.9	-20.9	-20.4	-80.6	-78.9	18.9	19.8
2022/2/15	3.6	4.8	4.4	4.8	4.4	4.9	-20.7	-20.4	-80.6	-79.5	18.8	19.5
2022/2/16	3.6	5.1	4.1	5.1	4.3	5.2	-20.9	-20.4	-80.6	-79.7	19	19.6
2022/2/17	3.7	4.9	4.1	4.6	4.9	5.3	-21.1	-20.6	-80.6	-79.8	19	20.1
2022/2/18	3.6	4.9	4.2	4.8	4.9	5.3	-21	-20.4	-80.9	-78.7	19.2	19.8
2022/2/19	3.7	4.8	4.4	4.7	4.9	5.1	-20.9	-20.6	-81	-80	19	19.4
2022/2/20	3.8	4.8	4.3	4.7	4.9	5.1	-20.9	-20.8	-81.1	-79.9	18.4	19
2022/2/21	3.6	4.8	4.3	4.7	4.9	5.2	-20.9	-20.7	-81	-79.7	18.2	19
2022/2/22	3.7	4.8	4.3	4.7	4.9	5.2	-20.9	-20.6	-81	-79.9	18.6	19.2
2022/2/23	3.4	4.7	4.3	4.8	4.7	5.6	-20.8	-18.3	-81.1	-78.2	18.7	19.3
2022/2/24	3.7	4.9	4.3	4.8	4.7	5.1	-20.7	-20.2	-80.7	-79.8	18.6	19.1
2022/2/25	3.7	4.8	4.2	4.8	4.7	5.1	-21.1	-19.6	-80.7	-79.7	18.6	19.2
2022/2/26	3.9	4.8	4.2	4.6	4.7	4.9	-21.1	-20.8	-80.7	-79.5	18.8	18.9
2022/2/27	3.9	4.8	4.2	4.6	4.7	5	-21.1	-20.8	-80.7	-79.5	18.7	18.8
2022/2/28	3.9	4.8	4.2	4.7	4.7	4.9	-21.1	-20.8	-80.7	-79.7	18.8	18.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥  
師(三)級藥師 藥學部 邱子芹  
師(三)級藥師

總藥師: 藥學部 廖志峰  
師(二)級藥師

業管主管: 藥學部 何沁沁  
師(一)級藥師

溫度均在正常範圍內