

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111年 03 月

	零號冰箱		四號冰箱		五號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		1715-0116	
備用	2052-0129		1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max
2022/3/1	5.1	5.8	4.9	5.4	3.9	5.6	20.1	20.6
2022/3/2	4.9	5.5	4.8	5.4	4.4	5.8	20.5	20.9
2022/3/3	4.8	5.5	4.8	5.4	4.3	5.4	20.7	21.2
2022/3/4	4.9	5.4	4.9	5.6	4.3	5.7	20.9	21.2
2022/3/5	4.9	5.4	5.1	5.5	4.5	5.5	20.6	21.1
2022/3/6	4.9	5.4	5	5.5	4.5	5.5	20.3	20.6
2022/3/7	4.9	5.4	4.9	5.5	4.5	5.7	20.2	20.4
2022/3/8	4.9	5.4	4.9	5.4	4.4	5.8	20.1	20.3
2022/3/9	4.9	5.5	4.8	5.5	4.3	5.7	20.1	20.7
2022/3/10	4.9	5.6	4.8	5.4	4.3	5.6	20.5	21
2022/3/11	5.1	5.6	4.8	5.5	4.4	5.6	20.7	21.1
2022/3/12	5.1	5.4	4.8	5.3	4.4	5.6	20.9	21.1
2022/3/13	5.1	5.4	4.7	5.3	4.7	5.6	20.7	20.9
2022/3/14	5.1	5.4	4.7	5.8	4.3	5.6	20.6	21.1
2022/3/15	5.1	5.5	4.8	5.4	4.4	5.7	20.9	21.4
2022/3/16	5.1	5.4	4.8	5.4	4.7	5.8	21.2	21.7
2022/3/17	4.7	5.4	4.8	5.6	4.6	5.9	21.4	21.7
2022/3/18	4.7	5.3	4.3	5.8	4.8	5.9	21.6	21.8
2022/3/19	4.7	5.3	4.3	5	4.8	5.7	20.8	21.6
2022/3/20	4.7	5.3	4.3	4.9	4.7	5.7	20.2	20.7
2022/3/21	4.7	5.3	4.3	5	4.4	5.5	19.9	20.5
2022/3/22	4.7	5.6	4.3	5.1	4.4	5.5	20.1	20.7
2022/3/23	4.9	5.6	4.3	4.9	4.4	5.4	20.6	20.7
2022/3/24	4.6	5.5	4.3	4.9	4.3	6.7	20.4	20.9
2022/3/25	4.6	5.3	4.3	5	4.2	5.9	20.7	21
2022/3/26	4.6	5.3	4.4	4.9	4.4	5.4	20.7	22.4
2022/3/27	4.6	5.3	4.4	4.9	4.4	5.5	20.6	20.9
2022/3/28	4.6	5.3	4.3	5.3	4.4	5.7	20.4	20.8
2022/3/29	4.6	5.3	4.2	4.9	4.7	5.7	20.2	20.8
2022/3/30	4.5	5.3	4	6.1	4.7	6.1	20.3	20.7
2022/3/31	4.5	5.3	3.9	4.7	4.9	5.8	20.5	20.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥
師(三)級藥師 藥學部 邱子芹
師(三)級藥師

總藥師: 藥學部 廖志峰
師(二)級藥師

業管主管: 藥學部 何元元
師(一)級藥師

溫度均在正常範圍內