

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111 年 04 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/4/1	5.1	6.4	3.8	4.4	4	4.6	-20.7	-19.4	-81.4	-79.4	17.9	18.5
2022/4/2	5.2	5.8	3.9	4.4	4	4.5	-20.7	-20.4	-81.5	-80.2	17.8	18.1
2022/4/3	5.2	5.8	3.9	4.3	4	4.5	-20.7	-20.4	-81.5	-80.2	17.5	17.8
2022/4/4	5.2	5.8	3.9	4.3	4	4.5	-20.7	-20.4	-81.5	-80.1	17.5	17.6
2022/4/5	5.2	5.8	3.9	4.3	4	4.5	-20.7	-20.4	-81.4	-80.1	17.5	17.7
2022/4/6	5.1	5.9	3.7	4.3	4	4.6	-20.7	-20.1	-81.2	-79	17.5	18.3
2022/4/7	5.1	5.8	3.7	4.2	4.1	4.7	-20.4	-19.9	-81.2	-78.9	17.8	18.4
2022/4/8	5.1	5.9	3.7	4.2	4.1	4.7	-20.4	-19.9	-81.2	-79.9	17.9	18.9
2022/4/9	5.3	5.7	3.8	4.2	4.1	4.7	-20.4	-20.2	-81.4	-79.7	18.4	18.6
2022/4/10	5.3	5.7	3.7	4.2	4.1	4.7	-20.4	-20.2	-81.2	-79.7	18.4	18.6
2022/4/11	5.3	5.8	3.7	4.4	4	4.7	-20.4	-20.2	-81.2	-79.6	18.4	19.3
2022/4/12	5.2	5.8	3.5	4.2	4	4.5	-20.4	-19.9	-80.9	-79.7	18.6	19.3
2022/4/13	4.8	5.7	3.5	4.3	4	4.5	-20.8	-20	-80.9	-79.7	18.7	19.6
2022/4/14	5.1	5.9	3.7	4.2	4	4.5	-20.8	-20	-80.9	-78.8	18.8	19.4
2022/4/15	5.1	5.7	3.7	4.8	3.9	4.5	-20.4	-19.7	-80.9	-78.5	18.8	19.4
2022/4/16	5.1	5.6	3.8	4.3	3.9	4.4	-19.9	-19.9	-80.9	-78.2	18.6	18.9
2022/4/17	5.1	5.7	3.8	4.3	3.9	4.4	-20.1	-19.9	-80.9	-79	18.4	18.6
2022/4/18	4.9	5.7	3.8	4.3	3.9	4.6	-20.4	-19.9	-80.9	-79.9	18.4	19
2022/4/19	4.9	5.6	3.7	4.3	4.2	4.8	-20.4	-19.7	-80.9	-79	18.6	19.2
2022/4/20	4.9	6	3.7	6.7	4.2	4.7	-19.9	-19.5	-80.9	-78.2	18.6	19.2
2022/4/21	5	5.8	4.1	4.5	4.2	4.6	-20.4	-19.6	-80.9	-78.9	18.5	19.4
2022/4/22	4.9	5.7	4	4.5	4.2	4.7	-20.4	-19.7	-80.9	-79.1	18.9	19.6
2022/4/23	5.2	5.6	4	4.5	4.2	4.6	-19.9	-19.7	-80.9	-79.6	18.9	19.1
2022/4/24	5.2	5.6	4.1	4.5	4.2	4.6	-19.9	-19.7	-80.9	-79.8	18.8	18.9
2022/4/25	5.1	5.8	4	4.5	4	4.6	-20.4	-19.7	-80.9	-79.5	18.8	19.4
2022/4/26	4.9	5.9	4	4.5	4	4.5	-20.4	-19.6	-80.9	-79.4	18.7	19.3
2022/4/27	4.8	5.7	4	4.5	4	4.7	-20.2	-19.6	-80.9	-79.3	18.8	19.4
2022/4/28	5	5.7	4	4.6	4.1	4.6	-20.1	-19.6	-81.2	-79.1	18.7	19.4
2022/4/29	5.2	5.8	4	4.5	4.1	4.6	-20.6	-19.7	-81.4	-80.2	18.8	19.4
2022/4/30	5.2	5.7	4	4.4	4.1	4.6	-20.6	-20.3	-81.2	-80.2	18.9	19

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 111  
師(三級藥師) 0802 藥學部 邱子芹 111  
師(三級藥師) 0802

總藥師: 藥學部 廖志峰  
師(二級藥師) 0802  
1600

業管主管: 藥學部 何心心 0803  
師(一級藥師) 1130  
21