

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111 年 05 月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	0520-362	0520-292	0520-293	1504-0305	1504-0339	1213-0010	1330-0064	1405-0193	1215-0006	1213-0007	1405-0194	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/5/1	5.2	5.7	4	4.4	4.1	4.6	-20.6	-20.3	-81.4	-79.9	18.8	19
2022/5/2	5.1	6.2	3.9	4.5	4.2	4.7	-20.8	-19.3	-81.5	-79.6	18.9	19.4
2022/5/3	4.7	5.6	3.9	4.4	4	4.7	-20.8	-20.1	-81.5	-78.9	18.8	19.5
2022/5/4	4.6	5.4	3.9	4.5	3.9	4.6	-20.4	-19.9	-81.5	-79.8	18.9	19.7
2022/5/5	4.6	5.4	3.9	4.4	3.9	4.7	-20.8	-19.9	-81.5	-79.9	19	19.6
2022/5/6	4.5	5.4	4	4.7	3.9	4.6	-20.8	-19.8	-81.2	-79.6	19	19.8
2022/5/7	4.5	5.3	4.4	4.7	4.1	4.6	-20.2	-19.9	-80.8	-79.9	19.2	19.3
2022/5/8	4.5	5.3	4.4	4.7	4.1	4.6	-20.2	-19.9	-80.8	-79.9	19	19.2
2022/5/9	4.5	5.4	4.2	4.8	4	4.6	-20.2	-19.8	-80.9	-80.1	18.9	19.7
2022/5/10	4.5	5.4	4.2	4.9	3.9	4.6	-20.2	-19.4	-80.9	-79.8	19.2	19.9
2022/5/11	4.5	5.4	4.1	4.9	3.9	4.4	-20.2	-19.6	-80.9	-78.1	19.2	20.2
2022/5/12	4.5	6.4	4	4.6	4	4.6	-20.2	-19.6	-81.2	-77.9	19.3	20.2
2022/5/13	4.4	5.4	4	4.7	4	4.5	-20.2	-19.7	-80.9	-79.1	19.2	20.1
2022/5/14	4.4	5.3	4.3	4.7	4	4.4	-20.1	-19.8	-80.9	-79.4	19.2	19.3
2022/5/15	4.5	5.3	4.3	4.7	4	4.4	-20.1	-19.8	-80.9	-78.8	18.9	19.2
2022/5/16	4.5	6.1	4.3	4.9	4.1	4.5	-20.8	-20	-80.9	-79.4	18.8	20.2
2022/5/17	4.7	5.6	4.2	4.7	4	4.4	-20.8	-20.1	-81.2	-79.4	18.9	19.6
2022/5/18	4.7	5.6	4.2	4.6	4	4.8	-20.4	-19.8	-81.2	-79.6	18.9	19.7
2022/5/19	4.6	6.1	4.2	4.6	4.1	4.5	-20.2	-19.6	-81.2	-79.6	19	19.8
2022/5/20	4.5	5.7	4.2	4.7	3.6	4.4	-20.1	-19.6	-80.9	-79.6	19.1	20.1
2022/5/21	4.5	5.4	4.3	4.6	3.7	4.2	-20.2	-19.8	-81.2	-78.9	19	19.2
2022/5/22	4.5	5.3	4.2	4.5	3.7	4.2	-20.2	-19.8	-80.9	-77.8	18.8	19
2022/5/23	4.3	5.3	4.2	4.6	3.6	4.2	-20.1	-19.3	-81.2	-77.9	18.8	19.8
2022/5/24	4.3	5.3	4.2	4.8	3.6	4.2	-21.1	-19.3	-81.2	-79.8	19.4	20.1
2022/5/25	4.4	5.3	4.2	5.1	3.6	4.5	-20.8	-20.3	-81.5	-79.6	19.5	20.1
2022/5/26	4.4	5.3	4.2	5.1	3.5	4.2	-20.9	-20.4	-81.5	-78.3	19.5	20.1
2022/5/27	4.3	5.3	4.2	4.8	3.6	4.2	-21.2	-20.7	-81.5	-78.9	19.5	20.2
2022/5/28	3.7	5.2	4.2	4.5	3.6	4.2	-21.2	-20.8	-81.5	-79	19.5	20.1
2022/5/29	4	5.1	4.2	4.5	3.6	4.2	-21.2	-20.9	-81.5	-78.9	19.4	19.5
2022/5/30	4.1	5.1	4.2	4.6	3.5	4.2	-21.2	-20.2	-81.4	-79.4	19.3	19.9
2022/5/31	4.1	5.1	4.2	4.6	3.5	4.2	-20.9	-20.2	-81.2	-77.6	19.4	20.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 0601  
師(三)級藥師 藥學部 邱子芳  
師(三)級藥師

總藥師: 藥學部 廖志峰 602  
師(二)級藥師

業管主管: 藥學部 何沁沁 0602  
師(一)級藥師

溫度均在正常範圍內