

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111年 06 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/6/1	4.2	5.1	4.2	4.7	3.4	4.1	-19.9	-19	-80.7	-79.3	19.8	20.2
2022/6/2	4.2	5.2	4.3	4.6	3.3	4.2	-20.4	-19.6	-80.6	-75.6	19.7	20.3
2022/6/3	4.2	5.3	4.3	4.6	3.4	4.1	-20.4	-20.2	-80.6	-78.6	19.7	19.8
2022/6/4	4.2	5.3	4.1	4.7	3.3	4.4	-20.4	-20.2	-80.6	-77.9	19.6	19.9
2022/6/5	4.2	5.3	4.3	4.6	3.3	4.1	-20.4	-20.2	-80.6	-77.7	18.6	19.6
2022/6/6	4.2	5.4	4.3	4.6	3.4	4.2	-20.6	-19.7	-81.5	-78.4	18.4	18.9
2022/6/7	4.3	5.5	4.2	4.6	3.5	4.3	-20.1	-19.2	-81.5	-79.6	18.4	19
2022/6/8	4.2	5.4	4.2	4.5	3.4	4.2	-19.5	-18.8	-81.5	-79.2	18.4	19.2
2022/6/9	4.5	5.5	4	4.7	3.3	4.2	-18.9	-17.8	-81.4	-79.7	18.6	19.2
2022/6/10	4.6	5.7	4	4.7	3.4	4.3	-18.5	-17.5	-81.5	-79.1	18.6	19.2
2022/6/11	4.5	5.3	4.1	4.3	3.5	4.3	-18.5	-18.5	-81.4	-78.9	18.5	18.7
2022/6/12	4.5	5.3	4	4.3	3.5	4.3	-18.5	-18.5	-81.4	-79	18.4	18.5
2022/6/13	4.1	5.7	3.9	4.6	3.5	4.3	-19.6	-18.5	-81.4	-77.4	18.4	18.9
2022/6/14	4.5	5.7	3.9	4.5	3.5	4.3	-19.6	-19.2	-80.5	-78.9	18.4	18.8
2022/6/15	4.5	5.4	4	4.4	3.5	4.3	-19.6	-19.2	-80.4	-77.9	18.3	18.9
2022/6/16	4.4	5.3	4.1	4.7	3.6	4.3	-19.8	-19.3	-80.4	-74.7	17.6	18.8
2022/6/17	4.4	5.5	4.2	4.8	3.6	4.7	-19.8	-19.2	-80.4	-77.7	18.2	18.8
2022/6/18	4.4	5.5	4.3	4.7	3.7	4.3	-19.6	-19.3	-80.6	-78.9	18.2	18.3
2022/6/19	4.4	5.5	4.3	4.7	3.7	4.3	-19.6	-19.3	-80.6	-79	18.1	18.2
2022/6/20	4.3	5.6	4.1	4.7	3.6	4.3	-19.6	-19.3	-80.6	-77.8	18	18.7
2022/6/21	4.3	5.6	4.1	4.4	3.6	4.3	-19.8	-19.3	-80.6	-79.2	18.2	18.7
2022/6/22	4.3	5.4	4.1	4.7	3.6	4.3	-19.8	-19	-80.6	-79.1	18.2	18.8
2022/6/23	4.3	5.4	4.3	4.8	3.7	4.3	-19.7	-19.3	-80.4	-79.2	18.3	18.9
2022/6/24	4.3	5.3	4.2	4.7	3.7	4.4	-19.7	-19.3	-80.4	-79	18.3	18.8
2022/6/25	4.2	5.3	4.2	4.5	3.7	4.3	-19.7	-19.6	-80.4	-78.9	18.3	18.6
2022/6/26	4.2	5.3	4.2	4.5	3.7	4.3	-19.7	-19.5	-80.6	-77.9	18.2	18.3
2022/6/27	4.2	5.7	4.2	4.6	3.7	4.3	-19.8	-19.4	-81.1	-78.9	18.1	18.7
2022/6/28	4.2	5.3	4.2	4.5	3.3	4.4	-19.8	-19.3	-80.6	-78.4	18.2	18.9
2022/6/29	4.3	5.3	4.1	4.5	3.6	4.3	-19.6	-18.7	-80.6	-79.4	18.4	18.9
2022/6/30	4.3	6.6	4.1	4.5	3.4	4.3	-19.1	-18.5	-80.6	-78.9	18.4	19.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 111/0101 藥學部 邱子堯 111/2004  
 總藥師: 藥學部 廖志峰 111/2030

業管主管: 藥學部 何沁沁 111/1330

溫度均在正常範圍內