

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111年 06月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/6/1	4.6	5.3	3.6	4.9	4.7	6.2	5.5	6.1	22.1	23.1
2022/6/2	4.7	5.3	3.6	4.7	5.2	5.9	5.6	6.1	22.2	22.6
2022/6/3	4.7	5.3	3.8	4.7	5.3	5.9	5.6	6	22.2	22.6
2022/6/4	4.7	5.3	3.8	4.7	5.3	5.9	5.5	5.9	21.9	22.2
2022/6/5	4.7	5.3	3.8	4.6	5.3	5.9	5.4	5.9	21	22.2
2022/6/6	4.7	5.3	3.9	4.8	5.2	5.9	5.4	5.9	20.4	21.2
2022/6/7	4.7	5.3	3.9	5.1	5.1	6	5.5	6	21.1	21.4
2022/6/8	4.7	5.4	3.9	4.8	4.9	5.9	5.5	6	21.1	21.3
2022/6/9	4.9	5.6	3.9	4.7	4.9	5.9	5.4	5.9	21.1	21.3
2022/6/10	4.9	5.6	3.9	4.7	4.9	5.8	5.4	5.9	21.1	21.3
2022/6/11	4.9	5.6	3.9	4.6	5	5.8	5.4	5.9	21.1	21.3
2022/6/12	4.9	5.6	3.9	4.6	5.1	5.8	5.4	5.9	20.9	21.1
2022/6/13	4.9	5.5	3.8	4.7	4.9	5.9	5.4	5.9	20.7	21.2
2022/6/14	4.9	5.5	3.8	4.6	4.8	5.7	5.4	5.9	21.1	21.3
2022/6/15	4.9	5.5	3.8	4.6	4.9	5.8	5.4	5.9	21.1	21.3
2022/6/16	4.9	5.5	3.7	4.6	4.8	5.9	5.4	5.9	20.9	21.2
2022/6/17	4.8	5.6	3.7	4.7	4.8	5.8	5.4	5.9	20.9	21.1
2022/6/18	4.8	5.3	4	4.7	4.9	5.9	5.4	5.9	20.8	21.1
2022/6/19	4.8	5.3	4	4.7	4.9	5.8	5.4	5.9	20.6	20.8
2022/6/20	4.7	5.4	4	5.5	4.9	5.9	5.4	5.9	20.4	20.8
2022/6/21	4.8	5.3	4	4.8	4.9	5.9	5.4	6	20.7	21.1
2022/6/22	4.7	5.4	4	4.7	4.6	6.1	5.4	6	20.9	21.3
2022/6/23	4.7	5.4	3.9	4.7	4.7	5.9	5.6	6.1	21.1	21.3
2022/6/24	4.7	5.4	3.9	4.8	4.7	5.9	5.6	6.2	21.1	21.4
2022/6/25	4.7	5.3	3.9	4.7	4.7	5.9	5.6	6	21.1	21.3
2022/6/26	4.7	5.3	4	4.7	4.7	5.9	5.6	5.9	20.7	21.1
2022/6/27	4.7	5.4	4	4.7	4.7	5.8	5.6	6.1	20.6	21.1
2022/6/28	4.8	5.4	3.9	4.9	4.7	5.9	5.6	6.2	20.9	21.3
2022/6/29	4.8	5.3	3.9	5.1	4.9	6.1	5.7	6.2	21.3	21.6
2022/6/30	4.8	5.4	4.2	5.1	4.9	5.9	5.6	6.1	21.3	21.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師 藥學部 邱保祥
師(三)級藥師 藥學部 邱子芹
師(三)級藥師

總藥師 藥學部 廖志峰
師(二)級藥師

業管主管: 藥學部 何沁沁
師(一)級藥師