

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111 年 07 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/7/1	4.5	5.4	3.9	4.4	3.6	4.3	-20.2	-18.2	-80.2	-78.9	18.6	18.9
2022/7/2	4.5	5.3	3.9	4.3	3.6	4.3	-20.2	-19.9	-80.2	-78.6	18.4	18.9
2022/7/3	4.5	5.3	3.9	4.3	3.6	4.3	-20.2	-19.9	-80.2	-78.5	18.6	18.8
2022/7/4	4.5	5.7	3.9	4.6	3.5	4.4	-20.2	-19.1	-80.4	-78.4	18.6	19
2022/7/5	4.4	5.6	4.1	4.4	3.4	4.2	-19.8	-19.6	-80.3	-78.8	18.4	18.9
2022/7/6	4.3	5.6	4.1	4.8	3.4	4.2	-20.3	-19.6	-80.3	-79	18.2	18.9
2022/7/7	4.4	5.4	3.9	4.3	3.4	4.5	-20.4	-19.6	-80.2	-79	18.6	18.9
2022/7/8	4.4	5.9	3.6	4.5	3.4	4.2	-20.4	-19.7	-80.2	-78.1	18.6	19
2022/7/9	4.4	5.4	3.6	4.2	3.4	4.2	-19.9	-19.8	-79.9	-78.8	18.6	18.6
2022/7/10	4.3	5.4	3.6	4.2	3.4	4.2	-19.9	-19.8	-79.9	-78.8	18.4	18.6
2022/7/11	4.4	5.7	3.5	4.2	3.4	4.2	-19.9	-19.7	-80.2	-78.8	17.7	19.8
2022/7/12	4.4	5.7	3.5	4.3	3.4	4.5	-20.3	-19.6	-80.2	-78.9	18.6	19.1
2022/7/13	4.8	5.7	3.5	4.2	3.4	4.4	-20.4	-19.3	-80.2	-78.3	18.4	19
2022/7/14	4.3	5.7	3.7	4.2	3.4	4.2	-20.7	-19.8	-80.2	-78.5	18.2	19.1
2022/7/15	4.4	5.5	3.7	4.2	3.4	4.2	-20.7	-20.2	-80.2	-78.3	17.9	18.3
2022/7/16	4.4	5.4	3.7	4.2	3.6	4.3	-20.7	-20.3	-80.2	-78.3	17.7	17.9
2022/7/17	4.4	5.5	3.7	4.2	3.6	4.3	-20.7	-20.3	-79.9	-78.4	17.6	17.7
2022/7/18	4.3	5.5	3.7	4.2	3.3	4.3	-20.7	-20	-80.1	-78.2	17.5	18.3
2022/7/19	4.4	5.7	3.7	4.2	3.3	4.2	-20.4	-19.4	-80.1	-78.4	17.7	18.3
2022/7/20	3.6	5.7	3.6	4.5	3.4	4.3	-20.4	-19.9	-80.2	-78.6	17.7	18.3
2022/7/21	4.3	5.6	3.6	4.2	3.3	4.5	-20.4	-19.9	-80.2	-78.3	17.8	18.2
2022/7/22	4.4	5.5	3.6	4.2	3.3	4.2	-20.6	-19.9	-80.2	-78.4	17.8	18.5
2022/7/23	4.4	5.3	3.7	4.2	3.4	4.2	-20.4	-19.9	-80.2	-78.8	18.4	18.4
2022/7/24	4.4	5.3	3.7	4.2	3.4	4.2	-20.4	-19.9	-80.1	-78.9	18.4	18.4
2022/7/25	4.4	5.4	3.7	4.2	3.4	4.2	-20.7	-20	-80.1	-78.9	18.3	19.3
2022/7/26	4.4	5.6	3.7	4.4	3.4	4.2	-20.7	-20	-80.1	-77.9	18.6	19.2
2022/7/27	4.8	5.7	3.5	4.3	3.4	4.2	-20.4	-19.9	-80.3	-78	18.6	19.2
2022/7/28	4.8	6	3.5	4.2	3.3	4.2	-20.4	-19.9	-80.7	-77.5	18.6	19.4
2022/7/29	4.8	5.9	3.6	4.6	3.4	4.2	-20.4	-19.9	-80.8	-79.3	18.7	19.4
2022/7/30	4.8	5.6	3.5	4.2	3.4	4.2	-20.4	-20.2	-80.7	-78.9	18.6	18.9
2022/7/31	4.8	5.6	3.5	4.2	3.4	4.2	-20.4	-20.2	-80.7	-78.8	18.4	18.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 (三)級藥師 藥學部 邱予芹 (三)級藥師

總藥師: 藥學部 廖志峰 (二)級藥師

業管主管: 藥學部 何沁沁 (一)級藥師

溫度均在正常範圍內