

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111 年 08 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/8/1	4.9	5.8	3.3	4.3	3.3	4.1	-20.9	-18.7	-80.4	-79	18.7	19.6
2022/8/2	4.9	5.9	3.9	4.4	3.3	4.1	-20.9	-20.4	-80.6	-79	18.4	19.2
2022/8/3	4.5	6	3.9	4.6	3.3	4.1	-21.4	-20.7	-80.6	-76.4	18.4	19.2
2022/8/4	4.4	5.8	3.8	4.3	3.3	4.1	-21.4	-20.8	-80	-78.6	18.4	19.4
2022/8/5	4.4	5.9	3.9	4.5	3.3	4.1	-21.4	-20.9	-80.1	-77.9	18.4	19.2
2022/8/6	4.7	5.6	3.9	4.2	3.3	4.1	-21.4	-21.2	-80.2	-77.8	18.4	18.7
2022/8/7	4.7	5.6	3.9	4.2	3.3	4.2	-21.4	-21.1	-80.1	-77.9	18.3	18.4
2022/8/8	4.6	5.7	3.7	4.3	3.3	4.2	-21.7	-21.2	-80.1	-77.9	18.2	19.1
2022/8/9	4.5	5.7	3.7	4.2	3.4	4.2	-21.7	-21.2	-80.1	-78.3	18.3	18.9
2022/8/10	4.5	5.6	3.7	5	3.3	4.2	-21.7	-20.7	-80.1	-78.6	18.3	19.1
2022/8/11	4.5	5.6	3.8	4.4	3.3	4.2	-21.2	-20.7	-80.1	-78.7	18.4	19.4
2022/8/12	4.4	6.1	3.8	4.7	3.3	4.2	-21.2	-20.3	-80.7	-78.3	18.8	19.4
2022/8/13	4.4	5.3	3.9	4.4	3.3	4.2	-20.9	-20.6	-80.2	-78.9	18.6	18.8
2022/8/14	4.4	5.3	3.9	4.4	3.4	4.2	-20.9	-20.6	-80.2	-78.3	18.4	18.6
2022/8/15	4.4	5.6	3.9	4.4	3.3	4.2	-20.9	-20.4	-80.2	-78.3	18.3	19
2022/8/16	4.6	5.5	3.9	4.5	3.2	4.3	-20.8	-20.4	-80.2	-78.4	18.6	19.2
2022/8/17	4.4	5.7	3.9	4.5	3.3	4.2	-20.8	-19.7	-80.1	-78.8	18.6	19.6
2022/8/18	4.4	5.6	3.8	4.4	3.3	4.3	-20.6	-19.9	-80.1	-78.9	18.6	19.2
2022/8/19	4.4	5.4	3.9	4.8	3.3	4.3	-20.8	-19.9	-80.1	-78.8	18.6	19.4
2022/8/20	4.4	5.3	3.9	4.4	3.3	4.2	-20.9	-20.4	-80.4	-78.8	18.6	18.8
2022/8/21	4.4	5.4	3.9	4.4	3.3	4.2	-20.9	-20.7	-80.3	-78.3	18.4	18.6
2022/8/22	4.4	5.4	3.8	4.4	3.3	4.2	-20.9	-20.6	-80.4	-78.3	18.4	19.6
2022/8/23	4.4	5.7	3.8	4.4	3.3	4.1	-20.9	-20.1	-80.4	-78.8	19.1	19.7
2022/8/24	4.3	5.4	3.8	4.5	3.3	4.1	-21.1	-20.2	-80.3	-77.8	18.6	19.5
2022/8/25	4.3	5.4	3.8	4.3	3.2	4.2	-21.2	-20.4	-79.7	-77.3	18.4	21.7
2022/8/26	4.2	5.3	3.8	4.4	3.4	4.2	-21.2	-20.5	-80.2	-77.6	18.3	18.8
2022/8/27	4.2	5.1	3.8	4.3	3.4	4.2	-21.2	-20.8	-79.4	-78.3	18.1	18.4
2022/8/28	4.2	5.2	3.8	4.3	3.4	4.2	-21.2	-20.8	-79.4	-78.3	17.9	18.1
2022/8/29	4.2	5.2	3.8	4.4	3.3	4.2	-21.2	-20.4	-79.4	-78.4	17.9	18.9
2022/8/30	4.1	5.2	3.9	4.4	3.3	4.2	-21.2	-20.7	-79.4	-78.3	18.6	19.2
2022/8/31	4	5.1	3.8	4.5	3.3	4.2	-21.6	-20.8	-79.9	-76.1	18.7	19.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 0601 (三級藥師) 藥學部 邱子芹 0901 (三級藥師)  
 總藥師: 藥學部 廖志峰 0902 (二級藥師)  
 業管主管: 藥學部 何沁沁 0905 (一級藥師)