

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111 年 09 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/9/1	4.8	5.5	5.1	6	4.9	5.7	5.6	5.9	21.6	21.9
2022/9/2	4.5	5.9	4.8	5.7	4.9	5.7	5.2	6.1	21.5	21.8
2022/9/3	4.5	5.3	4.8	5.5	4.9	5.7	5.2	5.7	21.8	21.9
2022/9/4	4.5	5.4	4.7	5.4	4.9	5.7	5.1	5.6	21.7	21.9
2022/9/5	4.4	5.3	4.4	5.9	4.9	5.7	5.2	5.8	21.7	22.1
2022/9/6	4.4	5.5	4.4	6	4.6	6.1	5.2	5.8	21.9	22.2
2022/9/7	4.6	5.3	4.6	5.5	4.6	5.8	5.1	5.7	22.1	22.5
2022/9/8	4.7	6	4.5	5.5	4.8	5.7	5.1	5.8	22.3	22.7
2022/9/9	4.8	5.4	4.5	5.3	4.8	5.7	5.2	5.7	22.2	22.6
2022/9/10	4.8	5.4	4.5	5.3	4.9	5.7	5.2	5.7	22.1	22.2
2022/9/11	4.8	5.4	4.5	5.3	4.9	5.7	5.2	5.6	21.9	22.1
2022/9/12	4.8	5.4	4.5	5.3	4.9	5.8	5.1	5.7	21.9	22.2
2022/9/13	4.8	5.4	4.5	5.6	4.9	5.9	5.1	6	22	22.2
2022/9/14	4.7	5.4	4.6	5.5	4.9	5.7	5.7	6.1	22.1	22.4
2022/9/15	4.7	5.9	4.5	5.4	4.9	5.7	5.9	6.1	22.2	22.8
2022/9/16	4.7	5.2	4.5	5.3	4.9	5.9	5.9	6.2	22.4	23
2022/9/17	4.8	5.3	4.6	5.3	4.9	5.8	5.9	6.1	22.6	22.9
2022/9/18	4.8	5.3	4.6	5.3	4.9	5.8	5.9	6.1	22.3	22.6
2022/9/19	4.8	5.4	4.5	5.4	4.9	5.8	5.8	6.1	22.2	22.6
2022/9/20	4.8	5.6	4.4	6.7	4.9	5.9	5.8	6.3	22.4	22.8
2022/9/21	4.8	5.6	4.4	5.3	4.7	5.9	5.8	6.1	22.4	23
2022/9/22	4.9	5.4	4.5	5.5	4.7	6.2	5.8	6.1	22.6	22.9
2022/9/23	4.9	5.8	4.4	6.2	4.7	5.7	5.7	6.2	22.4	22.8
2022/9/24	4.8	5.4	4.4	5.3	4.7	5.6	5.7	5.9	22.6	22.8
2022/9/25	4.8	5.3	4.4	5.2	4.7	5.6	5.7	5.9	22.3	22.6
2022/9/26	4.8	5.5	4.4	5.4	4.7	5.7	5.4	6	22.2	22.6
2022/9/27	4.7	5.3	4.5	5.3	4.7	5.7	5.4	6.3	22.4	22.6
2022/9/28	4.9	5.5	4.4	5.3	4.9	6.1	5.9	6.3	22.4	22.6
2022/9/29	4.8	5.5	4.3	5.3	5.1	5.8	5.6	6.1	22.4	22.6
2022/9/30	4.8	5.3	4.3	5.1	5.1	5.9	5.6	5.9	22.3	22.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師 藥學部 邱保祥 1003  
藥學部 莊承芳 1003

總藥師: 藥學部 廖志峰 1603

業管主管: 藥學部 何沁沁 1005

溫度均在正常範圍內