

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表111年 09月

	六號冰箱		七號冰箱		室溫	
序號	1436-0086		2123-0026		1118-0011	
備用	0436-049		2123-0027		0742-0169	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2022/9/1	4.7	6.3	4.9	5.3	20.4	21.3
2022/9/2	5.1	5.4	4.9	5.3	20.5	20.9
2022/9/3	4.9	5.3	4.9	5.3	20.3	20.7
2022/9/4	4.9	5.3	4.9	5.3	20.2	20.6
2022/9/5	4.9	5.3	4.9	5.3	20.1	21
2022/9/6	4.8	5.3	4.8	5.2	20.3	20.7
2022/9/7	4.8	5.3	4.8	5.3	20.3	20.7
2022/9/8	4.9	5.3	4.8	5.2	20.3	20.5
2022/9/9	4.9	5.3	4.8	5.2	20.2	20.4
2022/9/10	4.9	5.3	4.8	5.2	20.1	20.3
2022/9/11	4.9	5.3	4.8	5.2	20.1	20.3
2022/9/12	4.9	5.7	4.8	5.2	20.1	20.9
2022/9/13	4.8	5.3	4.8	5.2	20.2	20.5
2022/9/14	4.6	5.1	4.7	5.2	20.2	20.5
2022/9/15	4.7	5.1	4.7	5.2	20.3	20.6
2022/9/16	4.7	5.1	4.7	5.2	20.3	20.4
2022/9/17	4.7	4.9	4.7	5.2	20.1	20.4
2022/9/18	4.7	4.9	4.7	5.2	20.1	20.3
2022/9/19	4.7	5.3	4.7	5.2	20.1	20.2
2022/9/20	4.6	4.9	4.8	5.2	20.1	20.2
2022/9/21	4.6	4.9	4.7	5.2	19.9	20.4
2022/9/22	4.4	5.3	4.7	5.2	19.7	20.1
2022/9/23	4.3	5.2	4.7	5.2	19.4	19.9
2022/9/24	4.7	5.2	4.7	5.2	19.5	19.8
2022/9/25	4.7	5.3	4.7	5.2	19.4	19.7
2022/9/26	4.8	5.4	4.7	5.2	19.4	20
2022/9/27	4.8	5.4	4.7	5.2	19.5	20
2022/9/28	4.8	5.4	4.7	5.2	19.5	20.3
2022/9/29	4.9	5.4	4.7	5.2	20.1	20.4
2022/9/30	4.9	5.4	4.8	5.2	19.9	20.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 1003
 藥學部 莊承芳 1003

總藥師: 藥學部 廖志峰 1003
 師二級藥師 1600

業管主管: 藥學部 何沁沁 1005
 師一級藥師 190