

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111 年 10 月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/10/1	4.7	5.3	3.9	4.3	3.4	4.2	-21.2	-20.9	-79.2	-78.1	19	19.2
2022/10/2	4.8	5.3	3.9	4.3	3.4	4.3	-21.3	-20.9	-79.2	-78.1	18.3	19
2022/10/3	4.8	5.5	3.9	4.4	3.4	4.3	-21.2	-19.1	-79	-76.9	18.4	18.8
2022/10/4	4.7	5.6	3.8	4.6	3.4	4.3	-21.2	-19.5	-79	-76.6	18.3	18.8
2022/10/5	4.8	6.4	3.6	4.5	3.3	4.3	-19.9	-19.6	-79.1	-76.5	18.3	18.9
2022/10/6	4.4	7.2	3.7	4.7	3.3	4.2	-20.2	-19.6	-78.9	-76.7	18.3	18.9
2022/10/7	4.2	5.4	3.7	4.2	2.9	5.7	-20.2	-19.6	-79.2	-77.1	18.4	18.8
2022/10/8	4.2	5.2	3.7	4.2	3.3	4.2	-20.2	-19.9	-78.9	-77.8	18.4	19.4
2022/10/9	4.2	5.3	3.7	4.2	3.3	4.2	-20.2	-19.9	-78.9	-77.9	19.4	19.4
2022/10/10	4.2	5.3	3.7	4.2	3.3	4.1	-20.2	-19.9	-79.1	-77.9	19.3	19.4
2022/10/11	3.9	5.2	3.7	4.4	3.2	4.2	-20.2	-19.6	-78.9	-77.9	19.2	19.8
2022/10/12	3.9	5.2	4	4.4	3.3	4.2	-20.1	-19.7	-78.9	-76.6	19.4	19.9
2022/10/13	3.9	5.3	4	4.7	3.4	4.4	-19.9	-19.6	-78.9	-76.9	19.4	19.9
2022/10/14	4	5.3	3.7	4.4	3.4	5	-19.9	-19.3	-78.9	-77.2	19.5	20.2
2022/10/15	4	5.1	3.7	4.1	4.5	4.9	-20.1	-19.8	-78.9	-76.9	19.6	20.1
2022/10/16	4	5.1	3.7	4.1	4.5	4.9	-20.1	-19.8	-78.9	-77.4	19.4	20.2
2022/10/17	4.1	5.3	3.7	4.1	4.3	4.9	-20.1	-19.6	-79.1	-77	19.2	20.1
2022/10/18	3.7	5.1	3.7	4.5	4.3	4.7	-20.2	-19.3	-78.9	-77.8	19.3	20.4
2022/10/19	3.4	5.1	3.9	4.5	4.3	4.6	-20.7	-19.9	-78.9	-77.3	19	19.6
2022/10/20	4	5.2	3.9	4.4	4.3	4.7	-20.7	-18.9	-79.8	-77.5	18.8	19.4
2022/10/21	4.1	5.3	3.1	6.2	4.3	4.7	-20.1	-19.7	-79.2	-77.8	18.8	19.4
2022/10/22	4.1	5.1	3.4	4.1	4.4	4.7	-19.9	-19.9	-78.9	-76.8	18.8	19
2022/10/23	4.1	5.1	3.4	4.1	4.4	4.7	-19.9	-19.9	-79.1	-76.4	18.6	18.8
2022/10/24	4.1	5.2	3.4	4.1	4.4	4.9	-20.6	-19.9	-79.1	-76.9	18.6	19.1
2022/10/25	4.1	5.2	3.4	4.7	4.5	4.9	-20.8	-20.3	-79.1	-76.9	18.6	20.4
2022/10/26	3.8	5.4	3.6	4.4	4.5	5	-20.9	-20.6	-79.2	-77.7	20.2	21.1
2022/10/27	3.5	5.1	3.9	4.4	4.4	5.1	-20.9	-20.5	-78.9	-77.8	20.3	20.8
2022/10/28	3.5	5.1	3.9	4.7	4.5	5	-20.8	-20.5	-78.9	-77.8	20.1	20.8
2022/10/29	3.6	5.1	4	4.7	4.6	4.9	-20.9	-20.6	-78.9	-76.7	18.4	20
2022/10/30	3.6	5.1	4	4.6	4.6	4.9	-20.9	-20.7	-78.9	-76.7	18.2	18.4
2022/10/31	3.2	5.1	3.9	4.7	4.6	5	-21.2	-20.7	-79.2	-77.8	18.2	18.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 (11)
藥學部 邱予芹 (10)

總藥師: 藥學部 廖志峰 (10)
藥學部 廖志峰 (10)

業管主管: 藥學部 何沁沁 (10)

溫度均在正常範圍內