

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表111年10月

	六號冰箱		七號冰箱		室溫	
序號	1436-0086		2123-0026		1118-0011	
備用	0436-049		2123-0027		0742-0169	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2022/10/1	4.9	5.4	4.8	5.2	19.5	19.9
2022/10/2	5	5.4	4.8	5.2	19.4	19.6
2022/10/3	5.4	5.9	4.7	5.2	19.4	19.7
2022/10/4	4.8	5.8	4.7	5.2	19.5	20.1
2022/10/5	4.7	5.2	4.7	5.2	19.8	20.2
2022/10/6	4.7	5.2	4.7	5.3	19.8	20.2
2022/10/7	4.7	5.6	4.8	5.3	19.8	20.1
2022/10/8	5.3	5.6	4.9	5.3	19.7	20.1
2022/10/9	5.3	5.6	4.9	5.2	19.5	19.9
2022/10/10	5.3	5.6	4.8	5.2	19.2	19.5
2022/10/11	4.9	5.6	4.8	5.2	19.1	19.4
2022/10/12	4.9	5.3	4.8	5.3	19.2	19.9
2022/10/13	4.9	5.9	4.8	5.3	19.4	20.8
2022/10/14	5.2	5.5	4.8	5.2	20.1	20.3
2022/10/15	5.2	5.4	4.9	5.3	20.2	20.9
2022/10/16	5.1	5.4	4.8	5.3	20.2	21.2
2022/10/17	5.2	5.4	4.8	5.2	20.1	20.4
2022/10/18	5.2	6.1	4.8	5.3	19.7	20.1
2022/10/19	5.1	5.7	4.8	5.2	19.6	19.9
2022/10/20	5.1	5.7	4.8	5.3	19.7	20.2
2022/10/21	5.1	5.6	4.8	5.3	19.9	20.4
2022/10/22	5.3	5.7	4.8	5.3	19.9	20.7
2022/10/23	5.4	5.7	4.8	5.3	19.7	20.4
2022/10/24	5.3	5.9	4.8	5.3	19.7	19.9
2022/10/25	5.4	6.3	4.8	5.2	19.7	19.9
2022/10/26	5.3	6	4.9	5.3	19.8	20.7
2022/10/27	5.3	5.5	4.9	5.3	20.1	20.7
2022/10/28	5.3	5.6	4.9	5.3	20.3	20.7
2022/10/29	5.3	5.6	4.9	5.3	20.3	20.7
2022/10/30	5.3	5.6	4.9	5.3	20.4	20.8
2022/10/31	5.3	6.7	4.9	5.3	20.1	20.6

註1：溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2：溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師 : 藥學部邱保祥
師(三級藥師) 1101

總藥師 : 藥學部廖志峰
師(二級藥師) 1102

業管主管 :

藥學部何沁沁
師(一級藥師) 1102