

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111 年 11 月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362	0520-292	1504-0305	1405-0193	1215-0006	1213-0007							
0850-0835	0520-293	1504-0339	1213-0010	1330-0064	1405-0194							
範圍	2~8°C	2~8°C	2~8°C	負15~負25°C	≤負70°C	15~25°C						
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/11/1	3.9	5.1	3.8	4.5	4.6	4.9	-20.8	-19.1	-79	-71.5	18.3	18.9
2022/11/2	3.9	5.1	3.8	4.5	4.6	4.9	-20.9	-20.5	-78.9	-77.8	18.2	18.9
2022/11/3	4	5.3	3.8	4.5	4.6	4.9	-20.9	-20.5	-78.9	-77.8	18.2	18.9
2022/11/4	3.7	5.1	3.6	4.4	4.6	4.9	-20.9	-20.4	-78.9	-77.6	18.2	19.2
2022/11/5	3.7	5.4	3.6	4.2	4.7	5	-20.9	-20.6	-78.9	-77.6	18.1	18.9
2022/11/6	4.2	5.4	3.6	4.2	4.7	4.9	-20.9	-20.6	-78.9	-77.2	18	18.6
2022/11/7	4.1	5.5	3.5	4.4	4.7	5.2	-21.2	-20.4	-78.9	-77.3	18.4	19.3
2022/11/8	4.1	5.3	3.5	4.2	4.7	5.1	-21.2	-20.4	-78.9	-77.3	18.7	19.6
2022/11/9	4	5.3	3.4	4.2	4.7	5.1	-21.2	-20.5	-78.9	-77.4	18.8	19.7
2022/11/10	3.9	5.5	3.6	4.1	4.7	5.2	-21.2	-19.9	-78.9	-77.4	18.8	19.6
2022/11/11	3.9	5.5	3.6	4.2	4.7	5.2	-20.7	-20.3	-78.9	-77.3	18.8	19.5
2022/11/12	3.9	5.5	3.6	4	4.7	5	-20.7	-20.4	-78.9	-77.6	18	18.9
2022/11/13	4	5.5	3.6	4.1	4.7	5	-20.8	-20.5	-78.9	-77.2	17.7	18.8
2022/11/14	3.5	5.6	3.6	4.1	4.7	5	-20.9	-20.5	-78.9	-77.7	18.2	19.2
2022/11/15	3.7	5.1	3.5	4.2	4.7	5.1	-20.9	-20.2	-79.1	-77.4	18.4	19.2
2022/11/16	3.6	5.1	3.6	4.3	4.7	5.1	-20.8	-20.3	-79.1	-77	18.6	19.3
2022/11/17	3.6	5.2	3.7	4.3	4.7	5.1	-20.8	-20.5	-79.2	-77.5	18.7	19.4
2022/11/18	3.9	5.1	3.7	4.3	4.7	5.1	-20.9	-20.3	-79.2	-77.6	18.7	19.6
2022/11/19	3.9	5	3.7	4.3	4.7	5.1	-20.9	-20.7	-79.2	-77.4	18.7	18.8
2022/11/20	3.9	5	3.7	4.2	4.7	5.1	-20.9	-20.7	-78.9	-77.7	18.6	18.7
2022/11/21	3.9	5.2	3.7	4.4	4.7	5.1	-20.9	-20.5	-78.9	-77.6	18.4	19.4
2022/11/22	3.9	4.9	3.9	4.3	4.6	4.9	-20.9	-20	-78.9	-76.5	18.6	19.5
2022/11/23	4.2	5.1	3.9	4.3	4.6	4.9	-20.4	-19.9	-78.9	-77.7	18.8	19.5
2022/11/24	4.2	6.1	3.9	4.4	4.6	5.1	-20.3	-19.8	-78.9	-77.3	18.8	19.9
2022/11/25	4.2	5.2	3.9	4.5	4.4	4.8	-20.4	-19.8	-78.9	-77.1	18.8	19.6
2022/11/26	4.2	5.2	4.2	4.5	4.4	4.7	-20.4	-20.2	-78.9	-76.7	18.8	19
2022/11/27	4.2	5.2	4.2	4.4	4.4	4.7	-20.7	-18.5	-78.9	-77.2	18.6	18.8
2022/11/28	4.1	5.4	4.1	4.4	4.4	4.8	-20.8	-19.9	-78.9	-77.4	18.6	19.2
2022/11/29	3.8	5.9	4.1	4.4	4.4	4.7	-20.3	-19.9	-78.9	-77.6	18.6	19.3
2022/11/30	3.8	5.3	4.1	4.6	4.4	4.9	-20.1	-20	-78.9	-77.7	18.6	19.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 201
藥學部 邱子芹 1245

總藥師: 藥學部 廖志峰 1330

業管主管: 藥學部 何沁沁 1205 0830

溫度均在正常範圍內