

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 111 年 12 月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362	0520-292	1504-0305	1405-0193	1215-0006	1213-0007							
0850-0835	0520-293	1504-0339	1213-0010	1330-0064	1405-0194							
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2022/12/1	4	5.4	3.8	4.7	4.7	5.1	-20.6	-19.6	-79.3	-77.4	19.6	19.9
2022/12/2	4.1	5.4	3.8	4.4	4.6	5	-20.6	-19.9	-80.1	-77.2	19.6	19.9
2022/12/3	4.1	5.3	3.9	4.4	4.6	4.9	-20.3	-20	-80.1	-77.7	19.6	19.8
2022/12/4	4.2	5.3	3.9	4.4	4.6	4.9	-20.4	-20.1	-80.1	-77.8	19.3	19.6
2022/12/5	4.1	5.3	3.9	4.4	4.6	5.1	-20.7	-20.1	-80.1	-79.2	19.3	19.6
2022/12/6	4.1	5.3	3.9	4.4	4.6	5	-20.7	-20.4	-80.1	-77.9	19.3	19.7
2022/12/7	4	5.1	3.9	4.5	4.7	4.9	-20.7	-19.6	-80.1	-77.7	19.4	19.8
2022/12/8	3.9	5	4	5.2	4.6	4.9	-20.6	-20.1	-80.1	-77.9	17.9	19.8
2022/12/9	3.9	5.1	3.9	5.2	4.6	5.1	-20.7	-20	-80.1	-79.2	19.4	19.8
2022/12/10	3.9	5	3.9	4.3	4.7	4.9	-20.7	-20.4	-80.1	-79.2	19.2	19.4
2022/12/11	3.9	4.9	3.9	4.3	4.7	4.9	-20.7	-20.4	-80.1	-79	19	19.3
2022/12/12	3.9	5.1	3.9	4.3	4.3	5.1	-20.8	-20.4	-80.1	-79.2	18.9	19.4
2022/12/13	3.8	5.1	3.8	4.5	4.3	4.8	-20.8	-20.4	-80.1	-79.2	19.1	19.4
2022/12/14	3.9	5.1	3.7	4.3	4.3	5.1	-21.1	-19.9	-80.2	-78	19.2	19.5
2022/12/15	4	5.1	3.7	4.3	4.6	5.2	-21.2	-20.8	-80.1	-78.6	19.2	19.7
2022/12/16	3.9	5.1	3.7	4.2	4.2	5.1	-21.2	-20.7	-80.1	-78.2	19.7	20.1
2022/12/17	4	5.1	3.7	4.2	4.2	4.6	-21.1	-20.8	-80.1	-78.3	19.4	19.7
2022/12/18	4.1	5.1	3.7	4.2	4.3	4.6	-21.2	-20.8	-80.2	-78.7	18.4	19.5
2022/12/19	3.8	5.1	3.7	4.2	4.2	4.6	-21.2	-20.7	-80.3	-73.1	18.3	18.6
2022/12/20	4.1	5.2	3.7	4.3	4.2	4.6	-21.2	-20.7	-80.3	-78.2	18.4	19.1
2022/12/21	4.2	5.2	3.7	4.6	4.2	4.7	-21.2	-20.7	-80.2	-78.8	18.8	19.2
2022/12/22	4.2	5.2	3.8	4.3	4.3	5.2	-21.2	-20.7	-80.2	-79.2	18.6	19
2022/12/23	4.2	5.2	3.8	4.3	4.4	5	-21.1	-20.8	-80.3	-77.2	18.4	18.7
2022/12/24	4.3	5.2	3.9	4.3	4.4	4.9	-21.1	-20.8	-80.3	-78.1	18.1	18.4
2022/12/25	4.3	5.2	3.9	4.3	4.4	4.9	-21.1	-20.8	-80.3	-77.8	18	18.1
2022/12/26	4.3	5.3	3.8	4.4	4.3	5.5	-21.2	-20.8	-80.3	-78.3	18.1	18.6
2022/12/27	4.2	5.6	3.9	4.7	4.3	4.7	-21.4	-20.8	-80.3	-77.7	18.4	19.2
2022/12/28	4.1	5.1	3.7	4.4	4.3	4.9	-21.4	-20.9	-80.1	-78.3	19.1	19.9
2022/12/29	4.1	5.1	3.7	4.3	4.3	4.7	-21.4	-20.9	-80.1	-78.8	19.9	20.4
2022/12/30	4.1	4.9	3.7	4.4	4.3	4.6	-21.4	-20.9	-80.1	-78.8	20.3	20.7
2022/12/31	4.2	4.9	3.8	4.3	4.3	4.7	-21.4	-21	-80.1	-78.7	20.3	20.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 12/30/2022 藥學部 邱子芹 12/30/2022

總藥師: 藥學部 廖志峰 12/30/2022

業管主管: 藥學部 何沁沁 12/30/2022