

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112 年 01 月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/1/1	3.9	4.9	3.8	4.3	4.4	4.8	-21.4	-20.7	-80.1	-77.7	20.3	22.6
2023/1/2	4	5.1	3.8	4.3	4.4	4.7	-21.4	-20.9	-80.1	-78.8	20.6	22.2
2023/1/3	4.1	4.8	3.9	4.5	4.4	4.7	-21.9	-18.8	-80.3	-79.9	18.8	19.4
2023/1/4	4.1	5.1	3.8	4.7	4.4	4.8	-21.9	-21.4	-80.3	-77.8	18.7	19.4
2023/1/5	4.1	5.1	3.9	4.6	4.4	4.8	-21.7	-21.2	-80.2	-78.3	18.6	19.3
2023/1/6	4.2	4.8	3.7	4.4	4.4	4.7	-21.9	-21.2	-80.2	-79.3	18.7	19.3
2023/1/7	4.2	4.9	3.8	4.2	4.4	4.8	-21.5	-21.2	-80.2	-78.4	18.4	18.8
2023/1/8	4.2	4.9	3.7	4.2	4.4	4.7	-21.4	-21.2	-80.2	-78.1	18.3	18.4
2023/1/9	4.1	4.9	3.8	4.5	4.4	4.9	-21.4	-20.8	-80.2	-78.2	18.3	19.2
2023/1/10	3.9	4.8	3.7	4.3	4.4	4.8	-21.4	-20.5	-80.3	-79.1	18.4	19.3
2023/1/11	3.9	4.9	3.7	4.5	4.3	4.9	-21.7	-21.2	-80.1	-79.4	18.6	19.3
2023/1/12	3.9	4.9	3.7	4.3	4.4	4.8	-21.9	-21.5	-80.1	-79.4	18.7	19.5
2023/1/13	4.2	5.1	3.7	4.7	4.4	4.8	-21.9	-20.7	-80.1	-78.2	18.7	19.4
2023/1/14	4.2	5.1	3.7	4.2	4.4	4.7	-21.2	-20.9	-80.2	-78.6	18.7	18.9
2023/1/15	4.2	5.1	3.7	4.2	4.4	4.7	-21.2	-20.9	-80.2	-78.3	18.3	18.8
2023/1/16	4.2	5.2	3.7	4.3	4.4	4.8	-21.2	-20.3	-80.2	-78.4	18.1	18.6
2023/1/17	4.2	5.3	3.6	4.4	4	4.7	-20.9	-20.1	-80.2	-78.2	18	18.8
2023/1/18	4.2	5.3	3.6	4.4	4.2	4.7	-20.7	-20.1	-80.1	-79.4	18.1	18.9
2023/1/19	4.2	5.2	3.4	4.2	4.3	4.7	-20.4	-20	-80.1	-77.9	18.9	19.7
2023/1/20	4.2	5.1	3.5	4	4.3	4.7	-20.6	-20.2	-80.3	-78.3	19.2	19.3
2023/1/21	4.2	5.1	3.5	4	4.3	4.7	-20.4	-20.2	-80.2	-78.6	19.2	19.3
2023/1/22	4.2	5.1	3.5	4	4.3	4.7	-20.6	-20.2	-80.2	-79	19.2	19.3
2023/1/23	4.2	5.1	3.5	4	4.3	4.7	-20.6	-20.2	-80.2	-79.1	19.1	19.3
2023/1/24	4.2	5.1	3.5	4	4.3	4.7	-20.7	-20.3	-80.2	-78.4	17.9	19.1
2023/1/25	4.2	5.1	3.5	4	4.2	4.7	-21.1	-20.3	-80.2	-78.4	17.7	17.9
2023/1/26	4.2	5.1	3.4	4	4.3	4.7	-21.1	-20.7	-80.1	-79.4	17.7	18.3
2023/1/27	4.2	5.2	3.5	4.1	4.2	4.7	-20.9	-20.5	-80.2	-78.9	17.7	18.1
2023/1/28	4.2	5.2	3.6	4	4.2	4.6	-20.9	-20.7	-80.2	-78.7	17.5	17.8
2023/1/29	4.2	5.2	3.5	4.1	4.2	4.6	-20.9	-20.6	-80.3	-79.4	17.7	19.1
2023/1/30	4.1	5.2	3.5	4.2	4.2	4.7	-21.1	-20.6	-80.3	-78.3	18.3	19.3
2023/1/31	4.1	5.2	3.2	4.4	3.8	5.1	-21.1	-20.5	-80.2	-79	18.6	19.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 201
藥學部 邱子荊 202

總藥師: 藥學部 廖志峰 201

業管主管: 藥學部 何沁心 202