

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112 年 02 月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362												
0850-0835												
2~8°C												
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/2/1	3.9	5.3	3.7	4.4	3.7	4.4	-20.4	-19.6	-80.1	-77.8	19.2	19.9
2023/2/2	4	5.1	3.8	4.7	3.7	4.3	-20.4	-19.6	-80.1	-78.4	19.4	20.1
2023/2/3	3.7	5.1	3.8	4.6	3.8	4.4	-20.1	-19.4	-79.9	-78.9	19.4	20.1
2023/2/4	3.7	5	3.7	4.3	3.8	4.3	-20.1	-19.8	-79.9	-78.7	19.6	19.7
2023/2/5	3.7	5	3.8	4.3	3.8	4.3	-20.1	-19.8	-79.9	-78.6	19.6	19.7
2023/2/6	3.8	5.1	3.6	4.4	3.8	4.4	-20.4	-19.8	-79.9	-78.4	19.5	20.3
2023/2/7	3.9	5.2	3.6	4.4	3.7	4.4	-20.4	-19.6	-80.1	-78.9	19.5	20.2
2023/2/8	3.8	5.6	3.7	4.7	3.7	4.3	-20.2	-19.6	-80.1	-78	19.6	20.2
2023/2/9	3.5	5.2	3.6	4.3	3.7	4.6	-19.9	-19.7	-79.9	-78	19.7	20.6
2023/2/10	3.4	5.1	3.6	4.5	3.7	4.4	-19.9	-19.6	-79.9	-78	19.7	20.3
2023/2/11	3.6	4.9	3.6	4.2	3.8	4.3	-19.9	-19.8	-79.9	-78.1	19.7	19.8
2023/2/12	3.6	4.9	3.6	4.2	3.8	4.3	-19.9	-19.8	-79.9	-78.2	19.6	19.7
2023/2/13	3.6	4.9	3.4	4.4	3.8	4.5	-20.9	-19.8	-79.9	-78.4	19.4	20.2
2023/2/14	3.5	4.9	3.6	4.4	3.8	4.3	-20.9	-20.4	-79.9	-78	19.4	19.8
2023/2/15	3.6	4.8	3.6	4.4	3.9	4.3	-20.9	-20.3	-79.9	-78.4	19.1	19.5
2023/2/16	3.6	5	3.7	4.3	3.9	4.3	-20.7	-20.3	-80.1	-78.9	19	19.6
2023/2/17	3.3	5.1	3.6	4.3	3.9	4.6	-20.9	-20.3	-79.9	-78.9	19.2	20
2023/2/18	3.6	4.8	3.7	4.4	3.8	4.4	-20.9	-20.6	-80.1	-78.6	19.6	20.4
2023/2/19	3.6	4.7	3.8	4.3	3.8	4.3	-20.9	-20.7	-79.9	-78.4	19.4	19.7
2023/2/20	3.6	5.2	3.9	4.4	3.8	4.3	-20.9	-20.7	-79.9	-78.4	19.2	19.7
2023/2/21	3.7	5.1	3.8	4.4	3.8	4.3	-21.1	-20.7	-79.9	-78.4	19.2	19.7
2023/2/22	3.8	4.9	3.7	4.5	3.7	4.4	-21.1	-20.1	-80.1	-77.9	19.2	20.9
2023/2/23	3.8	4.8	3.7	4.3	3.6	4.3	-20.7	-20.1	-81.4	-77.6	19.5	20.2
2023/2/24	3.9	4.8	3.6	4.4	3.6	4.9	-20.8	-20.4	-81.4	-79.9	19.4	20.1
2023/2/25	3.9	4.8	3.7	4.3	3.7	4.3	-20.8	-20.5	-81.4	-79.8	19	19.4
2023/2/26	3.9	4.8	3.7	4.2	3.7	4.3	-20.8	-20.4	-81.4	-79.5	18.7	19
2023/2/27	3.9	4.8	3.7	4.2	3.7	4.3	-20.8	-20.5	-81.2	-79.6	18.6	18.8
2023/2/28	3.9	4.9	3.7	4.2	3.7	4.3	-20.8	-20.4	-81.2	-79.8	18.8	18.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 112 0301
藥學部 邱子芹 112 0301

總藥師: 藥學部 廖志峰 112 0310

業管主管: 藥學部 何沁沁 112 0302