

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112 年 02 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/2/1	5.5	5.8	4.1	5	4.9	5.9	6	6.3	20.6	21.2
2023/2/2	5.3	5.8	4.1	5.1	4.9	6.7	5.9	6.2	21.2	21.5
2023/2/3	5.3	5.8	4.1	5.1	4.6	6.3	5.9	6.3	21.4	21.9
2023/2/4	5.4	5.8	4.1	5.1	4.6	5.3	5.9	6.2	21.5	21.9
2023/2/5	5.4	5.8	4.1	5.1	4.6	5.3	5.9	6.1	21.9	21.9
2023/2/6	5.4	5.8	4.1	5.2	4.6	5.7	5.9	6.1	21.7	22
2023/2/7	5.3	6.1	4.1	5.1	4.7	5.9	5.8	6.1	21.9	22.2
2023/2/8	5.3	5.7	4.2	5.2	4.7	6.1	5.8	6.2	21.9	22.2
2023/2/9	5.4	5.7	4.2	5.2	4.9	6	5.9	6.2	21.9	22.3
2023/2/10	5.1	6.6	4.2	5.2	4.9	5.9	5.9	6.2	22.1	22.4
2023/2/11	5.6	5.8	4.2	5.1	4.9	5.9	5.9	6.2	22.1	22.3
2023/2/12	5.6	5.8	4.2	5.1	4.9	5.9	5.9	6.2	21.9	22.1
2023/2/13	5.6	5.9	4.2	5.2	4.9	5.9	5.9	6.2	21.7	22.1
2023/2/14	5.7	5.9	4.2	5.3	4.9	6.2	5.9	6.2	21.6	22
2023/2/15	5.7	5.9	4.3	5.4	4.9	6	5.9	6.2	21.1	21.6
2023/2/16	5.4	5.9	4.3	5.2	5.1	5.9	6	6.2	21.1	21.2
2023/2/17	5.4	5.9	4.2	5.5	5.1	6.2	5.7	6.2	20.9	21.7
2023/2/18	5.7	5.9	4.2	5.1	5.1	5.9	5.7	6.6	21.6	21.9
2023/2/19	5.6	5.9	4.2	4.9	5.2	5.9	6.4	6.6	21.5	21.9
2023/2/20	5.6	5.9	4.2	5	5.1	5.9	6.2	6.4	21.1	21.5
2023/2/21	5.6	5.9	4.2	5.1	5.2	6.7	6.1	6.3	20.9	21.2
2023/2/22	5.6	6.2	4.2	5.1	4.9	6.1	6.2	6.9	20.9	21.5
2023/2/23	5.9	6.3	4.2	5.1	4.9	6	5.9	6.9	21.3	21.8
2023/2/24	5.9	6.2	4.2	5.3	4.5	6.2	5.8	6.3	21.6	21.8
2023/2/25	5.9	6.2	4.2	4.9	4.6	5.6	5.9	6.1	21.1	21.7
2023/2/26	5.9	6.2	4.2	4.9	4.6	5.6	5.9	6.1	20.4	21
2023/2/27	5.9	6.1	4.2	4.9	4.6	5.7	5.8	6.1	20.3	20.5
2023/2/28	5.9	6.2	4.2	4.9	4.6	5.7	5.9	6.1	20.5	20.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師 **藥學部 邱保祥** 112  
師(三)級藥師  
**藥學部 邱子芹** 112  
師(三)級藥師

總藥師: **藥學部 廖志峰** 112  
師(二)級藥師

業管主管: **藥學部 何沁沁** 112  
師(一)級藥師