

# 臺北榮民總醫院藥學部

## 臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表112年 02月

	六號冰箱		七號冰箱		室溫	
序號	1436-0086		2123-0026		1118-0011	
備用	0436-049		2123-0027		0742-0169	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2023/2/1	4.4	5.2	4.2	4.6	16.6	17.6
2023/2/2	4.5	5.3	4.2	4.7	17.4	17.7
2023/2/3	4.6	5.3	4.3	4.8	17.4	18.2
2023/2/4	4.7	5.3	4.4	4.8	17.9	18.6
2023/2/5	4.7	5.3	4.4	4.8	18.4	19.1
2023/2/6	4.4	5.3	4.4	4.8	18.4	19.1
2023/2/7	4.4	5.3	4.3	4.8	18.4	19.1
2023/2/8	4.2	5.3	4.3	4.7	18.6	18.9
2023/2/9	4.4	5.3	4.3	4.7	18.6	19.2
2023/2/10	4.4	5.3	4.2	4.7	18.8	19.3
2023/2/11	4.4	5.3	4.2	4.6	18.8	19.4
2023/2/12	4.4	5.2	4.2	4.6	18.8	19.4
2023/2/13	4.4	5.2	4.2	4.6	18.7	19
2023/2/14	4.4	5.3	4.2	4.8	17.9	18.9
2023/2/15	4.4	5.2	4.1	4.6	17.3	18.1
2023/2/16	4.4	5.3	4.2	4.5	17.3	17.7
2023/2/17	4.6	5.3	4.2	4.7	17.6	18.8
2023/2/18	4.7	5.3	4.3	4.7	18.7	18.8
2023/2/19	4.7	5.3	4.2	4.7	18.4	18.9
2023/2/20	4.5	5.3	4.2	4.8	18.2	18.4
2023/2/21	4.5	5.3	4.4	5	18.2	18.4
2023/2/22	4.5	5.3	4.4	4.8	18.2	18.7
2023/2/23	4.5	5.2	4.4	4.7	18.6	18.9
2023/2/24	4.5	5.2	4.4	4.8	18.6	19.1
2023/2/25	4.5	5.3	4.5	4.7	17.9	18.6
2023/2/26	4.5	5.3	4.5	4.7	17.4	17.9
2023/2/27	4.5	5.3	4.4	4.7	17.8	18.2
2023/2/28	4.5	5.3	4.4	4.8	18	18.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 112 301  
藥學部 邱子芹 112 100

總藥師: 藥學部 廖志峰 112 310

業管主管: 藥學部 何沁沁 112 1200

溫度均在正常範圍內