

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112年 03月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/3/1	3.9	4.9	3.3	4.5	3.6	4.3	-20.7	-19.1	-81.2	-79.8	19.2	19.6
2023/3/2	4	5.1	3.3	4.2	3.7	4.3	-20.9	-20.3	-81.4	-79.5	19	19.7
2023/3/3	4	4.8	3.3	4.2	3.6	4.4	-21.2	-20.6	-81.4	-79.7	19.1	19.7
2023/3/4	4.1	4.8	3.4	4	3.7	4.3	-21.3	-20.9	-81.4	-80	19.2	19.3
2023/3/5	4.1	4.8	3.4	4	3.7	4.3	-21.3	-20.9	-81.4	-79.6	18.8	19.2
2023/3/6	4	4.8	3.4	4.3	3.7	4.4	-21.3	-20.9	-81.4	-79.5	18.8	19.5
2023/3/7	3.8	4.8	3.4	4.3	3.6	4.4	-21.4	-20.9	-81.3	-79.6	18.9	19.6
2023/3/8	3.8	4.7	3.3	4.1	3.6	4.3	-21.4	-21.1	-81.4	-79.9	19.2	19.8
2023/3/9	3.7	6.8	3.3	4.3	3.3	6.3	-21.4	-21.1	-81.4	-79	19.4	19.9
2023/3/10	3.7	4.7	3.3	4	3.7	4.4	-21.3	-19.3	-81.3	-80	19.4	20.2
2023/3/11	4.2	4.8	3.3	3.9	3.7	4.4	-19.7	-19.5	-81.4	-80.1	19.6	19.7
2023/3/12	4.2	4.8	3.3	3.9	3.7	4.4	-19.7	-19.5	-81.3	-80.1	19.5	19.7
2023/3/13	4	5.7	3.3	3.9	3.7	4.4	-20.8	-19.6	-81.4	-79.9	19.3	19.9
2023/3/14	4.1	5	3.3	4.1	3.6	4.3	-20.8	-20.2	-81.4	-79.7	19.3	20.2
2023/3/15	4.2	4.9	3.3	4.7	3.6	4.2	-20.6	-20	-81.3	-80.1	19.6	20.3
2023/3/16	4.1	5.1	3.1	4.7	3.4	4.2	-20.6	-19.9	-81.4	-80.1	19.7	21.1
2023/3/17	4.1	5.3	3.3	4.1	3.3	4.2	-21.1	-20.3	-81.4	-80.1	19.8	20.7
2023/3/18	4.2	4.9	3.4	4.1	3.4	4.1	-21.1	-20.8	-81.3	-80	19.8	20.1
2023/3/19	4.2	4.9	3.4	4.1	3.4	4.1	-21.1	-20.8	-81.2	-80	19.7	19.8
2023/3/20	4.1	5.1	3.3	4.4	3.4	4.2	-21.1	-20.4	-81.2	-79.9	19.5	20.2
2023/3/21	3.9	5.1	3.3	4	3.4	4.1	-20.7	-20	-81.2	-79.8	19.5	20.3
2023/3/22	3.9	4.7	3.3	4	3.5	4.1	-21.2	-20.2	-81.2	-80	19.8	20.4
2023/3/23	4	4.7	3.3	4	3.4	5.1	-21.4	-20.9	-81.3	-80	19.9	20.4
2023/3/24	3.6	4.6	3.2	4	3.7	4.3	-21.7	-20.9	-81.3	-80.1	19.9	20.6
2023/3/25	3.7	4.5	3.1	4.2	3.7	4.3	-21.7	-21.2	-81.2	-78.4	19.9	20.8
2023/3/26	3.7	4.5	3.1	3.9	3.9	4.3	-21.7	-21.4	-81.2	-79.9	19.7	20.1
2023/3/27	3.7	4.7	3.1	4.1	3.9	4.3	-21.7	-20.9	-81.2	-79.9	19.6	20.2
2023/3/28	3.8	4.7	2.9	4	3.9	4.4	-21.2	-20.7	-81.4	-79.9	19.7	20.2
2023/3/29	3.7	4.8	3.2	4.3	3.8	4.3	-20.9	-20.4	-81.5	-79.7	19.8	20.3
2023/3/30	3.7	5.2	3.5	4.1	3.8	4.3	-20.9	-20.5	-81.5	-80	19.8	20.4
2023/3/31	3.4	4.9	3.4	4.2	3.4	4.7	-20.9	-20.4	-81.5	-79.2	19.9	20.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 0406
藥學部 邱子芹 0406

總藥師: 藥學部 廖志峰 0457

業管主管: 藥學部 何沁沁 0407