

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112 年 04 月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362												
0850-0835												
2~8°C	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/4/1	3.7	4.5	3.4	3.9	3.8	4.2	-20.8	-20.5	-81.5	-79.9	19.9	19.9
2023/4/2	3.7	4.5	3.4	3.9	3.8	4.2	-20.9	-20.5	-81.5	-79.9	19.7	19.9
2023/4/3	3.7	4.5	3.4	3.9	3.8	4.2	-20.9	-20.6	-81.5	-79.9	19.6	19.7
2023/4/4	3.7	4.5	3.4	3.9	3.8	4.2	-20.9	-20.5	-81.5	-79	19.6	19.7
2023/4/5	3.7	4.5	3.4	3.9	3.8	4.2	-20.8	-20.4	-81.5	-78.9	19.7	19.8
2023/4/6	3.4	4.5	2.8	3.9	3.7	4.3	-20.7	-18.8	-81.5	-79	19.6	21.5
2023/4/7	3.9	4.7	2.9	4.4	3.8	4.4	-20.8	-20	-81.5	-79.5	18	19.6
2023/4/8	4.1	4.8	3.7	4.3	3.9	4.3	-20.7	-20.3	-81.5	-79.3	17.1	18.1
2023/4/9	4.2	4.8	3.7	4.3	3.9	4.4	-20.7	-20.4	-81.5	-79.9	16.7	17.3
2023/4/10	4.2	4.9	3.7	4.3	3.9	4.7	-21.2	-20.3	-81.5	-78.9	16.9	18.3
2023/4/11	4	4.9	3.6	4.3	3.9	4.4	-21.2	-20.5	-81.3	-79.5	17.3	18.3
2023/4/12	3.8	4.8	3.6	4.2	3.7	4.4	-21.1	-20.2	-81.4	-79.7	17.7	18.7
2023/4/13	4	4.8	3.6	4.3	3.9	4.3	-22	-20.8	-81.4	-79.8	17.9	19.1
2023/4/14	3.8	4.8	3.7	4.3	3.9	4.4	-22.2	-21.7	-81.4	-78.6	17.5	18.3
2023/4/15	3.9	4.8	3.7	4.2	3.9	4.3	-22.4	-21.9	-81.2	-79.9	17.5	18.2
2023/4/16	3.9	4.8	3.7	4.2	3.9	4.3	-22.4	-21.9	-81.2	-80.1	17.4	18.1
2023/4/17	3.9	4.9	3.7	4.2	3.9	4.4	-22.3	-21.9	-81.2	-79.8	17.6	18.8
2023/4/18	3.9	4.8	3.5	4.2	4	4.4	-22.4	-21.9	-81.2	-79.8	17.7	18.7
2023/4/19	4	5.4	3.4	4.4	4	4.5	-22.4	-21.9	-81.2	-75.4	17.9	18.8
2023/4/20	4.1	5.1	3.7	4.4	3.9	4.7	-22.4	-21.6	-81.1	-79.8	17.9	18.3
2023/4/21	4.3	5.2	3.7	4.3	4	4.6	-22.2	-21.7	-81.1	-79.9	17.7	18.3
2023/4/22	4.4	5.2	3.8	4.3	4.1	4.4	-22.3	-21.9	-81.2	-79.9	17.1	18.2
2023/4/23	4.3	5.2	3.8	4.3	4.2	4.4	-22.4	-21.9	-81.2	-79.6	16.8	18
2023/4/24	4.4	5.3	3.8	4.4	4.1	4.5	-22.4	-21.9	-81.2	-79.6	16.4	17.5
2023/4/25	4.3	5.3	3.7	4.4	4	4.4	-22.5	-21.9	-81.2	-79.7	17.1	20.8
2023/4/26	4.3	5.1	3.8	4.4	4	4.4	-22.5	-20.6	-81.4	-79	20.2	20.8
2023/4/27	4.3	5.1	3.8	4.3	3.9	4.4	-20.9	-19.6	-81.5	-79.1	19.9	21.6
2023/4/28	4.4	5.2	3.8	4.4	3.9	4.4	-20.8	-20	-81.4	-79.9	19.6	20.2
2023/4/29	4.4	5.1	3.8	4.3	4	4.4	-20.8	-20.6	-81.5	-79.9	19.3	19.6
2023/4/30	4.4	5.1	3.8	4.3	4	4.4	-20.8	-20.6	-81.4	-79.9	19.2	19.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 112 0501 師(三)級藥師 藥學部 邱子堯 112 0502 師(三)級藥師

總藥師: 藥學部 廖志峰 112 0503 師(二)級藥師

業管主管: 藥學部 何沁沁 112 0503 師(一)級藥師