

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112年 05月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362												
0850-0835												
備用												
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/5/1	4.2	5.1	3.7	4.3	4.1	4.4	-21.2	-19.6	-81.5	-79.2	19.3	19.8
2023/5/2	3.7	5	3.7	4.4	4.1	4.5	-21.4	-20.8	-81.5	-78.9	19.2	20.3
2023/5/3	3.7	4.9	3.7	4.3	4	4.4	-21.4	-20.9	-80.4	-78.2	19.3	20.1
2023/5/4	3.4	4.7	3.7	4.3	4.1	4.4	-21.4	-20.7	-80.4	-77.8	19.1	19.6
2023/5/5	3.7	4.8	3.6	4.4	4.1	4.6	-21.4	-20.1	-80.3	-78	19.1	19.7
2023/5/6	3.9	4.8	3.6	4.3	4.1	4.4	-20.8	-20.5	-80.3	-77.8	19	19.2
2023/5/7	3.9	4.8	3.6	4.3	4.1	4.4	-20.9	-20.5	-80.4	-78.1	18.8	19
2023/5/8	3.9	4.7	3.5	4.3	3.9	4.6	-20.9	-20.5	-80.4	-78.1	18.8	19.3
2023/5/9	3.9	6.1	3.6	4.5	4	4.5	-21.4	-20.7	-80.4	-77.8	18.8	19.3
2023/5/10	4	4.9	3.6	4.4	4	4.4	-21.9	-20.9	-80.5	-77.9	18.7	19.4
2023/5/11	4.1	5.6	3.6	4.9	4	4.5	-21.9	-20.6	-80.4	-78.8	18.8	19.4
2023/5/12	4.1	5.1	3.7	4.3	4	4.8	-21.2	-19.8	-80.3	-78.9	18.8	19.4
2023/5/13	4.1	4.9	3.7	4.3	4.2	4.6	-21.1	-20.9	-80.4	-78.9	18.8	19
2023/5/14	4.1	4.9	3.7	4.3	4.2	4.6	-21.1	-20.8	-80.3	-78.9	18.7	18.8
2023/5/15	4.1	6.2	3.7	4.6	3.6	4.6	-21.4	-20.8	-80.4	-77.9	18.6	19.3
2023/5/16	3.9	5.1	3.7	4.3	4	4.4	-21.7	-21.1	-80.4	-77.9	18.7	19.3
2023/5/17	3.9	4.8	3.7	4.3	4	4.6	-21.7	-20.9	-80.4	-78.9	18.8	19.4
2023/5/18	3.8	4.8	3.7	4.8	4	4.5	-21.5	-20.9	-80.4	-77.7	18.8	19.7
2023/5/19	3.9	4.9	3.5	4.5	4	5.3	-21.7	-21.2	-80.4	-77.4	19	19.8
2023/5/20	4.2	5.1	3.7	4.4	4.2	4.5	-21.7	-21.5	-80.4	-77.8	19	19.2
2023/5/21	4.2	5.1	3.7	4.4	4.2	4.5	-21.7	-21.5	-80.4	-77.4	18.8	19
2023/5/22	4.2	5.1	3.7	4.4	4.2	4.5	-21.9	-21.1	-80.4	-77.4	18.7	19.4
2023/5/23	4	5.2	3.7	4.4	3.9	4.5	-22.2	-21.1	-80.4	-78.9	18.9	19.5
2023/5/24	4.2	5.1	3.4	4.6	4.1	4.9	-22.4	-21.3	-80.4	-77.2	18.9	19.7
2023/5/25	4	5.2	3.3	4.6	4.6	4.9	-22.5	-22.2	-80.4	-78	19	19.7
2023/5/26	4.2	5.1	3.2	4.3	4.6	4.9	-22.5	-22.2	-80.3	-78.3	18.9	19.6
2023/5/27	4.2	5	3.3	4.2	4.6	4.8	-22.6	-22.3	-80.4	-78.2	18.7	18.9
2023/5/28	4.3	5	3.3	4.2	4.6	4.9	-22.6	-22.3	-80.3	-78	18.6	18.7
2023/5/29	4.3	5.1	3.2	4.2	4.2	5.5	-22.6	-21.9	-80.4	-78.3	18.6	19.1
2023/5/30	4.2	4.9	3.3	4.4	4.3	4.7	-22.2	-20.4	-80.4	-77.9	18.7	19.4
2023/5/31	3.9	5.1	3.4	4.4	3.9	4.9	-20.8	-20.2	-80.4	-77.8	18.8	19.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥 112 0601
 藥學部 邱子芹 112 0601

總藥師

藥學部 廖志峰 112 0601

業管主管:

藥學部 周千澄 112 0606
 科主任