

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年11月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-293		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-292		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2018/11/1	3.6	4.4	4.3	5.3	4.6	5.1	-21.8	-18.9	-81.8	-79.8	19.1	19.9
2018/11/2	3.5	4.4	4.3	5.3	4.5	5.1	-21.9	-20.8	-81.7	-79.5	18.8	19.8
2018/11/3	4	4.4	4.6	5.3	4.7	5.1	-21.9	-21.6	-81.8	-79.8	18.7	19.1
2018/11/4	4	4.4	4.6	5.3	4.7	5.1	-21.9	-21.6	-81.8	-79.9	18.6	19.3
2018/11/5	3.6	4.5	4.1	5.4	4.5	5.1	-22	-18.7	-81.8	-79.8	18.4	20.3
2018/11/6	3.5	4.7	4	5.5	4.5	5.3	-21.6	-19.6	-81.8	-79.4	19.2	20.2
2018/11/7	3.3	4.6	4.3	5.4	4.6	5.1	-21.2	-20.4	-81.8	-79.8	19.3	20.3
2018/11/8	3.4	4.7	4.1	5.4	4.4	5.2	-21.4	-20.6	-81.9	-79.6	19.3	20.9
2018/11/9	3.7	4.7	4.1	5.3	4.4	5.2	-21.4	-20.4	-81.8	-79.6	19.4	20.7
2018/11/10	4.3	4.5	4.4	5.4	4.7	5.2	-21.5	-21.2	-81.8	-79.4	19.3	19.7
2018/11/11	4.3	4.6	4.4	5.4	4.7	5.2	-21.5	-21.2	-81.9	-79.7	19.2	19.7
2018/11/12	3.7	4.6	4.1	5.4	4.3	5.1	-21.5	-20.3	-81.9	-79.6	19.2	20.6
2018/11/13	3.7	4.8	4.1	5.4	4.4	5.3	-21.1	-20.4	-81.8	-79.6	19.3	20.4
2018/11/14	3.8	4.8	4.1	5.4	4.5	5.3	-20.9	-18.4	-81.8	-79.7	19.5	20.9
2018/11/15	3.8	5.2	4.1	5.4	4.6	5.5	-19.9	-18.6	-81.8	-79.5	19.5	20.4
2018/11/16	4.1	5.2	4.1	5.4	4.7	5.5	-20.4	-19.8	-81.8	-78.9	19.4	20.6
2018/11/17	4.9	5.2	4.4	5.4	5.2	5.5	-20.6	-20.2	-81.8	-79.6	19.4	19.6
2018/11/18	4.9	5.2	4.3	5.4	5.1	5.5	-20.6	-20.2	-81.8	-79.6	19.3	19.4
2018/11/19	4	5.2	4.2	5.4	4.7	5.6	-20.6	-19.7	-81.8	-79.8	19.2	20.8
2018/11/20	4.1	5.3	4.1	5.4	4.7	5.6	-20.8	-20.1	-81.7	-79.6	19.8	20.9
2018/11/21	4.2	5.3	4.1	5.4	4.7	5.6	-20.9	-20.4	-81.8	-79.5	20.1	21.3
2018/11/22	4.1	5.4	3.9	5.4	4.7	5.5	-21.1	-20.4	-81.7	-79.7	20.2	21.3
2018/11/23	4	5.3	4	5.4	4.7	5.5	-21.2	-20.6	-81.7	-79.8	19.8	20.8
2018/11/24	5.1	5.2	4.4	5.5	5.1	5.5	-21.2	-20.9	-81.8	-79.8	19.7	19.9
2018/11/25	4.9	5.3	4.4	5.5	5.1	5.6	-21.2	-20.7	-81.7	-79.6	19.4	19.7
2018/11/26	3.9	5.4	4.1	5.5	4.7	5.5	-21.2	-20.7	-81.7	-79.7	19.4	21.2
2018/11/27	3.9	5.1	4.1	5.5	4.7	5.4	-21.2	-20.7	-81.7	-79.7	19.6	20.8
2018/11/28	4	5.4	4.1	5.5	4.4	5.5	-21.2	-20.2	-81.5	-79.7	19.7	20.9
2018/11/29	3.9	5.5	4.1	5.5	4.4	5.6	-23.2	-20.4	-81.7	-76.9	19.6	20.8
2018/11/30	3.9	5.2	4.2	5.5	4.3	5.1	-22.9	-21.5	-81.7	-79.6	19.6	21.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 107
師(三)級藥師 (203)
藥學部 溫雅如 107