

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年10月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-293		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-292		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2018/10/1	4.1	5	5.1	6.1	4.3	4.9	-20.9	-19.7	-81.5	-79.7	20.1	20.9
2018/10/2	3.9	5.1	4.7	5.8	4.2	4.9	-21	-20.6	-81.7	-79.5	20.1	21.7
2018/10/3	3.6	4.8	4.7	5.8	4.3	4.9	-21.2	-20.7	-81.6	-79.5	20.3	21.6
2018/10/4	3.5	4.7	4.7	5.8	4.3	4.9	-21.2	-20.7	-81.5	-79.7	19.9	21.7
2018/10/5	3.4	4.6	4.7	5.7	4.3	4.9	-21.2	-20.9	-81.5	-79.8	19.9	21.1
2018/10/6	4.3	4.7	4.7	5.4	4.6	4.9	-21.4	-20.9	-81.7	-79.3	19.1	20.2
2018/10/7	4.3	4.6	4.7	5.4	4.5	4.9	-21.4	-21	-81.5	-79.8	18.8	19.1
2018/10/8	3.1	4.6	4.2	6.2	4.3	4.9	-21.4	-20.8	-81.5	-79.6	18.8	19.8
2018/10/9	3.1	4.5	4.1	6.1	4.3	5.2	-21.4	-20.8	-81.6	-79.3	19	21.4
2018/10/10	4.2	4.5	4.1	4.8	4.7	4.9	-21.5	-21.2	-81.5	-79.8	20.2	20.4
2018/10/11	3.3	4.6	4.1	5.7	4.6	5.3	-21.6	-20.9	-81.5	-79.7	20.1	21.4
2018/10/12	3.3	4.7	4.4	5.8	4.7	5.3	-21.7	-20.8	-81.5	-79.5	20.3	21.4
2018/10/13	4	4.4	4.4	5.2	4.9	5.3	-21.4	-21.1	-81.5	-79.3	20.3	20.6
2018/10/14	4	4.4	4.4	5.2	4.9	5.3	-21.4	-21.1	-81.7	-79.4	20.2	20.3
2018/10/15	3.1	4.5	4.4	5.8	4.7	5.3	-22.3	-20.6	-81.5	-79.8	20.1	21.8
2018/10/16	3.2	4.4	4.5	5.8	4.6	5.3	-22.4	-21.4	-81.6	-79.2	20.3	21.5
2018/10/17	3.3	4.4	4.6	5.7	4.6	5.2	-22.5	-21.9	-81.7	-79.6	19.6	21.4
2018/10/18	3.4	4.5	4.6	5.9	4.4	5.2	-22.5	-21.9	-81.7	-79.7	19.1	21.3
2018/10/19	3.4	4.5	4.6	5.7	4.3	5.1	-22.5	-21.7	-81.7	-79.8	18.8	21
2018/10/20	4.2	4.5	4.6	5.2	4.7	5.1	-22.1	-21.7	-81.6	-79.7	19.3	20.3
2018/10/21	4.2	4.5	4.6	5.2	4.7	5.1	-22.2	-21.7	-81.7	-79.8	19.1	19.3
2018/10/22	3.4	4.6	4.6	5.5	4.4	5.7	-22.2	-21.4	-81.7	-77.5	19	20.2
2018/10/23	3.5	4.7	4.6	5.5	4.4	5.2	-22.2	-21.4	-81.8	-79.8	19.2	20.7
2018/10/24	3.4	4.7	4.6	5.5	4.4	5.2	-22.2	-21.6	-81.7	-79.8	19.3	20.6
2018/10/25	3.4	4.5	4.6	5.6	4.2	5.1	-22.4	-21.5	-81.7	-79.8	19.4	20.5
2018/10/26	3.7	4.5	4.6	5.6	4.2	5.8	-22.4	-21.3	-81.7	-79.5	19.4	20.4
2018/10/27	3.4	4.4	4.6	5.2	4.5	4.9	-22.4	-21.7	-81.8	-79.5	19.2	19.6
2018/10/28	3.5	4.5	4.6	5.2	4.5	4.9	-22	-21.7	-81.7	-79.8	18.9	19.1
2018/10/29	3.4	4.4	4.6	5.6	4.2	4.9	-22	-20.9	-81.8	-79.4	18.8	20.2
2018/10/30	3.5	4.5	4.4	5.4	4.4	5.1	-21.4	-20.8	-81.8	-76.2	19.6	21.1
2018/10/31	3.6	4.4	4.3	5.3	4.4	5.1	-21.4	-19.9	-81.8	-79.6	19.3	20.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 溫度均在正常範圍內  
藥學部 邱保祥 (10)  
藥學部 溫雅如 (10)  
藥學部 溫雅如 (10)