

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年10月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2018/10/1	4.4	5.1	5.6	6.1	21.3	22.2
2018/10/2	4.3	5.8	5.7	6.1	21.9	22.6
2018/10/3	4.3	6.4	5.7	6.2	22.2	23
2018/10/4	4.3	5.1	5.7	6.2	22.1	22.9
2018/10/5	4.3	4.9	5.7	6.1	22	22.6
2018/10/6	4.3	4.9	5.6	6	21.4	22.4
2018/10/7	4.3	4.9	5.6	6	20.8	21.4
2018/10/8	4.3	5.2	5.4	6	20.6	21.3
2018/10/9	4.4	5.3	5.2	6.2	21.1	22.4
2018/10/10	4.4	5.1	5.2	5.7	21.9	22.3
2018/10/11	4.4	5.1	5.2	6.6	21.7	22.4
2018/10/12	4.4	5.1	5.4	5.9	22	22.6
2018/10/13	4.4	5	5.4	5.9	22	22.4
2018/10/14	4.4	5.1	5.3	5.8	21.7	22.1
2018/10/15	4.4	5.2	5.3	6.1	21.6	22.3
2018/10/16	4.4	5.2	5.6	6.2	21.9	22.8
2018/10/17	4.4	5.2	5.8	6.1	22	22.6
2018/10/18	4.4	5.7	5.6	6.3	21.3	22
2018/10/19	4.5	5.7	5.4	5.9	21.1	21.9
2018/10/20	4.5	5.1	5.3	5.8	21.4	21.9
2018/10/21	4.4	5.1	5.3	5.8	21.1	21.4
2018/10/22	4.4	5.1	4.8	5.6	20.8	21.7
2018/10/23	4.4	5.2	4.7	5.3	21.2	21.8
2018/10/24	4.4	5.2	4.5	5.2	21.4	22.1
2018/10/25	4.4	5.1	4.7	5.3	21.6	22.1
2018/10/26	4.4	5.2	4.3	5.1	21.6	22.2
2018/10/27	4.5	5.1	4.4	4.9	21.4	22.1
2018/10/28	4.4	5.1	4.4	4.9	21.1	21.4
2018/10/29	4.4	5.2	4.3	5.1	20.7	21.7
2018/10/30	4.5	5.3	4.4	5.1	21.4	22.1
2018/10/31	4.5	5.2	4.3	5.1	21.9	22.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

- 溫度均在正常範圍內

承辦藥師: 藥學部 邱保祥 (01)  
師(三)級藥師 (10)  
藥學部 溫雅如 (10)  
師(三)級藥師 (10)