

臺北榮民總醫院藥學部
臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年07月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-292		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2018/7/1	4.4	4.7	4.9	5.7	4.4	4.8	-19.9	-19.7	-81.7	-79.9	19.9	20.1
2018/7/2	3.9	4.7	4.2	5.7	4.3	4.9	-20.4	-19.2	-81.7	-79.9	20.2	21.3
2018/7/3	3.8	5.1	4.2	5.7	4.2	4.9	-20.7	-19.9	-81.7	-79.9	19.9	21.4
2018/7/4	4	5.2	4.2	5.7	4.2	4.8	-20.8	-20.3	-81.7	-79.9	20.2	21.6
2018/7/5	4	4.9	4.2	5.7	4.3	4.8	-20.8	-19.6	-81.6	-79.9	20.4	22.1
2018/7/6	3.9	4.7	4.3	5.5	4.3	4.7	-20.7	-19.9	-81.6	-79.9	20.1	21.2
2018/7/7	4.3	4.7	4.7	5.5	4.3	4.7	-20.7	-20.2	-81.7	-79.7	19.7	20.1
2018/7/8	4.3	4.7	4.7	5.5	4.3	4.7	-20.7	-20.3	-81.7	-79.9	19.5	19.7
2018/7/9	3.7	4.8	4.3	5.6	4.2	4.7	-20.8	-20.4	-81.7	-79.8	19.4	20.8
2018/7/10	3.8	4.7	4.4	5.7	4.3	4.9	-21.4	-19.6	-81.7	-79.9	19.6	20.8
2018/7/11	3.8	4.7	4.2	5.7	4.4	5.2	-21.8	-20.6	-81.7	-79.9	19.7	21
2018/7/12	4	4.7	4.2	5.7	4.5	5.1	-22.1	-20.9	-81.7	-77.3	19.7	20.9
2018/7/13	4	5	4.2	5.5	4.3	5	-22.2	-20.9	-81.7	-79.6	19.6	20.8
2018/7/14	4.4	4.8	4.6	5.5	4.3	4.7	-21.4	-20.9	-81.7	-79.8	19.5	19.9
2018/7/15	4.4	4.8	4.7	5.5	4.3	4.7	-21.5	-20.9	-81.6	-79.7	19.4	19.5
2018/7/16	4.1	4.9	4.2	5.6	4.1	4.8	-21.4	-20.9	-81.8	-79.6	19.3	20.8
2018/7/17	4.2	4.9	4.1	5.5	4.4	5.3	-21.9	-20.9	-81.7	-76.9	19.6	21.2
2018/7/18	4.2	5.1	4.1	5.5	4.4	5.4	-22.2	-21.1	-81.8	-79.9	19.7	20.9
2018/7/19	4.2	4.9	4	5.5	4.7	5.7	-22.2	-21.2	-81.8	-79.9	19.7	21
2018/7/20	4.3	5.1	3.8	5.5	4.7	5.4	-21.7	-21	-81.8	-79.8	19.7	21.3
2018/7/21	4.8	5.1	4.4	5.4	5.3	5.5	-21.7	-21.2	-81.8	-79.9	19.6	20.1
2018/7/22	4.7	5.1	4.4	5.3	5.3	5.4	-21.7	-21.3	-81.8	-79.9	19.4	19.7
2018/7/23	4.3	5.2	3.9	5.4	4.5	5.4	-22.2	-20.9	-81.9	-79.9	19.3	20.8
2018/7/24	4.2	5.3	3.8	5.4	4.4	5.1	-22.1	-20.7	-81.8	-79.9	19.4	20.8
2018/7/25	4.3	5.3	3.8	5.4	4.1	5.2	-22.1	-21	-81.8	-77.1	19.1	20.6
2018/7/26	4.3	5.4	3.9	5.4	4.1	4.9	-21.4	-20.9	-81.6	-79.8	19.4	20.6
2018/7/27	4.3	5.3	3.8	5.4	4.3	4.9	-21.7	-19.8	-81.7	-79.7	19.4	20.7
2018/7/28	5.1	5.4	4.6	5.5	4.5	4.9	-21.9	-20.2	-81.5	-79.6	19.4	19.8
2018/7/29	5.1	5.3	4.6	5.5	4.5	4.9	-22	-21.4	-81.7	-79.8	19.2	19.4
2018/7/30	4.3	5.4	3.8	5.4	4.3	4.9	-22.1	-21.2	-81.6	-79.6	19.2	20.7
2018/7/31	4.2	5.5	3.8	5.5	4.4	4.9	-22.4	-20.8	-81.5	-79.4	19.3	20.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師 : **藥學部邱保祥** 107
 職稱 : **(三級)藥師** 10801

藥學部溫雅如 107
 職稱 : **(三級)藥師** 10801