

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年06月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-292		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2018/6/1	3.4	4.5	3.9	5.3	4.3	4.9	-20.3	-19.7	-81.5	-79.9	20	21.2
2018/6/2	4.2	4.5	4.3	5.3	4.6	4.9	-19.9	-19.9	-81.7	-79.9	19.8	20.1
2018/6/3	4.2	4.5	4.3	5.3	4.6	4.9	-20.1	-19.8	-81.8	-79.9	19.7	20.1
2018/6/4	3.5	4.7	3.9	5.3	4.6	5.3	-20.4	-19.9	-81.7	-79.9	19.6	20.9
2018/6/5	3.3	4.7	3.9	5.3	4.8	5.6	-20.6	-20.1	-81.7	-79.9	19.8	21.1
2018/6/6	3.4	4.6	4.2	5.4	4.7	5.6	-20.6	-19.3	-81.8	-79.7	19.9	21.2
2018/6/7	3.4	4.5	4.2	5.5	4.8	5.7	-20.6	-19.9	-81.7	-79.9	20.1	21.4
2018/6/8	3.4	4.8	4.2	5.8	4.8	5.6	-20.2	-19.7	-81.7	-79.7	20.1	21.4
2018/6/9	4.4	4.8	4.6	5.4	5.2	5.6	-20.4	-19.9	-81.8	-79.7	18.6	20.3
2018/6/10	4.5	4.8	4.6	5.4	5	5.4	-20.4	-20.2	-81.8	-79.9	18.1	18.6
2018/6/11	3.9	5	4.3	5.5	4.7	5.7	-20.4	-19.2	-81.8	-79.9	18.1	20.3
2018/6/12	3.8	5	4.3	5.5	4.7	5.4	-19.9	-19.7	-81.7	-79.9	19.2	20.4
2018/6/13	3.9	4.8	4.4	5.7	4.8	5.4	-19.9	-19.5	-81.7	-79.9	19.3	20.6
2018/6/14	3.9	5.3	4.2	5.7	4.9	5.5	-20.1	-19.2	-81.7	-79.6	19.4	20.6
2018/6/15	4.3	5.3	4.4	6	4.9	5.6	-20.4	-19.5	-81.7	-79.9	19.4	21.6
2018/6/16	4.7	5.1	4.9	5.7	5.1	5.4	-20.8	-20.3	-81.8	-79.9	19.8	20.5
2018/6/17	4.7	5.1	4.9	5.7	5.1	5.4	-20.7	-20.4	-81.8	-79.8	19.6	19.8
2018/6/18	4.7	5.1	4.9	5.7	5.1	5.4	-20.7	-20.5	-81.7	-80	19.4	19.6
2018/6/19	4	5.3	4.4	5.7	4.9	5.4	-20.7	-19.8	-81.8	-79.8	19.4	20.8
2018/6/20	3.7	5	4.3	5.7	4.8	5.3	-20.4	-19.9	-81.7	-79.8	19.1	21.2
2018/6/21	3.8	5.1	4.3	5.7	4.9	5.3	-20.7	-20.2	-81.7	-79.7	20.3	21.4
2018/6/22	3.8	4.7	4.3	5.7	4.8	5.3	-20.7	-20.1	-81.7	-79.6	20.5	22.6
2018/6/23	4.4	4.7	4.8	5.7	4.9	5.3	-20.4	-20.2	-81.7	-79.7	22.3	22.6
2018/6/24	4.4	4.7	4.9	5.7	4.9	5.3	-20.4	-20.4	-81.6	-79.7	21.9	22.4
2018/6/25	3.7	4.7	4.3	5.7	4.3	5.3	-20.6	-20.2	-81.6	-79.8	20.3	21.9
2018/6/26	3.7	4.7	4.2	5.7	4.3	5.1	-20.9	-19.8	-81.7	-79.6	20.1	21.3
2018/6/27	3.8	4.8	4.2	5.7	4.4	5.1	-21.1	-19.4	-81.7	-79.7	20.1	21.3
2018/6/28	3.7	4.7	4.2	5.7	4.3	4.9	-20.9	-20.4	-81.7	-79.9	20.2	21.3
2018/6/29	3.8	4.8	4.3	5.9	4.2	5	-20.9	-20.1	-81.7	-79.9	20.2	21.4
2018/6/30	4.3	4.7	4.9	5.6	4.4	4.8	-20.4	-20.2	-81.7	-79.7	20.1	20.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 2102
藥學部 溫雅如 2102