

# 臺北榮民總醫院藥學部

## 臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 107年06月

	六號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2018/6/1	3.9	4.6	21.6	21.9
2018/6/2	3.9	4.2	21.4	21.7
2018/6/3	3.9	4.2	21.1	21.5
2018/6/4	3.9	4.2	21.1	21.4
2018/6/5	3.9	4.3	21.2	21.5
2018/6/6	3.8	4.1	21.3	21.7
2018/6/7	3.9	4.8	21.6	21.9
2018/6/8	4.3	4.4	21.4	21.9
2018/6/9	4.3	4.4	21.2	21.6
2018/6/10	4.3	4.4	20.9	21.3
2018/6/11	4.3	4.4	20.9	21.8
2018/6/12	4.3	4.6	20.9	21.4
2018/6/13	4.2	4.5	21.1	21.8
2018/6/14	4.3	4.4	21.2	21.6
2018/6/15	4.3	4.4	21.3	21.7
2018/6/16	4.3	4.4	21.2	21.4
2018/6/17	4.3	4.4	20.8	21.1
2018/6/18	4.3	4.4	20.8	20.9
2018/6/19	4.2	4.5	20.8	21.2
2018/6/20	4.2	4.3	21.1	21.3
2018/6/21	4.2	4.4	21.1	21.7
2018/6/22	4.3	4.4	21.3	21.6
2018/6/23	4.3	4.4	21.1	21.5
2018/6/24	4.2	4.4	20.9	21.3
2018/6/25	4.3	4.4	20.9	21.4
2018/6/26	4.2	4.5	21.3	22.6
2018/6/27	4.2	4.4	21.3	21.8
2018/6/28	4.2	6.1	21.4	22.2
2018/6/29	4.5	4.7	21.7	21.9
2018/6/30	4.4	4.7	21.3	21.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥 107  
師(三)級藥師 01702  
藥學部 溫雅如 107  
師(一)級藥師 01702