

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年05月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362												
0850-0835												
2~8°C												
2~8°C												
2~8°C												
負15~負25°C												
≤負70°C												
15~25°C												
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2018/5/1	4.4	5.1	3.9	5.5	4.3	5.7	-21.1	-19.2	-81.8	-79.9	20.1	22.1
2018/5/2	4.4	5.6	4	5.5	4.3	5.7	-21	-20.5	-81.7	-79.8	19.4	22.1
2018/5/3	4.4	5.2	4.1	5.6	4.3	5.7	-21.2	-20.6	-81.8	-79.9	18.7	20.1
2018/5/4	4.5	5.3	4.1	5.7	4.3	5.7	-21.1	-20.6	-81.7	-79.9	18.7	20.1
2018/5/5	5.1	5.3	4.8	5.6	4.3	4.8	-21.1	-20.8	-81.8	-80	18.7	19.1
2018/5/6	5.1	5.3	4.7	5.7	4.3	5.1	-21.2	-20.9	-81.7	-79.7	18.6	18.9
2018/5/7	4.3	5.3	4.2	5.7	4.3	5.6	-21.2	-20.4	-81.7	-79.9	18.4	19.6
2018/5/8	3.8	5.1	4.2	5.6	4.3	5.6	-21.1	-20.8	-81.7	-80	18.4	20.8
2018/5/9	3.9	4.8	4.2	5.7	4.2	5.4	-21.2	-20.4	-81.8	-79.9	18.2	19.6
2018/5/10	4.2	5.3	4.3	5.6	4.2	4.8	-20.9	-20.7	-81.8	-79.8	18.1	19.4
2018/5/11	4.2	5.3	4.2	5.6	3.9	4.8	-21.1	-20.5	-81.8	-79.9	18	19.3
2018/5/12	4.8	5.1	4.8	5.7	3.9	4.4	-20.9	-20.6	-81.7	-80	17.9	18.3
2018/5/13	4.8	5.1	4.9	5.7	3.9	4.4	-20.9	-20.6	-81.7	-79.9	17.7	17.9
2018/5/14	3.8	5	4.2	5.7	3.9	5.1	-20.9	-20.5	-81.7	-80	17.7	19.3
2018/5/15	3.8	4.7	4.2	5.7	4.2	4.9	-20.9	-19.6	-81.7	-79.9	17.8	19.2
2018/5/16	3.9	4.7	4.2	5.7	4.2	5.6	-20.9	-20	-81.8	-79.9	17.9	19.2
2018/5/17	3.9	4.9	4.2	5.6	4.9	5.6	-20.9	-20.4	-81.8	-79.9	18.1	19.3
2018/5/18	3.9	5.3	3.8	5.7	4.9	5.6	-20.9	-19.9	-81.8	-79.9	18.2	19.4
2018/5/19	5.1	5.3	4.5	5.4	5.3	5.5	-20.4	-20.2	-81.8	-79.9	17.9	18.4
2018/5/20	5	5.3	4.5	5.5	5.3	5.5	-20.4	-20.2	-81.8	-80	17.8	17.9
2018/5/21	4.2	5.3	3.7	5.4	4.9	5.5	-20.8	-20.2	-81.8	-79.9	17.7	19.4
2018/5/22	4.3	5.3	3.8	5.3	4.9	5.7	-20.9	-20.6	-81.8	-79.9	18.4	20.2
2018/5/23	4.1	5.4	3.7	5.4	4.9	5.7	-20.9	-20.4	-81.8	-79.9	19.1	20.6
2018/5/24	4	5.3	3.9	5.3	4.9	6.1	-20.9	-20.6	-81.7	-79.9	19.3	20.8
2018/5/25	3.8	5.2	3.9	5.3	4.7	5.3	-20.9	-20.7	-81.7	-79.9	19.2	20.4
2018/5/26	4.7	5.1	4.3	5.3	5.1	5.3	-21.1	-20.8	-81.8	-79.9	19.4	20.2
2018/5/27	4.7	5	4.3	5.3	5.1	5.3	-21.1	-20.8	-81.7	-79.9	19.4	19.6
2018/5/28	3.8	4.9	4.3	5.3	4.8	5.4	-21.1	-20.7	-81.8	-79.8	19.4	20.8
2018/5/29	3.8	4.9	3.8	5.3	4.6	5.3	-21.1	-20.6	-81.8	-79.8	19.6	20.7
2018/5/30	3.4	4.9	4.3	5.3	4.4	5.2	-20.9	-20.6	-81.7	-79.9	19.3	20.4
2018/5/31	3.5	4.6	3.8	5.3	4.6	5.2	-20.9	-20.6	-81.7	-79.9	19.3	20.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 1070601

藥學部 陳盈如 1070601

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年05月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2018/5/1	4.3	5.1	4.8	5.7	22.8	23.4
2018/5/2	4.2	5.9	4.7	5.4	23	24
2018/5/3	4.2	4.9	4.7	5.3	21.8	23.1
2018/5/4	4.3	5.3	4.7	5.2	21.7	22.2
2018/5/5	4.4	4.9	4.7	5.2	21.7	22.2
2018/5/6	4.4	4.9	4.6	5.2	21.2	21.7
2018/5/7	4.3	4.9	4.7	5.2	20.7	21.9
2018/5/8	4.3	4.9	4.7	5.7	21.3	22.8
2018/5/9	4.3	5.4	4.7	5.2	21.3	22.1
2018/5/10	4.3	4.9	4.5	5.2	21.1	21.8
2018/5/11	4.3	4.9	4.5	5.6	21.1	21.7
2018/5/12	4.3	4.9	4.7	5.3	21.1	21.4
2018/5/13	4.3	4.9	4.7	5.2	20.7	21.1
2018/5/14	4.3	4.9	4.6	5.2	20.5	21.3
2018/5/15	3.6	4.9	4.7	5.2	20.7	21.5
2018/5/16	4.1	4.9	4.4	5.2	20.8	21.7
2018/5/17	4.1	4.8	4.5	5.2	21.1	22.3
2018/5/18	4.1	4.8	4.6	5.3	21.3	21.9
2018/5/19	4.1	4.7	4.7	5.2	21.2	21.8
2018/5/20	4.1	4.7	4.7	5.2	20.7	21.2
2018/5/21	4.1	4.9	4.4	5.2	20.5	21.6
2018/5/22	4.2	4.9	4.4	5.1	21.1	22.2
2018/5/23	4.2	4.9	4.5	5.2	21.7	22.6
2018/5/24	4.1	4.9	4.6	5.2	21.8	22.5
2018/5/25	4.2	4.9	4.6	5.2	21.9	22.7
2018/5/26	4.2	4.9	4.7	5.2	22.1	22.4
2018/5/27	4.3	4.9	4.6	5.2	21.9	22.2
2018/5/28	4.2	4.9	4.6	5.2	21.7	22.6
2018/5/29	4.3	5.1	4.5	5.2	22	22.8
2018/5/30	4.3	5	4.5	5.2	21.9	22.6
2018/5/31	4.3	5.2	4.6	5.2	21.9	22.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 1070601
師(三)級藥師 藥學部 陳盈如 1070601
師(三)級藥師