

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年03月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-292		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2018/3/1	4.7	5.7	4.2	5.4	4.3	5.1	-20.9	-19.9	-81.5	-72.9	19.7	20.3
2018/3/2	3.8	5.9	4.1	5.4	4.3	4.9	-20.9	-20.7	-81.7	-79.7	19.5	20.4
2018/3/3	4.3	4.6	4.8	5.4	4.4	4.9	-20.9	-20.7	-81.6	-79.9	19.5	19.7
2018/3/4	4.4	4.6	4.7	5.4	4.4	4.9	-20.9	-20.7	-81.7	-79.7	19.4	19.6
2018/3/5	3.8	4.9	4.1	5.4	4.3	4.9	-21.1	-20.1	-81.7	-79.9	19.4	20.7
2018/3/6	3.7	4.7	4.2	5.5	4.3	5.2	-21.1	-20.7	-81.7	-79.7	19.4	20.8
2018/3/7	3.7	5.2	4.2	5.4	4.4	5.2	-21.2	-20.7	-81.6	-79.9	18.6	21.5
2018/3/8	4	5.2	4.3	5.5	4.4	5.2	-21.2	-20.7	-81.7	-79.9	18.4	20.6
2018/3/9	3.9	5.1	4.3	5.5	4.5	5.2	-21.2	-20.7	-81.7	-79.9	19.2	20.7
2018/3/10	4.8	5.1	4.9	5.5	4.8	5.3	-21.1	-20.8	-81.7	-79.6	19.9	20.3
2018/3/11	4.9	5.2	4.9	5.4	4.8	5.2	-20.9	-20.7	-81.6	-79.9	19.9	20.8
2018/3/12	3.9	5.1	4.3	5.5	4.5	5.2	-21.2	-20.8	-81.7	-79.7	19.5	20.9
2018/3/13	3.8	4.9	4.4	5.5	4.4	5.2	-21.2	-20.8	-81.7	-79.8	19.7	20.8
2018/3/14	3.9	5.2	4.4	5.8	4.4	5.1	-21.1	-20.7	-81.7	-79.9	19.7	20.8
2018/3/15	4.3	5.8	4.4	5.8	4.4	5.1	-20.9	-20.7	-81.5	-79.8	20.2	21.3
2018/3/16	4.2	5.4	4.3	5.7	4.4	5.1	-21.1	-20.7	-81.7	-79.9	19.9	20.8
2018/3/17	4.9	5.2	5.1	5.7	4.6	5.1	-21.1	-20.7	-81.7	-79.9	19.6	19.9
2018/3/18	4.9	5.2	5.1	5.7	4.6	5.1	-21.1	-20.8	-81.7	-80	19.4	19.6
2018/3/19	3.9	5.2	4.3	5.8	4.4	5.1	-21	-20.7	-81.7	-79.8	19.2	20.6
2018/3/20	4	5	4.3	5.9	4.4	5.3	-21	-20.6	-81.7	-79.9	19	20.7
2018/3/21	4.1	5.1	4.3	5.9	4.5	5.3	-20.9	-20.6	-81.8	-79.9	19.5	20.6
2018/3/22	4.2	5.1	4.3	5.8	4.2	5.2	-20.9	-19.7	-81.8	-79.7	19.3	20.2
2018/3/23	4.1	5.1	4.3	5.9	4.2	4.9	-21.1	-19.9	-81.8	-79.8	19.2	20.3
2018/3/24	4.7	5.1	5.3	5.9	4.6	4.9	-21.1	-20.8	-81.7	-79.9	19.4	19.6
2018/3/25	4.7	5.1	5.3	5.9	4.6	4.9	-21.1	-20.8	-81.7	-79.8	19.2	19.4
2018/3/26	4.2	5.1	4.3	5.9	4.2	4.9	-21.1	-20.3	-81.7	-79.8	19.2	20.7
2018/3/27	4.3	5.4	4.3	5.8	4.2	4.8	-20.9	-20.5	-81.7	-79.9	19.6	20.9
2018/3/28	4.4	5.5	4.4	5.7	4.2	4.7	-21	-20.7	-81.7	-79.9	19.8	21.1
2018/3/29	4.2	5.1	4.3	5.8	4.2	4.7	-21.1	-20.7	-81.7	-79.9	19.9	21.2
2018/3/30	4.2	4.9	4.3	5.8	4.3	4.8	-21.1	-20.5	-81.7	-79.8	20.1	21.4
2018/3/31	4.7	4.9	5.1	5.8	4.4	4.8	-21.1	-20.8	-81.7	-79.9	20.1	20.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

107
藥學部 邱保祥
師(三)級藥師

107
0103