

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年02月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-292		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2018/2/1	4.5	5.3	4.6	5.7	4.5	5.3	-20.7	-19.7	-81.7	-80.1	18.6	19.6
2018/2/2	4.7	5.8	4.4	5.7	4.5	5.2	-20.9	-20.4	-81.7	-80	18.4	19.4
2018/2/3	5.6	5.9	5.1	5.7	4.5	5.1	-20.9	-20.6	-81.7	-80	18.6	19.4
2018/2/4	5.7	5.8	5.1	5.8	4.6	5.1	-20.9	-20.6	-81.7	-80	19.4	19.4
2018/2/5	4.4	5.8	4.5	5.8	4.4	5.3	-20.9	-19.9	-81.7	-79.9	19.2	20.7
2018/2/6	3.4	5.5	4.4	5.8	4.4	5.2	-20.9	-20	-81.5	-79.9	19.2	21.5
2018/2/7	3.9	5.2	4.4	5.9	4.3	5.3	-20.9	-20.3	-81.7	-80	19.5	20.5
2018/2/8	3.7	5.4	4.4	5.8	4.4	5.3	-21.1	-19.8	-81.7	-80.1	19.2	20.4
2018/2/9	4.4	5.6	4.5	5.8	4.4	5.6	-20.9	-19.9	-81.7	-79.9	19.6	21.1
2018/2/10	4.9	5.2	5.1	5.8	4.5	5.1	-21.1	-20.8	-81.5	-80.1	19.4	20.2
2018/2/11	4.8	5.2	5.1	5.8	4.5	5.1	-21.2	-20.8	-81.7	-80.1	18.8	19.3
2018/2/12	4.1	5.6	4.6	5.8	4.6	5.7	-21.2	-17.2	-81.7	-80.1	18.6	19.8
2018/2/13	4.2	5.6	4.6	5.8	4.6	5.6	-20.7	-19.6	-81.7	-80.1	18.6	19.9
2018/2/14	3.9	4.9	4.5	5.8	4.6	5.7	-20.8	-20.1	-81.7	-80.1	19	20.2
2018/2/15	4.5	4.9	5.1	5.7	4.6	5.1	-20.4	-20.4	-81.5	-79.8	19.3	19.5
2018/2/16	4.5	4.9	5.1	5.7	4.6	5.1	-20.4	-20.2	-81.7	-80.1	19.4	19.6
2018/2/17	4.5	4.9	5.1	5.7	4.6	5.1	-20.4	-20.4	-81.5	-80.1	19.4	19.6
2018/2/18	4.5	4.9	5.1	5.7	4.6	5.1	-20.6	-20.3	-81.7	-80.1	18.7	19.4
2018/2/19	4.5	4.9	5.1	5.7	4.6	5.1	-20.6	-20.3	-81.7	-80.1	18.6	18.7
2018/2/20	3.9	4.8	4.6	5.7	4.6	5.8	-20.7	-20.2	-81.5	-80.1	18.6	19.3
2018/2/21	3.9	5.1	4.4	5.7	4.4	5.7	-20.9	-20.6	-81.7	-79.9	18.6	19.8
2018/2/22	4.3	5.7	4.4	5.8	4.4	5.6	-21.1	-20.8	-81.5	-79.8	18.7	20.9
2018/2/23	4.4	5.7	4.4	5.7	4.7	5.6	-21.1	-20.7	-81.7	-79.9	18.8	21.1
2018/2/24	5.3	5.4	4.9	5.7	4.7	5.1	-21	-20.8	-81.7	-80.1	18.7	18.9
2018/2/25	5.3	5.4	4.9	5.7	4.7	5.1	-20.9	-18.9	-81.7	-80.1	18.6	18.7
2018/2/26	4.7	5.8	4.3	5.7	4.7	5.8	-20.9	-20.4	-81.6	-80.1	18.4	19.9
2018/2/27	5.1	6.1	4.4	5.6	4.5	5.5	-20.8	-20.5	-81.5	-80.1	19.2	20.3
2018/2/28	5.9	6.1	4.6	5.6	4.5	4.9	-20.8	-20.6	-81.7	-80.1	19.4	19.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 107
 藥學部 溫雅如 107
 師(三)級藥師 0301