

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年02月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1714-0116	
備用	1611-0052		1715-0119		1714-0117	
範圍	2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max
2018/2/1	4.5	5.3	3.3	4.3	20.3	20.9
2018/2/2	4.5	5.1	3.3	4.3	20.2	21
2018/2/3	4.5	5.1	3.3	4.3	20.5	20.9
2018/2/4	4.5	5	3.3	4.3	20.2	20.7
2018/2/5	4.5	6.3	3.3	4.5	19.9	20.5
2018/2/6	5.5	5.8	3.4	4.4	19.7	20.4
2018/2/7	5.5	5.8	3.7	4.7	20.1	20.7
2018/2/8	5.1	6.2	3.7	5.3	20.3	20.9
2018/2/9	4.9	5.3	3.7	5.1	20.6	21.7
2018/2/10	4.8	5.2	4.1	4.7	21.1	21.7
2018/2/11	4.8	5.1	4	4.7	20.2	21.1
2018/2/12	4.6	5.2	4	4.9	19.7	20.6
2018/2/13	4.6	5.7	4.3	4.9	19.9	21.1
2018/2/14	5.2	5.8	4.3	5	20.4	21.5
2018/2/15	5.1	5.4	4.3	4.9	20.9	21.4
2018/2/16	5.1	5.4	4.4	5	21.1	21.3
2018/2/17	5.1	5.4	4.4	4.9	20.9	21.3
2018/2/18	5.1	5.4	4.3	4.9	20.6	21.1
2018/2/19	4.9	5.3	4.3	4.9	20.3	20.6
2018/2/20	5.1	5.4	4.3	4.9	20.2	20.7
2018/2/21	4.9	5.9	4.4	5.1	20.3	21.2
2018/2/22	4.9	5.4	4.5	5.1	20.6	21.3
2018/2/23	4.9	5.5	4.5	5.2	20.7	21.7
2018/2/24	5.1	5.4	4.6	5.2	20.7	21.2
2018/2/25	4.9	5.4	4.6	5.2	20.6	21.1
2018/2/26	4.9	5.7	4.6	5.1	20.2	21.3
2018/2/27	5.1	5.6	4	5.1	20.8	21.9
2018/2/28	5.1	5.4	4.2	4.8	21.3	21.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師: 藥學部 邱保祥 1070301
 藥學部 溫雅如 1090301