

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年01月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-292		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2018/1/1	4.6	4.8	5.1	5.8	3.9	4.4	-20.6	-20.2	-81.5	-79.7	18.8	19.8
2018/1/2	4.3	5.2	4.8	5.8	4	4.9	-20.9	-20.2	-81.6	-79.8	18.7	19.6
2018/1/3	4.3	5.3	4.8	5.8	4	5.7	-20.9	-20.2	-81.7	-79.9	19.2	20
2018/1/4	4.4	5.4	4.4	5.8	4.4	5.6	-20.9	-20.5	-81.7	-79.7	19	19.9
2018/1/5	4.5	5.6	4.3	5.8	4.4	5.6	-20.9	-20.2	-81.7	-80.1	19	20
2018/1/6	5.1	5.2	4.7	5.7	4.4	4.9	-21.1	-20.8	-81.7	-80	18.8	19.2
2018/1/7	5.1	5.3	4.7	5.7	4.4	4.9	-20.9	-20.7	-81.5	-79.8	18.8	19.8
2018/1/8	4.5	5.3	4.3	5.7	4.5	5.3	-20.9	-20.7	-81.5	-80.1	19.4	20.4
2018/1/9	4.4	6	4.3	5.8	4.4	5.6	-20.9	-20.7	-81.7	-80	19.3	20.2
2018/1/10	4.6	5.8	4.6	5.8	4.2	5.6	-21.1	-20.6	-81.5	-80.1	18.8	20.4
2018/1/11	4.5	5.4	4.6	5.7	4.2	5.4	-21.1	-20.7	-81.6	-80.1	18.8	20.1
2018/1/12	4.2	5.3	4.6	5.7	4.2	5.4	-21.2	-20.8	-81.5	-80.1	18.4	19.7
2018/1/13	4.7	5.1	5.1	5.7	4.2	4.6	-21.2	-20.9	-81.7	-80.1	18.2	19.2
2018/1/14	4.7	5.1	5.2	5.8	4.2	4.6	-21.1	-20.8	-81.7	-79.8	19.2	19.6
2018/1/15	4.1	5.1	4.6	5.8	4.2	5.3	-21.1	-20.7	-81.5	-80	19.6	21.2
2018/1/16	4.3	5.1	4.4	5.8	4.2	5.3	-20.9	-20.7	-81.5	-80.1	19.3	20.4
2018/1/17	4.3	5.1	4.4	5.8	4.2	5.2	-21.1	-20.6	-81.5	-79.9	19	20.1
2018/1/18	4.3	4.9	4.4	5.7	4.2	4.9	-21.1	-20.7	-81.5	-79.8	19	20.2
2018/1/19	4.4	5.3	4.6	5.7	4.2	5.2	-21.1	-20.8	-81.5	-80.1	18.8	20.1
2018/1/20	5.1	5.4	5.1	5.7	4.2	4.7	-21.1	-20.8	-81.7	-79.9	18.8	19.1
2018/1/21	5.2	5.4	5.1	5.7	4.2	4.6	-21.1	-20.8	-81.5	-80.1	18.6	18.8
2018/1/22	4.2	5.4	4.5	5.7	4.2	5.1	-21.2	-20.7	-81.7	-79.9	18.6	19.6
2018/1/23	4.5	5.5	4.5	5.7	4.2	5.1	-21.2	-20.2	-81.5	-80	18.7	19.9
2018/1/24	4.7	5.6	4.4	5.7	4.2	5.2	-21.4	-20.4	-81.5	-80.1	18.8	19.8
2018/1/25	4.7	5.6	4.6	5.7	4.2	5.2	-21.2	-20.8	-81.5	-79.9	18.7	21.4
2018/1/26	4.5	5.7	4.5	5.8	4.2	5.3	-21.2	-20.7	-81.5	-80	18.8	21.5
2018/1/27	5.5	5.7	4.9	5.7	4.2	4.7	-21.2	-20.9	-81.6	-80.1	18.9	19.4
2018/1/28	5.4	5.7	4.9	5.7	4.2	4.7	-21.2	-20.9	-81.6	-79.9	18.7	18.9
2018/1/29	4.5	5.7	4.5	5.7	4.2	5.2	-21.4	-20.9	-81.7	-80.1	18.6	19.6
2018/1/30	4.7	5.7	4.5	5.7	4.2	5.3	-21.5	-21.1	-81.7	-79.9	18.3	19.6
2018/1/31	4.5	5.7	4.5	5.7	4.2	5.3	-21.5	-20.9	-81.7	-80.1	18.6	19.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

藥學部 邱保祥 (07) 201
 藥學部 溫雅如 (07) 201
 承辦藥師: