

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 106年12月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-292		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/12/1	4.5	6.4	5.1	5.9	4	5.1	-21.7	-20.3	-81.5	-80.1	20.3	20.8
2017/12/2	3.6	6.2	5.3	5.9	4.7	5.1	-21.7	-21.4	-81.5	-80.1	20.1	20.5
2017/12/3	3.6	6.1	5.3	5.9	4.7	5.1	-21.7	-21.4	-81.5	-80.2	19.9	20.1
2017/12/4	3.8	6.4	5.1	5.9	4.1	5.1	-21.7	-20.3	-81.5	-80.2	19.9	20.5
2017/12/5	4.1	6.6	4.9	5.9	3.9	5.1	-21.5	-20.9	-81.5	-80.1	19.9	20.3
2017/12/6	4.2	6.4	4.7	5.9	4.3	5.3	-21.5	-21.2	-81.5	-80.1	19.7	20.4
2017/12/7	3.8	6.4	4.9	5.9	4.6	5.6	-21.5	-20.9	-81.5	-80.2	20.1	20.6
2017/12/8	3.8	6.5	5.1	5.9	4.4	5.7	-21.7	-21.2	-81.5	-80.2	19.9	20.3
2017/12/9	3.3	6.2	5.2	5.9	5.3	5.6	-21.7	-21.3	-81.5	-80.2	19.6	19.9
2017/12/10	3.3	6.1	5.2	5.8	5.3	5.6	-21.7	-21.4	-81.5	-80.2	19.2	19.7
2017/12/11	4.1	6.4	4.8	6	4.4	5.4	-21.7	-21.1	-81.5	-80.1	19	19.5
2017/12/12	3.8	6.4	4.8	5.9	4.3	5.4	-21.6	-20.9	-81.5	-80.2	18.5	19.6
2017/12/13	3.7	6.6	4.7	5.8	3.6	5.5	-21.5	-21.1	-81.6	-80.2	18.6	20.6
2017/12/14	3.4	5.8	4.8	5.9	4.1	5.8	-21.6	-21.2	-81.5	-80	19.6	20.3
2017/12/15	3.5	5.1	4.7	5.9	4.6	5.7	-21.7	-21.2	-81.7	-80.2	19.4	20.4
2017/12/16	4.7	5.1	5.1	5.7	4.6	5.1	-21.6	-21.2	-81.7	-80.2	19.2	19.7
2017/12/17	4.7	4.9	5.1	5.7	4.5	5.1	-21.7	-21.4	-81.7	-80.1	18.8	19.2
2017/12/18	3.6	5.2	4.7	5.7	4.5	5.7	-21.6	-21.2	-81.5	-80.1	18.8	20
2017/12/19	3.4	4.5	4.7	5.7	4.6	5.8	-21.7	-21.1	-81.6	-80.2	18.8	20.1
2017/12/20	3.6	4.4	4.7	5.8	4.5	5.5	-21.9	-20.9	-81.7	-79.8	19.5	19.9
2017/12/21	3.3	5.2	4.9	5.9	4.7	5.6	-21.7	-20.9	-81.7	-80.1	19.4	20
2017/12/22	4.4	5.2	4.9	5.9	4.5	5.6	-21.4	-20.9	-81.5	-80.1	19.7	20.7
2017/12/23	4.7	5.1	5.1	5.7	4.5	5	-21.4	-21.2	-81.5	-80.2	20.2	20.4
2017/12/24	4.7	5.1	5.1	5.7	4.5	4.9	-21.6	-21.2	-81.5	-80.1	19.4	20.2
2017/12/25	4.4	5.3	4.8	6.4	4.5	5.8	-21.7	-21.2	-81.5	-76.7	19.4	20.2
2017/12/26	4.2	5.6	4.9	5.9	4.5	5.7	-21.5	-21	-81.5	-79.9	20.1	20.7
2017/12/27	4.2	5.2	4.8	5.9	4.7	5.9	-21.5	-21.1	-81.5	-80.2	20.3	20.7
2017/12/28	4.2	5.1	4.9	5.9	4.6	5.7	-21.5	-21.2	-81.5	-80.1	20.1	20.5
2017/12/29	4.2	5.1	4.8	5.9	4.2	5.7	-21.6	-21.1	-81.5	-80.1	20.1	20.6
2017/12/30	4.6	5.1	4.9	5.7	4.2	4.7	-21.4	-21.1	-81.5	-80.2	20.3	20.6
2017/12/31	4.6	5.1	4.9	5.7	4.2	4.7	-21.4	-21.1	-81.5	-80.1	20.4	20.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 107
 藥學部 溫雅如 107
 師(三)級藥師 邱保祥 102
 師(三)級藥師 溫雅如 102