

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 106年11月

	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1611-0062		1213-0010		1215-0006		1213-0007	
備用	0580-0835		0520-292		1504-0339		1611-0052		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/11/1	4.3	6.2	4.3	5.9	4.3	5.1	4	4.9	-22.1	-18.8	-80.9	-79.8	19.9	20.6
2017/11/2	4.3	6.3	4.3	5.9	4.2	5.1	4.1	5.6	-22.1	-21.4	-80.9	-79.8	19.7	20.9
2017/11/3	3.9	6.1	4.3	5.8	4.2	4.9	4.2	4.9	-22	-21.4	-80.9	-79.7	19	20.2
2017/11/4	4.1	6.1	5.1	5.7	4.5	4.9	4.3	4.8	-21.9	-21.4	-80.9	-79.8	19	19.7
2017/11/5	4.1	6.1	5.1	5.7	4.4	4.9	4.3	4.9	-21.9	-21.7	-80.9	-79.7	19.6	19.7
2017/11/6	3.8	6.1	4.3	5.7	4.1	4.9	4.3	4.9	-22	-21.4	-80.9	-79.8	19.4	20.3
2017/11/7	4.3	6.2	4.3	5.7	3.8	4.8	4.2	4.9	-22	-21.5	-80.9	-79.7	19.6	20.8
2017/11/8	4.4	6.5	4.2	5.7	4.2	4.8	4.2	5	-22.2	-21.4	-80.9	-79.7	19.8	21.1
2017/11/9	4.6	6.3	4.3	5.8	4.2	4.9	4.2	5.2	-22.2	-21.9	-80.9	-79.6	20.1	21.2
2017/11/10	4.5	6.4	4.2	5.7	4.2	4.8	4.2	4.9	-22.2	-21.9	-80.9	-79.7	19.7	20.8
2017/11/11	4.3	6.3	5	5.7	4.4	4.7	4.2	4.9	-22.2	-21.9	-80.9	-79.7	19.4	19.9
2017/11/12	4.2	6.2	5.1	5.7	4.4	4.7	4.2	4.9	-22.2	-21.9	-80.9	-79.8	19.3	19.5
2017/11/13	4.8	6.3	4.4	5.8	4	4.8	4.2	4.9	-22.2	-21.7	-80.9	-79.8	19.2	20.4
2017/11/14	4.2	6.4	4.5	6.1	4.1	4.9	4.2	4.9	-22.2	-21.8	-80.9	-79.8	19.4	20.6
2017/11/15	4.2	6.2	4.3	6.1	4.2	4.9	4.2	4.9	-22.2	-21.7	-80.9	-79.7	19.5	20.6
2017/11/16	4.6	6.2	4.4	5.8	4	4.9	4.2	5.6	-22	-21.6	-80.9	-79.7	19.5	20.8
2017/11/17	4.1	6.2	4.6	5.8	4.1	5.1	4.4	5.1	-22	-21.6	-80.9	-76.6	19.6	20.8
2017/11/18	3.9	6	5.2	5.8	4.6	4.9	4.4	5	-22	-21.7	-80.9	-79.8	19.3	19.8
2017/11/19	4	6	5.2	5.8	4.6	4.9	4.4	4.9	-22	-21.7	-80.9	-79.8	19.1	19.8
2017/11/20	4.7	6.1	4.6	5.7	4.1	5	4.4	4.9	-22.2	-21.7	-81.4	-79.8	19	20
2017/11/21	4.1	6.8	4.5	5.8	4.2	4.9	4.4	5.2	-22.2	-21.7	-81.4	-79.8	18.8	20.3
2017/11/22	4.9	6.5	4.6	5.8	4.2	4.9	4.9	5.5	-22.2	-21.8	-81.4	-79.9	19.1	20.2
2017/11/23	4.6	6.2	4.6	5.8	4.1	4.9	4.7	5.3	-22.4	-21.9	-81.4	-79.8	19.1	20.2
2017/11/24	4.5	6.4	4.6	5.9	4.2	4.9	4.7	5.2	-22.4	-21.9	-81.5	-79.9	18.9	20
2017/11/25	3.9	6.1	5.3	5.9	4.6	4.9	4.7	5.1	-22.4	-22	-81.4	-79.9	18.9	19.2
2017/11/26	3.9	6.1	5.4	5.9	4.6	4.9	4.7	5.1	-22.4	-22	-81.4	-79.9	18.8	18.9
2017/11/27	4.1	6.2	4.6	5.9	4.2	4.9	4.7	5.1	-22.4	-21.9	-81.4	-79.9	18.8	20
2017/11/28	4.4	6.7	4.6	5.9	4.2	5.1	4.6	5.1	-22.2	-20.3	-81.4	-79.8	19.1	20.2
2017/11/29	4.8	6.7	4.6	5.8	4.2	5.1	4.6	5.3	-20.9	-20.4	-81.3	-79.8	19.2	20.2
2017/11/30	4.6	6.8	4.6	5.9	4.1	5.1	4.6	5.3	-20.9	-20.3	-81.4	-79.9	19.4	20.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

藥學部 邱保祥 106
師(三)級藥師 201
藥學部 溫雅如 106
師(三)級藥師 201