

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 106年10月

	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1611-0052		1213-0010		1215-0006		1213-0007	
備用	0580-0835		0520-292		1504-0339		1611-0062		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/10/1	4.4	6.2	4.9	5.6	4.6	4.9	4.4	5.1	-21.2	-20.9	-80.9	-79.8	18.8	19.2
2017/10/2	5.1	6.5	4.2	5.4	4.2	4.9	4.4	5.2	-20.4	-19.6	-80.9	-79.8	19.1	19.9
2017/10/3	4.4	6.5	4.2	5.4	4.3	4.9	4.4	5.2	-20.6	-20	-80.9	-79.8	19	21.4
2017/10/4	4.2	6.2	4.7	5.4	4.6	4.9	4.5	5.2	-20.7	-20.3	-80.9	-79.8	18.8	19.4
2017/10/5	4.3	6.2	4.3	5.4	4.3	4.9	4.4	5.2	-20.8	-20.2	-80.9	-79.8	18.7	20
2017/10/6	4.4	6.2	4.4	5.5	4.3	4.9	4.4	5.2	-20.8	-20.4	-80.9	-79.7	18.8	19.9
2017/10/7	4.2	6.2	4.7	5.4	4.4	4.9	4.5	5.2	-20.9	-20.4	-80.9	-79.8	18.7	19.1
2017/10/8	3.9	6.1	4.7	5.4	4.4	4.8	4.4	5.2	-20.9	-20.6	-81	-79.8	18.6	18.7
2017/10/9	3.9	6.1	4.6	5.5	4.3	4.8	4.4	5.2	-20.9	-19.9	-80.9	-79.8	18.6	19.4
2017/10/10	4	6.1	4.9	5.5	4.4	4.8	4.4	5.1	-20.6	-20.3	-80.9	-79.8	18.6	18.7
2017/10/11	4.3	6.5	4.2	5.5	4.3	4.9	4.4	5.9	-21.2	-19.6	-80.9	-79.6	18.7	20.4
2017/10/12	4.5	6.4	4.3	5.4	4.3	4.9	4.4	5.2	-20.9	-20	-80.9	-79.7	19.5	20.8
2017/10/13	4.8	6.4	4.3	5.5	4.3	4.9	4.5	5.7	-20.8	-20.1	-80.9	-79.7	19.7	20.8
2017/10/14	4.5	6.4	4.7	5.4	4.6	4.9	4.8	5.2	-20.8	-20.4	-80.9	-79.7	19.7	20.1
2017/10/15	4.8	6.4	4.7	5.5	4.6	4.9	4.9	5.3	-20.9	-20.4	-80.9	-79.8	19.7	19.7
2017/10/16	4.1	6.4	4.2	5.5	4.2	4.9	4.9	5.3	-20.8	-20.4	-80.9	-79.7	19.5	20.8
2017/10/17	4.5	6.3	4.4	5.9	4.2	4.9	4.8	5.3	-20.9	-20.4	-80.9	-79.7	19.6	21.3
2017/10/18	4.6	6.4	4.4	5.9	4.1	4.8	4.5	5.3	-20.9	-20.6	-80.9	-79.7	19.7	20.8
2017/10/19	4.4	6.4	4.4	5.7	4.2	4.8	4.3	5.2	-20.9	-20.4	-80.9	-79.6	19.8	21.3
2017/10/20	4.7	6.3	4.5	5.7	4.2	5.7	4.3	5.1	-21.2	-20.5	-80.9	-79.6	19.7	20.8
2017/10/21	4.2	6.3	4.9	5.8	4.9	5.4	4.4	4.9	-20.9	-20.6	-80.9	-79.8	19.2	19.6
2017/10/22	4.2	6.2	4.9	5.8	4.9	5.3	4.4	4.9	-21	-20.7	-80.9	-79.8	19.1	19.4
2017/10/23	4.1	6.3	4.4	5.8	4.2	5.3	4.4	4.9	-21.1	-20.5	-80.9	-79.8	19	20.2
2017/10/24	4.3	6.3	4.7	5.9	4.2	5.4	4	4.9	-20.9	-19.9	-80.9	-79.8	19.2	20.2
2017/10/25	4.4	6.1	4.7	6	4.3	5.3	4	4.9	-20.3	-19.3	-80.9	-79.8	19.2	20.3
2017/10/26	3.9	6.1	4.7	6.1	4.3	5.3	4	4.8	-20.2	-19.7	-80.9	-79.8	19	20.1
2017/10/27	4	6	4.6	5.9	4.3	5.3	3.9	4.8	-20.7	-19.9	-80.9	-79.6	18.7	19.6
2017/10/28	3.8	6	5.3	6	4.9	5.3	4	4.6	-20.8	-20.4	-80.9	-79.7	18.4	18.8
2017/10/29	3.9	6	5.3	6	4.9	5.3	4	4.6	-20.8	-20.4	-80.9	-79.8	18.3	19.2
2017/10/30	4.6	6.1	4.7	6	4.3	5.3	3.9	4.6	-20.7	-20.3	-80.9	-79.8	19.2	20.4
2017/10/31	4.6	6.4	4.6	5.9	4.4	5.3	3.9	4.6	-20.7	-19.8	-80.9	-79.8	19.7	20.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部邱保祥¹⁰⁶₍₁₀₁₎
藥學部溫雅如¹⁰⁶₍₁₀₁₎