

臺北榮民總醫院藥學部
臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 106年10月

序號	六號冰箱		室溫	
	1436-0086	0436-049	1118-0011	0742-0169
備用	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2017/10/1	4	4.3	21.8	22.2
2017/10/2	3.9	5.7	21.8	22.2
2017/10/3	3.9	4.3	21.9	22.3
2017/10/4	4	4.3	22	22.2
2017/10/5	3.9	4.3	21.9	22.1
2017/10/6	4	4.6	21.9	22.5
2017/10/7	4.3	4.5	21.7	22
2017/10/8	4.3	4.5	21.6	21.9
2017/10/9	4.3	4.5	21.5	21.7
2017/10/10	4.3	4.5	21.5	21.9
2017/10/11	4.3	4.5	21.7	21.9
2017/10/12	4.3	4.7	21.7	22
2017/10/13	4.3	4.5	21.8	22.5
2017/10/14	4.3	4.4	21.7	22
2017/10/15	4.3	4.4	21.4	21.8
2017/10/16	4.3	4.4	21.3	21.7
2017/10/17	4.3	4.6	21.4	21.9
2017/10/18	4.3	4.5	21.3	21.9
2017/10/19	4.3	4.4	21.7	22.2
2017/10/20	4.2	4.4	21.8	22.2
2017/10/21	4.3	4.4	21.1	22
2017/10/22	4.3	4.4	20.6	21.1
2017/10/23	3.8	5.2	20.7	21.2
2017/10/24	3.8	4.1	20.4	21.3
2017/10/25	3.4	4.4	20.3	21.1
2017/10/26	3.4	4	20.6	21
2017/10/27	3.4	4	20.5	21.2
2017/10/28	3.4	3.9	20.4	20.8
2017/10/29	3.5	4	20.1	20.4
2017/10/30	3.5	4.1	19.9	20.6
2017/10/31	3.5	4.3	20.1	20.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄異動

承辦藥師: **藥學部邱保祥** (106)
(三)級藥師 MO

藥學部溫雅如
(三)級藥師 MO