

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 106年9月

	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1611-0052		1213-0010		1215-0006		1213-0007	
備用	0580-0835		0520-292		1504-0339		1611-0062		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/9/1	4.4	6.4	4.3	5.4	4.3	5.3	3.9	4.6	-21.2	-19.8	-80.6	-79.6	19.3	20.3
2017/9/2	4.3	5.6	4.7	5.5	4.9	5.4	4	4.6	-21.2	-20.9	-80.7	-79.6	19	19.6
2017/9/3	4.3	5.6	4.7	5.6	4.9	5.3	3.9	4.6	-21.2	-20.9	-80.7	-79.6	18.8	19
2017/9/4	3.8	5.7	4.3	5.5	4.2	5.3	3.9	4.6	-21.4	-20.8	-80.7	-79.6	18.7	19.8
2017/9/5	3.7	5.8	4.3	5.5	4.3	5.3	3.9	4.6	-21.4	-20.8	-80.7	-79.6	18.7	19.7
2017/9/6	4.3	5.7	4.3	5.5	4.3	5.3	3.9	4.6	-21.2	-20.8	-80.7	-79.6	18.8	19.9
2017/9/7	4.3	5.7	4.3	5.4	4.3	5.3	3.9	4.6	-21.1	-20.8	-80.7	-79.6	19	19.9
2017/9/8	3.9	5.7	4.3	5.5	4.4	5.3	3.9	4.6	-21.2	-20.4	-80.7	-79.7	19	20.2
2017/9/9	4.2	5.7	4.6	5.4	4.9	5.3	3.9	4.6	-21.1	-20.7	-80.7	-79.6	18.8	19.2
2017/9/10	4.3	5.6	4.6	5.4	4.9	5.3	3.9	4.6	-21.1	-20.8	-80.7	-79.7	18.6	18.8
2017/9/11	3.7	5.7	4.3	6	4.3	5.3	3.9	4.6	-21.1	-20.7	-80.8	-79.6	18.6	20.1
2017/9/12	3.9	5.7	4.3	5.6	4.3	5.3	3.9	4.8	-21.2	-20.6	-80.7	-79.6	18.8	19.9
2017/9/13	3.8	5.9	4.3	5.9	4.4	5.3	3.9	4.6	-21.2	-20.8	-80.7	-79.6	18.9	19.9
2017/9/14	4	5.8	4.3	5.6	4.4	5.3	3.9	4.8	-21.2	-20.9	-80.7	-79.6	18.9	19.9
2017/9/15	3.6	6.2	4.4	5.6	4.4	5.3	4	5.1	-21.3	-20.9	-80.7	-79.7	18.9	20.2
2017/9/16	4.4	6.2	4.7	5.5	5.1	5.4	4.6	5.1	-21.3	-20.9	-80.7	-79.6	18.8	19.1
2017/9/17	4.2	6.2	4.7	5.5	5.1	5.4	4.6	5.1	-21.2	-20.9	-80.7	-79.7	18.8	19.8
2017/9/18	2.9	6.2	4.4	5.4	4.3	5.4	4.5	5.2	-21.2	-20.8	-80.7	-79.6	19.7	20.9
2017/9/19	3.3	5.9	4.4	5.7	4.4	5.5	4.4	5.3	-21.2	-20.7	-80.7	-79.5	19.9	21.2
2017/9/20	4.2	6.1	4.6	5.8	4.4	5.7	4.4	5.1	-21.2	-20.8	-80.9	-79.7	20.1	21.4
2017/9/21	4.6	6.1	4.6	5.8	4.5	5.7	4.4	4.9	-21.1	-20.1	-80.9	-79.7	20.2	21.2
2017/9/22	4.3	6.1	4.3	5.5	4.4	5.8	4.4	4.9	-20.8	-20.2	-80.9	-79.7	20.2	21.4
2017/9/23	4.2	6.1	4.7	5.4	5.3	5.7	4.3	4.8	-20.9	-20.5	-80.9	-79.7	19.4	20.4
2017/9/24	3.9	6	4.7	5.4	5.3	5.6	4.3	4.8	-20.9	-20.7	-80.9	-79.8	18.9	19.4
2017/9/25	3.7	6	4.3	5.4	4.4	5.7	4.2	4.9	-20.9	-20.4	-80.9	-79.7	18.8	19.8
2017/9/26	3.6	6.4	4.4	5.5	4.4	5.6	4.5	5	-20.9	-19.3	-80.9	-79.8	18.9	19.9
2017/9/27	4.4	6.5	4.3	5.6	4.3	5.6	4.4	5	-21.1	-20.5	-80.9	-79.7	18.9	20.6
2017/9/28	4.7	6.5	4.3	5.6	4.3	4.9	4.4	5.5	-21.2	-20.7	-80.9	-79.7	19	20.1
2017/9/29	4.5	6.3	4.3	5.6	4.4	4.9	4.4	5.1	-21.2	-20.8	-80.9	-79.8	19	20.2
2017/9/30	4.4	6.4	4.9	5.6	4.6	4.9	4.4	5.1	-21.2	-20.8	-80.9	-79.8	19	20.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

藥學部 邱保祥 1061002
 藥學部 溫雅如 1061002