

**臺北榮民總醫院藥學部**  
**臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 106年8月**

	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1611-0052		1213-0010		1215-0006		1213-0007	
備用	0580-0835		0520-292		1504-0339		1611-0062		1405-0193		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/8/1	3.4	5.4	4.4	5.4	4.4	5.2	3.8	4.5	-20.3	-18.7	-80.7	-79.6	19.1	19.9
2017/8/2	3.2	5.6	4.5	5.4	4.3	5.2	3.8	4.5	-20.4	-19.5	-80.7	-79.5	18.8	20
2017/8/3	3.6	6.1	4.2	5.4	4.3	5.1	3.8	4.5	-20.2	-19.7	-80.7	-79.6	18.9	20.2
2017/8/4	3.4	5.9	4.1	5.4	4.2	5.1	3.8	4.5	-20.2	-19.8	-80.8	-79.6	19	21.3
2017/8/5	3.4	5.1	4.5	5.4	4.7	5.1	3.8	4.5	-20.2	-19.9	-80.7	-79.5	18.9	19.3
2017/8/6	3.4	5.1	4.5	5.4	4.7	5.1	3.8	6.7	-20.2	-19.9	-80.7	-79.6	18.8	18.9
2017/8/7	2.9	5.8	4.2	5.4	4.2	5.1	3.8	5.8	-20.2	-19.6	-80.7	-79.6	18.6	20
2017/8/8	2.7	5.2	4.2	5.5	4.2	5.1	3.8	4.3	-20.1	-19.7	-80.7	-79.5	18.9	20.1
2017/8/9	3.1	6.6	4.1	5.5	4.2	5.1	3.8	4.4	-20.2	-19.6	-80.8	-79.6	18.9	20.1
2017/8/10	4.7	6.2	4.1	5.5	4.2	5.1	3.8	4.4	-19.9	-19.6	-80.7	-79.6	19	20.2
2017/8/11	4.6	6.3	4.2	5.5	4.2	5.1	3.5	4.4	-19.9	-19.6	-80.7	-79.6	19.2	20.4
2017/8/12	4.7	5.9	4.6	5.5	4.6	5.1	3.6	4.3	-19.8	-19.6	-80.7	-79.7	19.1	19.3
2017/8/13	4.8	5.9	4.6	5.5	4.6	5.1	3.6	4.3	-19.8	-19.6	-80.8	-79.6	18.9	19.1
2017/8/14	4.5	6.2	4.2	5.6	4.2	5.1	3.6	4.6	-19.8	-19.4	-80.7	-79.6	18.8	20.1
2017/8/15	4.4	6.2	4.3	5.6	4.2	5.1	3.9	4.6	-19.7	-19.2	-80.7	-79.6	19	20.6
2017/8/16	4.6	6.2	4.3	5.7	4.2	5.1	3.9	4.7	-19.4	-18.8	-80.7	-79.6	19	20.2
2017/8/17	4.6	6.2	4.1	5.7	4.3	5	3.9	4.7	-19.9	-19.2	-80.7	-79.6	19.1	20.1
2017/8/18	4.6	6.5	4.1	5.6	4.2	5.1	3.8	4.6	-19.9	-19.6	-80.6	-79.6	19	20.1
2017/8/19	5.1	6.1	4.7	5.6	4.5	4.9	3.9	4.5	-20.2	-19.8	-80.7	-79.6	18.9	19.2
2017/8/20	5.1	6.1	4.7	5.6	4.5	4.9	3.9	4.6	-20.2	-19.9	-80.7	-79.6	18.8	18.9
2017/8/21	5	6.4	4.1	5.6	4.2	4.9	3.9	4.6	-20.2	-19.8	-80.6	-79.6	18.7	20.2
2017/8/22	4.7	6.3	4.3	5.9	4.3	5.1	3.9	4.7	-20.3	-19.8	-80.6	-79.5	19	20.1
2017/8/23	4.4	6.3	4.2	5.9	4.1	5.1	4	4.7	-20.6	-20.1	-80.7	-79.6	19.1	20.2
2017/8/24	4.4	6.3	4.2	5.5	3.9	5.1	4	4.7	-20.7	-20.2	-80.7	-79.6	19	20.2
2017/8/25	4.4	6.1	4.2	5.4	3.9	5.1	4.1	4.7	-20.8	-20.3	-80.7	-79.6	19.2	20.1
2017/8/26	4.3	5.7	4.6	5.4	4.7	5.1	4.1	4.6	-20.7	-20.3	-80.7	-79.6	19	19.2
2017/8/27	4.2	5.7	4.6	5.4	4.7	5.1	4.1	4.6	-20.7	-20.4	-80.7	-79.5	18.8	19
2017/8/28	4.5	6.6	4.2	5.5	3.9	5.1	4.1	4.6	-20.7	-20.3	-80.6	-79.4	18.8	19.8
2017/8/29	4.5	6.6	4.2	5.6	4	5.3	3.9	4.6	-20.6	-20.2	-80.7	-79.6	19	19.9
2017/8/30	4.8	6.6	4.4	5.7	4.2	5.4	3.8	4.6	-20.6	-20.3	-80.7	-79.6	19.1	20.2
2017/8/31	4	6.3	4.3	5.5	4.3	5.4	3.8	4.5	-24.5	-19.2	-80.7	-79.5	19.2	20.4

註1: 濕度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 濕度記錄器更換電池或送校正期間取備用紀錄器之資料

藥學部陳盈如  
師(三)級藥師

藥學部邱保祥  
師(三)級藥師

承辦藥師:

**臺北榮民總醫院藥學部**  
**臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 106年8月**

序號	六號冰箱		室溫	
	1436-0086	1118-0011	0436-049	0742-0169
備用	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2017/8/1	3.6	4.4	20.1	20.3
2017/8/2	3.5	4	19.9	20.7
2017/8/3	3.5	4.1	19.9	20.6
2017/8/4	3.6	3.9	19.7	20.6
2017/8/5	3.6	3.9	19.7	20.1
2017/8/6	3.6	3.9	19.7	20.1
2017/8/7	3.5	3.9	19.5	20.8
2017/8/8	3.5	3.9	19.8	20.7
2017/8/9	3.6	3.9	20.1	20.7
2017/8/10	3.5	3.9	20.1	20.9
2017/8/11	3.5	3.9	20.3	21.3
2017/8/12	3.5	3.9	20.3	20.9
2017/8/13	3.5	3.9	20.1	20.8
2017/8/14	3.5	3.9	20.1	21.1
2017/8/15	3.5	3.9	20.3	21.2
2017/8/16	3.5	3.9	20.4	21.2
2017/8/17	3.4	3.9	20.3	21.2
2017/8/18	3.4	3.9	20.3	21.2
2017/8/19	3.4	3.9	20.3	20.8
2017/8/20	3.4	3.9	20.1	20.7
2017/8/21	3.4	3.9	20.1	21.2
2017/8/22	3.3	3.9	20.4	21.1
2017/8/23	3.4	3.9	20.5	21.1
2017/8/24	3.4	4.4	20.6	21.4
2017/8/25	3.5	3.9	20.5	21.1
2017/8/26	3.6	3.9	20.5	20.8
2017/8/27	3.6	3.9	20.4	20.7
2017/8/28	3.4	4.1	20.3	21.2
2017/8/29	3.4	3.9	20.6	21.3
2017/8/30	3.4	3.9	20.6	21.3
2017/8/31	3.3	4	20.7	21.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

藥學部陳盈如  
0901  
0900  
(三)級藥師

承辦藥師:

藥學部邱保祥  
0901  
0900  
(二)級藥師