

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 106年7月

	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1611-0052		1213-0010		1215-0006		1213-0007	
備用	0580-0835		0520-292		1504-0339		1611-0062		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/7/1	4.4	5.6	4.2	5.5	4.6	4.9	4.2	4.9	-21.2	-20.8	-80.9	-79.9	18.1	18.4
2017/7/2	4.7	5.6	4.2	5.5	4.6	4.9	4.2	4.9	-21.2	-20.9	-80.9	-79.9	17.9	18.1
2017/7/3	4.3	5.6	4.2	5.8	4.4	4.9	4.2	4.9	-21	-17.9	-80.9	-79.9	17.9	19.8
2017/7/4	4.5	6.1	4.9	5.9	4.3	5.1	4.2	4.9	-20.9	-20.2	-80.9	-79.9	18.6	20.2
2017/7/5	4.7	5.7	4.8	5.8	4.2	5.1	4.2	4.9	-20.7	-20.4	-80.9	-79.9	18.7	20.1
2017/7/6	4.8	5.7	4.7	5.8	4.3	5.1	4.1	4.9	-20.8	-20.4	-80.9	-79.9	18.8	20
2017/7/7	4.7	6.1	4.7	5.8	4.4	5.2	4	4.7	-21.1	-20.5	-80.9	-79.9	19	21.5
2017/7/8	4.8	5.8	5.1	5.7	4.8	5.2	4.1	4.7	-21.2	-20.9	-80.9	-79.7	19.4	19.7
2017/7/9	5.1	5.7	5.1	5.7	4.8	5.2	4.1	4.7	-21.2	-20.9	-80.9	-79.9	19.2	19.4
2017/7/10	4.4	6.1	4.7	5.7	4.3	5.3	4.1	4.7	-21.2	-19.9	-80.9	-79.7	19.2	21.1
2017/7/11	4.6	5.7	4.7	5.8	4.3	5.3	3.8	4.7	-20.9	-20.2	-80.9	-79.9	19.4	21.3
2017/7/12	4.6	5.5	4.8	5.7	4.6	6.6	4.2	5.2	-20.9	-20.3	-80.9	-79.9	19.4	21
2017/7/13	4.6	5.4	4.8	5.8	5.6	6.6	4.3	4.8	-20.8	-20.2	-80.9	-79.7	19.6	20.9
2017/7/14	3.1	5.3	4.7	5.8	4.8	6.5	4.1	4.8	-20.8	-17.8	-80.9	-79.8	19.6	21.2
2017/7/15	3.5	6.1	5.2	5.9	5.2	6.2	3.9	5.2	-20.9	-20.6	-80.9	-79.9	18.9	19.8
2017/7/16	3.6	4.6	5.2	5.8	5.3	5.6	3.9	4.5	-20.9	-20.7	-80.9	-79.9	19.1	19.9
2017/7/17	3.3	7	4.9	6	4.7	5.7	3.9	4.5	-20.9	-19.9	-80.9	-79.8	19.1	20.2
2017/7/18	3.6	7.1	5.1	5.8	3.1	5.7	3.8	4.5	-20.9	-20.4	-80.9	-79.9	19.1	20.4
2017/7/19	4.4	6.3	4	5.8	3.6	4.7	3.7	4.4	-20.9	-20.2	-80.9	-79.9	19.4	20.8
2017/7/20	2.5	6.2	4	5.2	3.2	4.7	3.7	4.5	-20.6	-20.2	-80.9	-79.8	19.3	20.2
2017/7/21	3.3	5.9	3.9	5.2	3.2	4.7	3.8	4.4	-20.6	-18.8	-80.9	-79.9	18.9	20.4
2017/7/22	2.9	4.9	3.9	5.2	4	4.7	3.8	4.4	-20.7	-18.9	-80.9	-79.9	18.9	19.3
2017/7/23	2.9	4.9	3.9	5.2	4	4.7	3.8	4.4	-20.7	-20.4	-80.9	-79.9	18.7	18.9
2017/7/24	3.2	5.7	3.9	5.7	3.2	5.6	3.8	5.2	-22.9	-18.1	-80.9	-79.9	18.6	19.9
2017/7/25	2.5	5.2	4.1	5.4	4.5	5.6	4.1	4.7	-22.9	-19.9	-80.9	-78.8	18.9	20.2
2017/7/26	2.6	5.1	4	5.4	4.4	5.5	4.1	4.7	-20.7	-20.4	-80.9	-79.6	18.8	20.1
2017/7/27	2.7	5.1	4	5.4	4.6	5.5	4.1	4.7	-20.8	-19.9	-80.8	-79.7	18.9	20.3
2017/7/28	2.2	4.9	3.9	5.1	4.4	5.1	4.1	4.7	-20.9	-20.4	-80.8	-79.6	18.9	20.4
2017/7/29	2.2	4.5	3.9	5.1	4.6	4.9	4.3	4.7	-20.8	-20.5	-80.7	-79.6	19	19.3
2017/7/30	2.3	4.5	4	5	4.6	4.9	4.3	4.7	-20.8	-20.5	-80.8	-79.5	18.8	19.1
2017/7/31	2.2	4.6	4	5	4.3	4.9	4.2	4.7	-20.8	-20.4	-80.8	-79.6	18.8	19.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

藥學部 溫雅如 106
師(三)級藥師 0801
藥學部 邱保祥 106
師(三)級藥師 0801

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 106年7月

	六號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8℃		15~25℃	
日期	Min	Max	Min	Max
2017/7/1	3.8	4.1	18.8	19.3
2017/7/2	3.8	4.1	18.8	19
2017/7/3	3.7	4.3	18.9	19.6
2017/7/4	3.7	4.9	19	20
2017/7/5	3.9	4.2	18.9	19.5
2017/7/6	3.9	4.3	18.8	19.4
2017/7/7	4	4.3	18.8	19.2
2017/7/8	3.9	4.3	18.8	19.2
2017/7/9	3.9	4.2	18.7	19.5
2017/7/10	3.9	4.5	18.8	19.9
2017/7/11	4	4.4	19	19.8
2017/7/12	4.1	4.4	18.8	19.6
2017/7/13	4.1	4.4	19	19.8
2017/7/14	4	4.4	19.3	19.8
2017/7/15	4	4.4	19.4	19.7
2017/7/16	4.1	4.4	19.4	19.5
2017/7/17	3.7	4.4	19.3	19.9
2017/7/18	3.7	4.2	19.4	20.1
2017/7/19	3.7	4.2	19.6	19.9
2017/7/20	3.7	4.4	19.4	20.1
2017/7/21	3.9	4.3	19.5	20.8
2017/7/22	3.9	4.2	19.8	20.1
2017/7/23	3.9	4.2	19.6	20.1
2017/7/24	3.9	4.3	19.6	20.8
2017/7/25	3.9	4.4	19.8	20.6
2017/7/26	3.6	4.3	19.9	20.8
2017/7/27	3.6	4.1	19.9	20.7
2017/7/28	3.6	4	19.8	21.2
2017/7/29	3.7	4	20.1	20.4
2017/7/30	3.7	4	19.7	20.3
2017/7/31	3.7	4	19.7	20.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部溫雅如 1060801
藥學部邱保祥 1060801