

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 106年 6月

	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1611-0052		1213-0010		1215-0006		1213-0007	
備用	0580-0835		0520-292		1504-0339		1611-0062		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/6/1	2.3	4.7	3.2	5.1	4.3	5.3	4	4.9	-21.4	-19.8	-80.9	-79.8	19	20.2
2017/6/2	3.3	5.8	3.2	5.2	3.9	5	4	4.9	-21.6	-20.7	-80.9	-79.9	19	20.5
2017/6/3	5.1	5.8	3.3	5.2	4.7	5.1	4	4.8	-21.7	-21.3	-80.9	-79.9	19	19.4
2017/6/4	5.1	5.8	3.3	5.2	4.7	5.1	4	4.8	-21.7	-21.4	-80.9	-79.9	18.8	19.1
2017/6/5	5.1	5.8	3.3	5.2	4	5.1	4	4.8	-21.7	-21.2	-80.9	-79.9	18.8	20.1
2017/6/6	5.1	5.8	3.6	5.2	4.1	5.2	4	4.8	-21.7	-20.2	-80.9	-79.7	18.8	20.6
2017/6/7	5.2	6.1	3.9	5.2	4.3	5.2	3.9	4.8	-20.6	-19.9	-80.9	-79.8	19	20.6
2017/6/8	4.9	5.8	3.9	5.2	4.2	5.1	3.9	4.8	-20.9	-20.1	-80.9	-79.8	19.1	20.6
2017/6/9	4.8	5.8	3.9	5.2	4.2	5.2	4.2	4.9	-20.9	-20.2	-80.9	-79.8	19.2	20.2
2017/6/10	5.1	5.8	3.9	5.3	4.6	4.9	4.2	4.9	-20.6	-20.3	-80.9	-79.8	18.8	19.3
2017/6/11	4.9	5.9	4	5.3	4.6	4.9	4.2	4.9	-20.6	-20.3	-80.9	-79.8	18.6	18.8
2017/6/12	4.9	5.8	3.9	5.2	4.2	5.1	4.1	4.9	-21.1	-19.9	-80.9	-79.8	18.6	19.8
2017/6/13	5	5.8	3.8	5.2	4.2	5.1	4.1	5.6	-20.9	-20.2	-80.9	-79.8	18.7	19.8
2017/6/14	4.8	5.7	3.5	5.3	4.2	5.1	4.1	4.9	-20.7	-20.2	-80.9	-79.8	18.8	20.3
2017/6/15	4.9	5.8	3.4	5.2	4.2	5.1	4.2	5.1	-21.4	-20.3	-80.9	-79.8	19.3	20.8
2017/6/16	4.9	5.8	3.4	5.2	4.2	5.4	4.1	4.8	-21.4	-21.1	-80.9	-79.8	19.4	20.8
2017/6/17	4.9	5.8	3.4	5.1	5.1	5.3	4.1	4.9	-21.4	-21.1	-80.9	-79.9	19.2	19.6
2017/6/18	4.9	5.8	3.5	5.1	5.1	5.3	4.1	4.9	-21.4	-21.2	-80.9	-79.9	19	19.2
2017/6/19	4.9	5.9	3.5	5.5	4.5	5.3	4	4.9	-21.5	-20.9	-80.9	-79.8	18.8	20.3
2017/6/20	5.2	5.9	3.9	5.5	4.4	5.3	4	5.4	-21.7	-20.8	-80.9	-79.9	19.1	20.3
2017/6/21	4.9	5.9	3.9	5.4	4.7	5.4	4.1	4.8	-21.7	-18.9	-80.9	-79.9	19.2	20.7
2017/6/22	5.1	5.9	3.9	5.3	4.7	5.4	4.1	4.9	-19.6	-18.4	-80.9	-79.9	19.3	20.6
2017/6/23	5.1	5.9	3.9	5.3	4.5	5.4	4.1	4.9	-20.7	-19.1	-80.9	-79.9	19.3	20.6
2017/6/24	5.2	5.9	4.2	5.4	4.9	5.4	4.2	4.9	-20.8	-20.4	-80.9	-79.9	17.8	19.6
2017/6/25	4.8	5.8	4.2	5.4	4.9	5.3	4.2	4.9	-20.8	-20.5	-80.9	-79.9	17.5	18.3
2017/6/26	4.3	5.8	3.9	5.4	4.3	5.3	4.2	5.1	-21.1	-20.6	-80.9	-79.9	18.2	19.5
2017/6/27	4.3	5.6	4	5.5	4.3	5.2	4.2	5.1	-21.2	-20.5	-80.9	-79.8	18	19.3
2017/6/28	4.3	5.6	4.2	5.4	4.3	5.1	4.2	4.9	-21.1	-20.5	-80.9	-79.9	18.1	20.2
2017/6/29	4.3	5.6	4.2	5.7	4.3	5.1	4.2	4.9	-21.1	-20.5	-80.9	-79.9	18.3	19.5
2017/6/30	4.3	5.6	4.2	5.8	4.2	4.9	4.2	5.1	-21.2	-20.3	-80.9	-79.8	18.2	19.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部溫雅如 106
師(三)級藥師 0903

藥學部邱保祥 106
師(三)級藥師 0903

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 106年6月

	一號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8℃		15~25℃	
日期	Min	Max	Min	Max
2017/6/1	3.7	5.5	20.4	20.4
2017/6/2	3.4	4	20.6	20.6
2017/6/3	3.6	3.9	20.7	20.7
2017/6/4	3.5	3.9	20.9	20.9
2017/6/5	3.5	3.9	20.6	20.6
2017/6/6	3.5	3.9	20.4	20.4
2017/6/7	3.5	3.9	20.4	20.4
2017/6/8	3.7	4	19.7	19.7
2017/6/9	3.6	4	19.4	19.4
2017/6/10	3.7	3.9	19.4	19.4
2017/6/11	3.6	3.9	19.2	19.2
2017/6/12	3.7	4.2	19.5	20.3
2017/6/13	3.7	3.9	19.3	19.7
2017/6/14	3.6	3.9	19.4	19.8
2017/6/15	3.7	4.1	19.4	19.9
2017/6/16	3.7	4	19.4	19.9
2017/6/17	3.7	4	18.9	19.4
2017/6/18	3.7	4.1	18.2	19.4
2017/6/19	3.8	4.2	18.2	18.7
2017/6/20	3.8	4.1	18.4	18.7
2017/6/21	3.9	4.2	18.4	19.6
2017/6/22	3.9	4.7	18.9	19.4
2017/6/23	3.9	4.3	18.9	19.6
2017/6/24	3.9	4.2	18.8	19.3
2017/6/25	3.9	4.2	18.7	18.9
2017/6/26	3.9	4.2	18.6	19.2
2017/6/27	3.8	4.2	18.8	19.6
2017/6/28	3.8	4.2	18.9	19.8
2017/6/29	3.6	4.2	19.1	19.8
2017/6/30	3.6	4.2	19.1	19.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部溫雅如 106
師(三)級藥師 0903

藥學部邱保祥 106
師(三)級藥師 0903