

臺北榮民總醫院藥學部

臨床試驗藥局藥品冷藏/冷凍/室溫溫度管制表 106年 5月

	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1611-0052		1213-0010		1215-0006		1213-0007	
備用	0580-0835		0520-292		1504-0339		1611-0062		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/5/1	5.1	6	3.3	5.1	4.4	5.4	4.3	5.1	-20.3	-19.3	-81.2	-79.9	19.1	20.2
2017/5/2	5.1	6	3.3	5.1	4.3	5.5	4.3	5.1	-20.4	-19.9	-81	-79.9	19	20.6
2017/5/3	5.1	6.1	3.4	5.2	3.9	5.5	4.2	5.1	-20.4	-19.8	-81.1	-79.9	19.2	20.9
2017/5/4	4.8	5.8	3.4	4.9	4.2	5.2	4.2	4.9	-20.2	-19.7	-81.2	-79.8	19.4	21.3
2017/5/5	5.1	5.8	3.3	4.9	4.2	5.2	4.2	4.9	-20.3	-19.9	-81.1	-79.9	19.6	21.5
2017/5/6	5.1	5.7	3.3	4.9	4.6	5.3	4.2	4.9	-20.3	-20	-81.2	-80	18.9	19.7
2017/5/7	5.1	5.9	3.3	4.9	4.6	4.9	4.2	4.9	-20.3	-19.9	-81.1	-79.9	18.6	19.8
2017/5/8	4.9	5.9	3.3	5.3	4	5.1	4.2	4.9	-20.2	-19.2	-80.9	-79.8	19.7	21.3
2017/5/9	5.1	5.9	3.7	5.2	4.1	5.1	4.2	4.9	-19.6	-19.2	-80.9	-79.8	19.6	21.4
2017/5/10	4.9	5.9	3.8	5.2	4	4.9	4.2	5.3	-20.1	-19.3	-81	-79.9	18.4	20.2
2017/5/11	4.7	5.7	3.6	5.3	4.2	4.9	4.4	5.1	-20.1	-19.5	-81.2	-79.9	18	19.5
2017/5/12	5	5.7	3.7	5.2	4.1	4.9	4.4	5.1	-20.3	-19.7	-81.1	-79.8	18.3	19.5
2017/5/13	4.8	5.7	3.8	5.2	4.6	4.9	4.4	5.1	-20.4	-19.9	-81	-79.9	18.2	18.6
2017/5/14	4.7	5.8	3.8	5.3	4.6	4.9	4.4	5	-20.4	-20	-81.1	-79.9	18.2	18.9
2017/5/15	4.8	5.7	3.5	5.2	3.9	4.9	4.4	5.1	-20.4	-19.7	-81.1	-79.9	18.8	19.7
2017/5/16	5.1	6.1	3.6	5.2	3.7	4.9	4.4	4.9	-20.6	-19.8	-81	-79.8	18.6	20.3
2017/5/17	4.9	5.9	3.7	5.3	3.8	4.9	4.3	5.1	-20.8	-20.2	-81	-79.7	19.3	20.6
2017/5/18	5.1	6	3.8	5.3	3.9	4.9	4.3	4.9	-20.9	-18.9	-81.1	-79.9	19.4	20.8
2017/5/19	5.1	6	3.3	5.3	3.9	4.9	4.2	5.8	-20.7	-19.2	-81	-79.9	19.6	20.8
2017/5/20	4.9	5.8	3.4	5	4.6	5.1	4.2	4.9	-20.7	-20.4	-80.9	-79.9	19.6	19.9
2017/5/21	5.1	5.9	3.4	5	4.6	5.1	4.2	4.9	-20.7	-20.4	-81	-79.9	19.5	19.6
2017/5/22	5.1	5.9	3.2	5.1	3.9	4.9	4.2	4.9	-20.9	-20.4	-80.9	-79.8	19.4	22.1
2017/5/23	4.9	5.8	3.3	5.1	3.9	5.1	4.2	4.9	-20.9	-19.2	-80.9	-79.9	19.6	21.2
2017/5/24	4.8	5.9	3.3	5.1	4.2	5.1	4.1	4.9	-20.3	-19.3	-81	-79.9	19.7	21.2
2017/5/25	5.1	6.1	3.1	5.3	4.2	5.1	4.1	4.8	-20.3	-19.1	-80.9	-79.8	19	20.4
2017/5/26	4.4	5.8	3.7	5.3	4.2	5.1	4.1	4.9	-20.8	-19.1	-80.9	-79.9	19.2	20.3
2017/5/27	5.1	5.8	4.1	5.3	4.7	5.1	4.2	4.8	-21.4	-20.7	-80.9	-79.8	18.9	19.3
2017/5/28	4.9	5.7	4.1	5.3	4.7	5.1	4.2	4.8	-21.4	-21.1	-80.9	-79.9	18.8	18.9
2017/5/29	4.7	5.7	4.1	5.3	4.7	5.1	4.2	4.8	-21.4	-21.1	-80.9	-79.9	18.6	18.9
2017/5/30	4.7	5.7	4.1	5.3	4.7	5.1	4.2	4.9	-21.4	-21.1	-80.9	-79.9	18.6	18.8
2017/5/31	4.7	5.8	3.9	5.3	4.2	5.1	4	4.9	-21.4	-20.9	-80.9	-79.9	18.6	20.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部溫雅如 106
師(三)級藥師 0601

藥學部邱保祥 106
師(三)級藥師 0601