

臺北榮民總醫院藥學部

臨床試驗藥品冷藏/室溫溫度管制表

106 年 5 月

	一號冰箱		室溫	
序號	Serial No.1436-0086		Serial No.1118-0011	
備用	0436-049		0742-0169	
範圍	2 - 8 °C		15 - 25 °C	
日期	Min	Max	Min	Max
2017/5/1	3.3	4.1	19.3	19.7
2017/5/2	3.4	4.0	19.4	19.8
2017/5/3	3.3	4.2	19.3	19.7
2017/5/4	3.6	4.3	19.1	19.6
2017/5/5	3.6	4.7	18.7	19.3
2017/5/6	4.1	4.9	19.2	19.6
2017/5/7	3.7	4.5	18.7	19.5
2017/5/8	4.1	5.1	18.8	19.3
2017/5/9	3.8	4.2	18.8	20.0
2017/5/10	3.9	4.4	19.5	20.1
2017/5/11	3.8	4.0	19.4	20.1
2017/5/12	3.8	4.1	19.9	20.4
2017/5/13	3.8	4.1	19.8	20.1
2017/5/14	3.9	4.1	19.6	19.9
2017/5/15	3.8	4.1	19.4	20.1
2017/5/16	3.9	4.2	19.7	20.2
2017/5/17	3.8	4.2	20.1	20.4
2017/5/18	3.8	4.3	20.3	20.9
2017/5/19	3.9	4.6	20.4	20.9
2017/5/20	4.2	4.3	20.5	20.9
2017/5/21	4.2	4.3	20.7	20.9
2017/5/22	4.1	4.3	20.5	21.3
2017/5/23	4.1	4.3	20.2	21.0
2017/5/24	3.9	4.3	20.8	21.3
2017/5/25	3.9	4.2	21.1	21.4
2017/5/26	3.9	4.2	21.1	21.3
2017/5/27	3.9	4.2	20.9	21.2
2017/5/28	3.9	4.2	20.8	21.1
2017/5/29	3.9	4.2	20.8	21.1
2017/5/30	3.9	4.2	20.4	20.9
2017/5/31	3.9	4.2	20.6	20.9

註1: 溫度紀錄型號:ESCORT/Model No.: EI-IN-D-32-L

註2: 溫度紀錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

藥學部邵啟鈞
(三)級藥師